
Shear Modulus Testing Patterns





Special Use Sensors - Shear Modulus Testing Strain Gages

Shear Modulus Sensors are specifically designed to accommodate the unique specimen geometries and strainfield distributions encountered when testing composite materials for shear properties. Two popular specimens for in-plane shear modulus testing of composites are the losipescu and compact designs. The test section for both types is described as the area between two opposing notches. The losipescu specimen has a distance between the notch roots of 0.45 in [11.4 mm]; for the compact design this distance is 0.75 in [19 mm]. Both of these specimens have an inherently nonuniform shear-strain distribution in their test zone. Determining shear modulus requires extracting an average shear-strain value from this nonuniform strain field. Since strain gages have the unique characteristic of integrating the surface strain field under their grids, average specimen strain is automatically obtained by spanning the entire length of either specimen's test section.

Two 500Ω ±0.4%, ±45° shear-gage configurations are available for both the losipescu and compact specimen designs. The planar configuration, with side-by-side grids, is constructed with a standard N2 backing and Option SP61

(soft, copper lead ribbons and polyimide film encapsulation). The stacked configuration is produced with a special backing, A2. This backing is similar to N2 but is fully encapsulated with a polyimide film and includes integral soft, copper lead ribbons (like those provided by Option SP61). The stacked configuration is offered to best simulate strain measurement at a point. The stacked gages are supplied in a quarter-bridge arrangement so that independent gage measurements can be made if necessary. When connected in a half-bridge circuit, the stacked construction inherently provides temperature compensation and insensitivity to normal strains. Due to the increased stiffness of a stacked sensor, compared to one having only a single layer, an evaluation of the test conditions and requirements should be made to ensure that the gage will not compromise accuracy by significantly reinforcing low-modulus and/or thin specimens.

| GAGE PATTERN AND DESIGNATION Insert Desired S-T-C No. in Spaces Marked XX. ROHS COMPLIANT | | Pb-free | DIMENSIONS | | | | | inch millimeter |
|--|--|-----------|-----------------------------------|---------|----------------------|---------|--------|--------------------|
| | | RoHS | GAGE | OVERALL | GRID | OVERALL | MATRIX | |
| | | COMPLIANT | LENGTH | LENGTH | WIDTH | WIDTH | Length | Width |
| | N2A-XX-C032A-500/SP61 N2P-08-C032A-500/SP61 | | 0.032 | 0.462 | 0.031 | 0.197 | 0.500 | 0.257 |
| | | | 0.81 each section | 11.73 | 0.79 each section | 5.00 | 12.7 | 6.5 |
| | | | For use with losipescu specimens. | | | | | |
| | N2A-XX-C032B-500/SP61 N2P-08-C032B-500/SP61 | | 0.032 | 0.762 | 0.031 | 0.197 | 0.800 | 0.257 |
| | | | 0.81 each section | 19.35 | 0.79 each section | 5.00 | 20.3 | 6.5 |
| | | | For use with compact specimens. | | | | | |
| | A2A-XX-C085C-500 A2P-08-C085C-500 | | 0.085 | 0.445 | 0.070 | 0.200 | 0.500 | 0.260 |
| | | | 2.16 each section | 11.30 | 1.78 | 5.08 | 12.7 | 6.6 |
| | | | For use with losipescu specimens. | | | | | |
| | A2A-XX-C085D-500 A2P-08-C085D-500 | | 0.085 | 0.745 | 0.070 | 0.200 | 0.805 | 0.260 |
| | | | 2.16 each section | 18.92 | 1.78 | 5.08 | 20.4 | 6.6 |
| | | | For use with compact specimens. | | | | | |

Document Number: 11525 Revision: 02-Feb-10

Legal Disclaimer Notice



Vishay Precision Group

Disclaimer

All product specifications and data are subject to change without notice.

Vishay Precision Group, Inc., its affiliates, agents, and employees, and all persons acting on its or their behalf (collectively, "Vishay Precision Group"), disclaim any and all liability for any errors, inaccuracies or incompleteness contained herein or in any other disclosure relating to any product.

Vishay Precision Group disclaims any and all liability arising out of the use or application of any product described herein or of any information provided herein to the maximum extent permitted by law. The product specifications do not expand or otherwise modify Vishay Precision Group's terms and conditions of purchase, including but not limited to the warranty expressed therein, which apply to these products.

No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any conduct of Vishay Precision Group.

The products shown herein are not designed for use in medical, life-saving, or life-sustaining applications unless otherwise expressly indicated. Customers using or selling Vishay Precision Group products not expressly indicated for use in such applications do so entirely at their own risk and agree to fully indemnify Vishay Precision Group for any damages arising or resulting from such use or sale. Please contact authorized Vishay Precision Group personnel to obtain written terms and conditions regarding products designed for such applications.

Product names and markings noted herein may be trademarks of their respective owners.

Document Number: 63999 www.vishaypg.com Revision: 22-Feb-10