

# Supplement

Title: TH-1 Instruction Manual  
Part Number: 3311871  
Print Date: December 2011  
Revision/Date: 1, 12/11

Supplement Issue: 1  
Issue Date: 12/11  
Page Count: 1

---

This supplement contains information necessary to ensure the accuracy of the above manual

# **TH-1 Instruction Manual Supplement**

## **Change #1**

**On page #6, under Accuracy (after calibration) replace**

From: Humidity:  $\pm 3$  %RH (at 77 °F (25 °C), 10 ~ 95 %RH)

To: Humidity:  $\pm 4.0$  %RH (at 77 °F (25 °C), 10 ~ 95 %RH)

Add: Temperature Coefficient:  $0.1 \times$  (specified accuracy)/ °C ( < 25 °C or > 25 °C)