


Anritsu

Spectrum Master™

MS2711D

Fast. Accurate, Repeatable, Portable Spectrum Analysis



Programming Manual



SiteMaster

SpectrumMaster

CellMaster



MS2711D Programming Manual

MS2711D

Remote Mode Command Specification

Rev 1.45

May 2007
Copyright © 2007 Anritsu Co.

Part Number: 10580-00098
Revision: D

Table of Contents

TABLE OF CONTENTS	3
PROGRAMMING OVERVIEW	6
GENERAL DESCRIPTION.....	6
CABLES FOR THE SPECTRUM MASTER	6
SERIAL COMMUNICATION PARAMETERS	6
COMMUNICATIONS ERROR CHECKING	6
PARAMETER VALIDATION	6
ENTERING REMOTE MODE	7
EXITING REMOTE MODE	7
LIFETIME OF CHANGES TO SPECTRUM MASTER OPERATING PARAMETERS.....	7
WRITE CYCLE LIMITATION OF EEPROM.....	7
DOCUMENTATION CONVENTIONS.....	7
<i>Numeric Representation:</i>	7
<i>Bit Positions:</i>	7
CONTROL BYTE DESCRIPTIONS	8
SETUP SYSTEM – CONTROL BYTE #1 (01H).....	8
SELECT MEASUREMENT MODE – CONTROL BYTE #3 (03H)	8
READ TIME/DATE – CONTROL BYTE #7 (07H)	9
SET TIME/DATE – CONTROL BYTE #8 (08H).....	9
SET TRACE NAME (REFERENCE NUMBER) – CONTROL BYTE #9 (09H)	10
SERIAL PORT ECHO ON/OFF – CONTROL BYTE #10 (0AH).....	10
WATCH-DOG TIMER ON/OFF – CONTROL BYTE #12 (0CH)	11
STORE SWEEP TRACE – CONTROL BYTE #16 (10H).....	11
OBSOLETE: RECALL SWEEP TRACE – CONTROL BYTE #17 (11H)	11
SAVE SYSTEM SETUP – CONTROL BYTE #18 (12H)	15
RECALL SYSTEM SETUP – CONTROL BYTE #19 (13H)	16
OBSOLETE: QUERY SYSTEM STATUS – CONTROL BYTE #20 (14H)	16
TRIGGER SELF-TEST – CONTROL BYTE #21 (15H).....	21
READ FAIL COUNTER – CONTROL BYTE #22 (16H)	21
CLEAR FAIL COUNTERS – CONTROL BYTE #23 (17H).....	21
QUERY TRACE NAMES – CONTROL BYTE #24 (18H)	22
DELETE SWEEP TRACE – CONTROL BYTE #25 (19H).....	22
OBSOLETE - UPLOAD SPA SWEEP TRACE – CONTROL BYTE #26 (1AH).....	22
QUERY SWEEP MEMORY – CONTROL BYTE #27 (1BH).....	26
QUERY SYSTEM STATUS – CONTROL BYTE #29 (1DH)	26
SELECT PRINTER TYPE – CONTROL BYTE #30 (1EH).....	35
RECALL SWEEP TRACE – CONTROL BYTE #33 (21H).....	35
OBSOLETE: SET A/B TRACE – CONTROL BYTE #34 (22H)	46
SET SPA A/B TRACE – CONTROL BYTE #35 (23H).....	46
UPLOAD SPECTRUM MASTER SWEEP TRACE – CONTROL BYTE #36 (24H).....	47
GET OPTIONS – CONTROL BYTE #37 (25H)	56
QUERY POWER LEVEL – CONTROL BYTE #39 (27H).....	56
SET POWER METER UNITS – CONTROL BYTE #40 (28H)	57
POWER METER RELATIVE MODE ON/OFF – CONTROL BYTE #41 (29H).....	57
POWER METER OFFSET MODE ON/OFF – CONTROL BYTE #42 (2AH)	57
POWER METER ZERO MODE ON/OFF – CONTROL BYTE #43 (2BH)	58
POWER METER RMS AVERAGING ON/OFF – CONTROL BYTE #44 (2CH)	58
POWER METER CENTER FREQUENCY AND SPAN SETUP – CONTROL BYTE #45 (2DH)	59
SWEEP DATA ECHO ON/OFF – CONTROL BYTE #49 (31H)	59
CHECK BATTERY STATUS – CONTROL BYTE #50 (32H)	60
SET SPA MINIMUM SWEEP TIME – CONTROL BYTE #53 (35H)	61

SET TRIGGER POSITION – CONTROL BYTE #54 (36H)	61
SET VIDEO TRIGGER LEVEL – CONTROL BYTE #55 (37H)	61
AUTOMATICALLY SAVE RUNTIME SETUP – CONTROL BYTE #64 (40H)	62
DOWNLOAD SAVED SETUP – CONTROL BYTE #65 (41H)	62
UPLOAD SETUP – CONTROL BYTE #66 (42H)	72
READ PARAMETER LIMITS – CONTROL BYTE #67 (43H)	82
QUERY SAVED SETUPS – CONTROL BYTE #68 (44H)	89
ENTER REMOTE MODE – CONTROL BYTE #69 (45H)	90
ENTER REMOTE MODE IMMEDIATELY – CONTROL BYTE #70 (46H)	90
WRITE PROTECT SETUP – CONTROL BYTE #71 (47H)	91
CLEAR SETUP MEMORY LOCATION – CONTROL BYTE #72 (48H)	91
WRITE ANTENNA – CONTROL BYTE #82 (52H)	92
RECALL ANTENNA – CONTROL BYTE #83 (53H)	93
SET FIELD STRENGTH MEASUREMENT – CONTROL BYTE #84 (54H)	93
SET CHANNEL POWER – CONTROL BYTE #85 (55H)	94
READ CHANNEL POWER – CONTROL BYTE #86 (56H)	94
SET ADJACENT CHANNEL POWER RATIO (ACPR) – CONTROL BYTE #87 (57H)	95
READ ADJACENT CHANNEL POWER (ACPR) – CONTROL BYTE #88 (58H)	95
SELECT SIGNAL STANDARD – CONTROL BYTE #89 (59H)	96
SELECT CHANNEL IN CURRENT STANDARD – CONTROL BYTE #90 (5AH)	96
READ CURRENT SETTINGS OF SIGNAL STANDARD – CONTROL BYTE #91 (5BH)	96
UPLOAD USER'S SIGNAL STANDARD TABLE – CONTROL BYTE #92 (5CH)	97
MEASURE OCC BW % OF POWER – CONTROL BYTE #96 (60H)	97
MEASURE OCC BW dB DOWN – CONTROL BYTE #97 (61H)	98
SET BIAS TEE FUNCTION – CONTROL BYTE #98 (62H)	98
SET SPECTRUM ANALYZER START/STOP FREQUENCY – CONTROL BYTE #99 (63H)	99
SET SPECTRUM ANALYZER CENTER FREQ./SPAN – CONTROL BYTE #100 (64H)	99
SET SPECTRUM ANALYZER SCALE – CONTROL BYTE #101 (65H)	100
SET SPECTRUM ANALYZER MARKER – CONTROL BYTE #102 (66H)	100
SET SPECTRUM ANALYZER SINGLE LIMIT – CONTROL BYTE #103 (67H)	101
OBSOLETE: SET SPA RESOLUTION BANDWIDTH – CONTROL BYTE #106 (6AH)	101
OBSOLETE: SET SPECTRUM ANALYZER VIDEO BANDWIDTH – CONTROL BYTE #107 (6BH)	102
SET SPECTRUM ANALYZER SWEEP MODE – CONTROL BYTE #108 (6CH)	102
SET SPECTRUM ANALYZER MARKER TO PEAK – CONTROL BYTE #109 (6DH)	102
SET SPECTRUM ANALYZER MARKER TO CENTER – CONTROL BYTE #110 (6EH)	103
OBSOLETE: SET SPECTRUM ANALYZER ATTENUATION – CONTROL BYTE #111 (6FH)	103
SET SPECTRUM ANALYZER MULTIPLE LIMIT – CONTROL BYTE #113 (71H)	104
SET RETURN SPECTRUM ANALYZER SWEEP TIME – CONTROL BYTE #114 (72H)	105
SET REFERENCE LEVEL OFFSET – CONTROL BYTE #115 (73H)	105
SET SPECTRUM ANALYZER IMPEDANCE – CONTROL BYTE #116 (74H)	106
READ MARKER VALUE – CONTROL BYTE #117 (75H)	107
SET SWEEP AVERAGING – CONTROL BYTE #118 (76H)	107
NORMALIZE SPA – CONTROL BYTE #130 (82H)	108
SET SPA NORMALIZATION STATUS – CONTROL BYTE #131 (83H)	108
CHECK SPECTRUM ANALYZER EXTERNAL REFERENCE – CONTROL BYTE #134 (86H)	108
SET SA PREAMP STATE (ON/OFF/AUTO) – CONTROL BYTE #136 (88H)	109
SET SPECTRUM ANALYZER UNITS – CONTROL BYTE #140 (8CH)	109
SET SPECTRUM ANALYZER RESOLUTION BANDWIDTH – CONTROL BYTE #141 (8DH)	110
SET SPECTRUM ANALYZER VIDEO BANDWIDTH – CONTROL BYTE #142 (8EH)	110
SET SPECTRUM ANALYZER ATTENUATION – CONTROL BYTE #143 (8FH)	110
SET AM/FM DEMODULATION – CONTROL BYTE #145 (91H)	111
SET BAUD RATE – CONTROL BYTE #197 (C5H)	111
SET LANGUAGE – CONTROL BYTE #198 (C6H)	112
QUERY TIME – CONTROL BYTE #208 (D0H)	112
READ MAIN SERIAL NUMBER – CONTROL BYTE #221 (DDH)	112
READ ASCII SERIAL NUMBER – CONTROL BYTE #225 (E1H)	113

GPS POWER – CONTROL BYTE #237 (EDH).....	113
READ GPS POSITION – CONTROL BYTE #238 (EEH).....	113
AUTOMATIC CAL DISABLE – CONTROL BYTE #240 (F1H).....	114
RECALL SWEEP TRACE – CONTROL BYTE #243 (F3H).....	114
EXIT REMOTE MODE – CONTROL BYTE #255 (FFH)	114
SELECT SPA/POWER METER SIGNAL STANDARD – CONTROL WORD (A103H).....	115
SELECT SPA/POWER METER CHANNEL – CONTROL WORD (A104H).....	115
READ EXTERNAL MODULE NAME – CONTROL WORD (A201H).....	115
READ EXTERNAL MODULE SERIAL NUMBER – CONTROL WORD (A202H).....	116
READ EXTERNAL MODULE FREQUENCY RANGE – CONTROL WORD (A203H)	116
CLEAR MODULE FAIL COUNTER – CONTROL WORD (A205H)	117
PERFORM TRANSMISSION MODE CALIBRATION – CONTROL WORD (A301H)	117
TURN OFF TRANSMISSION MODE CALIBRATION – CONTROL WORD (A302H)	117
GET SIGNAL STANDARD NAME – CONTROL WORD (A501H)	118
PERFORM NOISE DIODE CAL – CONTROL WORD (A505H)	118
SET BIAS T VOLTAGE – CONTROL WORD (A506H).....	118
SELECT FUNCTION IN IA MEASUREMENT MODE – CONTROL WORD (A700H)	118
SPECTROGRAM: SET SWEEP INTERVAL – CONTROL WORD (A721H).....	119
SPECTROGRAM: SET TIME SPAN OF MEASUREMENT – CONTROL WORD (A722H)	119
SPECTROGRAM: TURN ON/OFF AUTO SAVE – CONTROL WORD (A723H)	119
SPECTROGRAM: GET TRACE NAME – CONTROL WORD (A724H)	120
SPECTROGRAM: RECALL TRACE – CONTROL WORD (A725H).....	121
SET FREQUENCY IN CW GENERATOR MODE - CODE WORD(A801H).....	121
RETRIEVE POWER IN CW GENERATOR MODE - CODE WORD(A802H).....	122
REMOTE SELF TEST – CONTROL WORD (AA15H)	122
TRIGGER SWEEP – CONTROL WORD (AA30H)	124
PARAMETER DEFINITIONS	125
PROGRAMMING EXAMPLES.....	126
EXAMPLES IN C:.....	126
EXAMPLE IN VISUAL BASIC.....	133
REVISION HISTORY	135

Programming Overview

Warning: This programming manual is written exclusively for Anritsu Spectrum Master Model MS2711D. For information on firmware upgrades, contact your local Anritsu Service Center. Commands listed in this manual are not all backward-compatible with earlier Anritsu models.

General Description

The Spectrum Master must first be set into “remote” mode for communication with a computer. Remote mode differs from normal repetitive sweep and single-sweep modes. During remote mode, the Spectrum Master suspends normal operations and attends to the serial port. The front panel display indicates when the Spectrum Master is in remote mode.

Once in remote mode, you send a series of control bytes and associated data to the Spectrum Master. These control byte sequences command the Spectrum Master to perform various functions and activities. The serial port supports virtually all features accessible from the keypad. The only exception is the printer, which requires connection to the same 9 pin connector on the Spectrum Master rear panel.

To complete the communication session, send the control byte to exit remote mode. Spectrum Master resumes normal operations. You may also exit the remote mode by using the ESCAPE/CLEAR key.

Cables for the Spectrum Master

Serial communications take place via the 9 pin connector on the back of the Spectrum Master. The Spectrum Master is a DTE-type serial device and therefore requires a “null modem” cable for communication with a computer, which is also a DTE device. We provide a suitable cable with your Spectrum Master. (Anritsu part number 800-441)

Serial Communication Parameters

The Spectrum Master communicates at a baud rate of 9600. It uses no parity bits, 8 data bits, and 1 stop bit (N-8-1). No hardware handshaking is used. The Set Baud Rate serial command Control Byte #197 (C5h) can be used to change the baud rate to other common baud rates.

Communications Error Checking

Since there is no hardware handshaking, byte level error handling must be done by the controlling program. The expected number of response bytes for each control byte (listed in the control byte description section of this manual) works well for responses coming from the Spectrum Master. For data streams going to the Spectrum Master, the “watch dog timer” protects against interrupted transmissions by aborting a control byte sequence if the inter-byte time limit is exceeded.

Parameter Validation

The Spectrum Master validates input parameters for each control byte sequence. If the input parameters are out of range or invalid, the Spectrum Master notifies the computer by sending Parameter Error Byte #224 (E0h). The Spectrum Master discards the received data and waits for the next control byte.

Entering Remote Mode

Send the Enter Remote Mode Byte #69 (45h) to the Spectrum Master to enter remote mode at the end of a sweep
Send the Enter Remote Mode Immediately byte #70 (46h) to enter remote mode in the middle of a sweep.

The Spectrum Master's serial port buffer is one byte wide. No internal buffer exists, so waiting for the unit's response is essential. If the Spectrum Master is not in remote, sending a second byte overwrites the original byte commanding it to go into remote. If you send control byte #69, you must wait until the end of the sweep. If you send control byte #70, the unit will enter remote mode as soon as it receives the byte. Note that this means that data stored for the current sweep may be incomplete.

Once you receive the response string from Spectrum Master, you are in remote mode.

Exiting Remote Mode

Send the Exit Remote control byte #255 (FFh) to the Spectrum Master. Spectrum Master sends a response byte of 255 (FFh) then exits remote mode. Remote mode can also be exited by pressing the ESCAPE/CLEAR key.

Lifetime of Changes to Spectrum Master Operating Parameters

System parameters changed during remote mode remain changed for normal operation. They are not automatically written to the non-volatile EEPROM. Turning off power erases the changed settings.

If you want the changes saved, you must save the change to one of the setup memories. Use either the run-time setup (location 0, which holds the power-on defaults) or one of the nine saved setups. See control byte #18 (12h) for details.

Write Cycle Limitation of EEPROM

The EEPROM, used to store calibrations, setups and traces has a guaranteed lifetime of at least 100,000 write cycles and an unlimited number of read cycles. The write cycle limitation is for a specific location. For example, you can store setup #1 100,000 times and setup #2 100,000 times, etc.

It is for this reason we do not automatically store the changed system parameters to EEPROM. Instead, we provide a means of changing the operating parameters independent of this limitation.

Be aware of the EEPROM write cycle limitation when programming the Spectrum Master. Keep the number of write cycles to a minimum.

Documentation Conventions

Through this manual the following conventions will be observed:

Numeric Representation:

Hexadecimal numbers are represented with the suffix h. For example, the decimal number 255 is represented in hexadecimal as FFh.

Binary numbers are represented with the suffix b. For example, the decimal number 2 is represented in binary as 10b.

Decimal numbers are represented with the prefix # when referring to a control byte (command byte) and without a prefix or suffix in all other cases.

Bit Positions:

When enumerating bits in a byte, bit 0 will always be the least significant bit (LSB).

Control Byte Descriptions

Setup System – Control Byte #1 (01h)

Description: Sets system status flags and switches. The current value of the flags can be obtained by executing command #29, Query System Setup, and parsing the values from the appropriate bytes. The Spectrum Master acts on the entire byte. So, the state of each of the bits must be defined every time the command is issued. See control byte #29 (1Dh) response bytes 275 and 276 for current Spectrum Master configuration.

Bytes to Follow: 1 byte

1) Status byte

- bit 0 = RBW Coupling (to span) (1b = auto 0b = manual)
- bit 1 = VBW Coupling (to RBW) (1b = auto 0b = manual)
- bit 2 = LCD Back Light ON/OFF (1b = ON 0b = OFF)
- bits 3-4 = Logarithmic Amplitude units (00b = dBm 01b = dBV 10b = dBmV 11b = dBuV)
- bits 5-6 = Detection algorithm (00b = positive peak 01b = RMS average 10b = negative peak 11b = Sampling)
- bit 7 = Attenuation Coupling (to ref level) (1b = auto 0b = manual)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 238 (EEh) Time-out Error
-

Select Measurement Mode – Control Byte #3 (03h)

Description: Sets the measurement mode of the Spectrum Master. The response byte will not be sent until the mode change is complete.

See control byte #29 (1Dh) response byte 3 for the current Spectrum Master measurement mode.

Bytes to Follow: 1 byte

1) Measurement Mode

- 30h: Spectrum Analyzer Mode
- 31h: Transmission Mode
- 39h: Channel Scanner Mode
- 3Bh: Interference Analyzer Mode
- 3Ch: CW Signal Generator Mode
- 40h: Power Meter Mode (narrow band)
- 41h: Power Monitor Mode (Option 5)
- 42h: High Accuracy Power Meter Mode

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid measurement mode
 - 238 (Eeh) Time-out Error
-

Read Time/Date – Control Byte #7 (07h)

Description: Reads the current time and date from the real time clock.

This Time/Date is stamped into all stored sweeps (for users' reference). The real time clock time and date can be set using control byte #8.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 7 bytes

- 1) Hour
 - 2) Minute
 - 3) Month
 - 4) Day
 - 5) Year (Higher byte)
 - 6) Year (Lower byte)
 - 7) Daylight Saving On/Off (01h = On, 00h = Off)
-

Set Time/Date – Control Byte #8 (08h)

Description: Sets the current time and date.

This Time/Date is stamped into all stored sweeps (for users' reference).

The Spectrum Master stores bytes as ASCII text. Recommended time form is "hh:mm:ss" (hour:minute:sec). Recommended date format is "mm/dd/yyyy" (month/day/year).

The current time setting can be found by using control byte #33 to recall trace 0 and examining response bytes 31-38.

The current date setting can be found by using control byte #33 to recall trace 0 and examining response bytes 21-30.

Bytes to Follow: 7 bytes

- 1) Hour
- 2) Minute
- 3) Month
- 4) Day
- 5) Year (Higher byte)
- 6) Year (Lower byte)
- 7) Daylight Saving (01h = On, 00h = Off)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 2) 238 (Eeh) Time-out Error
-

Set Trace Name (Reference Number) – Control Byte #9 (09h)

Description: Stores a Reference Number with the sweep trace.

The reference number is also known as the trace name. It is any combination of 16 letters, numbers and the characters “-“, “;”, “.” And “+”. This command stores a trace name with the sweep trace.

The current reference number is found by recalling trace 0 and examining response bytes 39 to 54.

Bytes to Follow: 16 bytes (ASCII text string)

Spectrum Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
 - 238 (Eeh) Time-out Error
-

Serial Port Echo On/Off – Control Byte #10 (0Ah)

Description: Sets the serial port echo mode On/Off.

Serial Port Echo Mode uses the **single sweep** mode (see control byte #11 (0Bh)). At the end of each sweep cycle, the Spectrum Master sends a Sweep Complete Byte #192 (C0h) to the serial port.

This mode activates once the Spectrum Master exits from the remote mode. Serial Port Echo status can't be saved to or recalled from saved setups. Cycling power resets the Serial port echo status to Off.

The Serial Port Echo Mode allows run-time handshaking between the Spectrum Master and computer by doing the following:

- 1) Enter remote mode. Set Serial Port Echo Mode On. Exit remote mode.
- 2) The Spectrum Master sweeps once and then sends the Sweep Complete Byte.
- 3) After you receive it. Enter remote mode. Recall sweep 0 (last sweep trace in RAM).
- 4) Exit remote mode. Send Sweep Triggering Byte #48 (30h) and wait for the next sweep cycle.
- 5) Repeat steps 2-4

Bytes to Follow: 1 byte

- 1) Serial Port Echo Status
 - 00h : Off
 - 01h : On

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid serial port echo status
 - 238 (Eeh) Time-out Error
-

Watch-Dog Timer On/Off – Control Byte #12 (0Ch)

Description: Enables or disables the Watch-dog timer. Default is Disabled.

The Spectrum Master incorporates a watch-dog timer for higher reliability in serial communication. In selected control bytes (see control byte summary), the Spectrum Master checks for the time interval between each byte received from the computer. If the time interval exceeds the set time limit (0.5 sec), the Spectrum Master notifies the computer by sending Time-out Byte #238 (EEh). The Spectrum Master discards the data it just received and then waits for the next control byte sequence.

Bytes to Follow: 1 byte

- 1) Watch-dog timer On/Off
 - 00h = Off
 - 01h = On

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid watch-dog timer status
-

Store Sweep Trace – Control Byte #16 (10h)

Description: Saves current trace to the next available memory location. Trace name can be set using control byte #9, “Set Trace Name (Reference Number)” before executing this command.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 5 bytes

- | | |
|------|---|
| 1-4) | Time/Date Stamp (In long integer format) |
| 5) | Operation result:
255 (FFh) Operation Complete Byte
224 (E0h) Out of memory (Memory full)
238 (Eeh) Time-out Error |
-

OBSOLETE: Recall Sweep Trace – Control Byte #17 (11h)

This command exists for backward compatibility with MS2711B models. To access the new features use Control Byte #33 (21h). This command cannot be used with a frequency extension module (option 6 required) attached.

Description: Queries the Spectrum Master for sweep trace data.

NOTE: Before you can recall a sweep stored in non-volatile memory (trace numbers 1-200) you must build a trace table in the Spectrum Master’s RAM. Use Control Byte #24 to build the trace table. Since the trace table exists in RAM, Control Byte #24 must be executed every time the Spectrum Master’s power is cycled.

Bytes to Follow: 1 byte

- 0 = Last sweep trace before entering remote mode (sweep trace in RAM)
- 1- 200 = Specific saved sweep number (stored sweeps in Flash memory)

Spectrum Master Returns: 1950 bytes

- 1-2) # of following bytes (1948 for a valid sweep)
- 3-4) Model ID (16h for the MS2711D)
- 5-11) Model Number (7 bytes in ASCII)
- 12-15) Software Version (4 bytes in ASCII)
- 16) Measurement Mode¹

¹ See Control Byte #3 for available measurement modes.

17-20)	Time/Date (long integer format)
21-30)	Date in String Format (mm/dd/yyyy)
31-38)	Time in String Format (hh:mm:ss)
39-54)	Reference Number/Trace Name (16 bytes in ASCII)
55-56)	# data points (400)
57)	Start Frequency (in Hz) (Highest byte)
58)	Start Frequency (in Hz)
59)	Start Frequency (in Hz)
60)	Start Frequency (in Hz) (Lowest byte)
61)	Stop Frequency (in Hz) (Highest byte)
62)	Stop Frequency (in Hz)
63)	Stop Frequency (in Hz)
64)	Stop Frequency (in Hz) (Lowest byte)
65)	Center Frequency (in Hz) (Highest byte)
66)	Center Frequency (in Hz)
67)	Center Frequency (in Hz)
68)	Center Frequency (in Hz) (Lowest byte)
69)	Frequency Span (in Hz) (Highest byte)
70)	Frequency Span (in Hz)
71)	Frequency Span (in Hz)
72)	Frequency Span (in Hz) (Lowest byte)
73)	Minimum Frequency Step Size (in Hz) (Highest byte)
74)	Minimum Frequency Step Size (in Hz)
75)	Minimum Frequency Step Size (in Hz)
76)	Minimum Frequency Step Size (in Hz) (Lowest byte)
77)	Ref Level ² (Highest byte)
78)	Ref Level
79)	Ref Level
80)	Ref Level (Lowest byte)
81)	Scale per div ³ (Highest byte)
82)	Scale per div
83)	Scale per div
84)	Scale per div (Lowest byte)
85)	Frequency Marker 1 ⁴ (Higher byte)
86)	Frequency Marker 1 (Lower byte)
87)	Frequency Marker 2 (Higher byte)
88)	Frequency Marker 2 (Lower byte)
89)	Frequency Marker 3 (Higher byte)
90)	Frequency Marker 3 (Lower byte)
91)	Frequency Marker 4 (Higher byte)
92)	Frequency Marker 4 (Lower byte)
93)	Frequency Marker 5 (Higher byte)
94)	Frequency Marker 5 (Lower byte)
95)	Frequency Marker 6 (Higher byte)
96)	Frequency Marker 6 (Lower byte)
97)	Single Limit ⁵ (Highest byte)
98)	Single Limit
99)	Single Limit

² “value” sent as (value in dBm * 1,000) + 270,000

³ “value” sent as (value * 1,000)

⁴ Display/Data Point

To convert from “point” to frequency:

((span / (#data points-1)) * point) + start frequency

where span is stored in bytes 69-72 and #data points is stored in bytes 55-56

⁵ “value” sent as (value in dBm * 1,000) + 270,000

- 100) Single Limit (Lowest byte)
- 101) Multiple Upper Limit 1 Start X (Frequency in Hz) (Highest byte)
- 102) Multiple Upper Limit 1 Start X (Frequency in Hz)
- 103) Multiple Upper Limit 1 Start X (Frequency in Hz)
- 104) Multiple Upper Limit 1 Start X (Frequency in Hz) (Lowest byte)
- 105) Multiple Upper Limit 1 Start Y (Power Level⁶) (Highest byte)
- 106) Multiple Upper Limit 1 Start Y (Power Level)
- 107) Multiple Upper Limit 1 Start Y (Power Level)
- 108) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 109) Multiple Upper Limit 1 End X (Frequency in Hz) (Highest byte)
- 110) Multiple Upper Limit 1 End X (Frequency in Hz)
- 111) Multiple Upper Limit 1 End X (Frequency in Hz)
- 112) Multiple Upper Limit 1 End X (Frequency in Hz) (Lowest byte)
- 113) Multiple Upper Limit 1 End Y (Power Level⁷) (Highest byte)
- 114) Multiple Upper Limit 1 End Y (Power Level)
- 115) Multiple Upper Limit 1 End Y (Power Level)
- 116) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 117-260) Multiple Upper Limits 2-5, Multiple Lower Limits 1-5 (see bytes 101-116 for format)
- 261) RBW Setting (Frequency in Hz) (Highest byte)
- 262) RBW Setting (Frequency in Hz)
- 263) RBW Setting (Frequency in Hz)
- 264) RBW Setting (Frequency in Hz) (Lowest byte)
- 265) VBW Setting (Frequency in Hz) (Highest byte)
- 266) VBW Setting (Frequency in Hz)
- 267) VBW Setting (Frequency in Hz)
- 268) VBW Setting (Frequency in Hz) (Lowest byte)
- 269) OCC BW Method (00h = % of power, 01h = dB down)
- 270) OCC BW % Value (0-99) (Highest byte)
- 271) OCC BW % Value (0-99)
- 272) OCC BW % Value (0-99)
- 273) OCC BW % Value (0-99) (Lowest byte)
- 274) OCC BW dBc (0-120) (Highest byte)
- 275) OCC BW dBc (0-120)
- 276) OCC BW dBc (0-120)
- 277) OCC BW dBc (0-120) (Lowest byte)
- 278) Attenuation⁸ (Highest byte)
- 279) Attenuation
- 280) Attenuation
- 281) Attenuation (Lowest byte)
- 282-297) Antenna Name (16 bytes in ASCII)
- 298) Reference Level Offset⁹ (Highest byte)
- 299) Reference Level Offset
- 300) Reference Level Offset
- 301) Reference Level Offset (Lowest byte)
- 302) Impedance¹⁰
- 303) Impedance Loss¹¹ (Highest byte)

⁶ “value” sent as (value in dBm * 1,000) + 270,000

⁷ “value” sent as (value in dBm * 1,000) + 270,000

⁸ “value” sent as (value * 1,000)

⁹ “value” sent as (value in dBm * 1,000) + 270,000

¹⁰ Impedance adapters:

00h = 50 Ω

0Ah = 75 Ω, adapter 12N50-75B

0Ch = 75 Ω, other adapter offset

¹¹ “value” sent as (value * 1,000)

304)	Impedance Loss
305)	Impedance Loss
306)	Impedance Loss (Lowest byte)
307)	N/A
308)	N/A
309)	N/A
310)	N/A
311)	N/A
312)	N/A
313)	N/A
314)	N/A
315)	Status Byte 1: (0b = Off, 1b = On) (LSB) bit 0 : Marker 1 On/Off bit 1 : Marker 2 On/Off bit 2 : Marker 3 On/Off bit 3 : Marker 4 On/Off bit 4 : Marker 5 On/Off bit 5 : Marker 6 On/Off bits 6-7 : Not Used
316)	Status Byte 2: (0b = Off, 1b = On) (LSB) bit 0 : Marker 2 Delta On/Off bit 1 : Marker 3 Delta On/Off bit 2 : Marker 4 Delta On/Off bits 3-7: Not Used
317)	Status Byte 3: (0b = Off, 1b = On) (LSB) bit 0 : Antenna Factor Correction On/Off bits 1-2 : Detection alg (00b = pos. peak 10b = neg. peak 11b = RMS averaging) bits 3-4 : Logarithmic Amplitude Units (00b = dBm 01b = dBV 10b = dBmV11b = dBuV) bit 5 : Channel Power On/Off bit 6 : Adjacent Channel Power On/Off bit 7 : Occupied Bandwidth On/Off
318)	Status Byte 4: (0b = Off/LOWER limit, 1b = On/UPPER limit) (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple) bit 1 : Not Used bit 2 : Single Limit On/Off bit 3 : Single Limit Level UPPER/ LOWER bit 4 : Multiple Limit Upper Segment 1 Status On/Off bit 5 : Multiple Limit Upper Segment 1 Limit Level UPPER / LOWER bit 6 : Multiple Limit Upper Segment 2 Status On/Off bit 7 : Multiple Limit Upper Segment 2 Limit Level UPPER / LOWER
319)	Status Byte 5: (0b = Off/LOWER limit, 1b = On/UPPER limit) (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off bit 1 : Multiple Limit Upper Segment 3 Limit Level UPPER / LOWER bit 2 : Multiple Limit Upper Segment 4 Status On/Off bit 3 : Multiple Limit Upper Segment 4 Limit Level UPPER / LOWER bit 4 : Multiple Limit Upper Segment 5 Status On/Off bit 5 : Multiple Limit Upper Segment 5 Limit Level UPPER / LOWER bit 6 : Multiple Limit Lower Segment 1 Status On/Off bit 7 : Multiple Limit Lower Segment 1 Limit Level UPPER / LOWER
320)	Status Byte 6: (0b = Off/LOWER limit, 1b = On/UPPER limit) (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off bit 1 : Multiple Limit Lower Segment 2 Limit Level UPPER / LOWER bit 2 : Multiple Limit Lower Segment 3 Status On/Off bit 3 : Multiple Limit Lower Segment 3 Limit Level UPPER / LOWER bit 4 : Multiple Limit Lower Segment 4 Status On/Off

bit 5 : Multiple Limit Lower Segment 4 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Limit Level UPPER / LOWER
 321) Status Byte 7
 (LSB) bits 0-6 : Number of Sweeps to Average (1-25, 1 implies averaging Off)
 bit 7 : Reserved
 322) Status Byte 8: (0b = Off, 1b = On)
 (LSB) bit 0 : Preamp On/Off
 bit 1 : Normalization On/Off
 bit 2 : Bias Tee On/Off
 bit 3 : Scale Type (0b = Log, 1b = Linear)
 bit 4 : Linear Units (0b = Watts, 1b = Volts)
 bits 5-7 : Date Format (000b = MMDDYYYY, 001b = DDMMYYYY, 010b =
 YYYYMMDD)
 323) N/A
 324) N/A
 325-350) Not Used
 351-1950) Sweep Data (400 points * 4 bytes/point = 1600 bytes)
 4 bytes for each data point
 1. dBm² (Highest byte)
 2. dBm
 3. dBm
 4. dBm (Lowest byte)

Spectrum Master Returns (For invalid sweeps/empty stored sweep locations): 11 bytes

- 1-2) Number of following bytes (9 bytes for invalid sweep recall)
- 3-4) Model # (unsigned integer, 16h for Spectrum Master model MS2711D)
- 5-11) Extended Model # (7 bytes in ASCII)

Spectrum Master Returns (Invalid sweep location): 1 byte

- 1) 224 Parameter Error: Invalid sweep location
-

Save System Setup – Control Byte #18 (12h)

Description: Saves current system setup parameters to a specific setup store location.

The Spectrum Master saves all parameters described in Query System Status - Control Byte #29 (1Dh), (except Serial Port Echo Status) to the specified store location. Store location 0 is the run-time setup of the Spectrum Master. It holds the power-on defaults of the Spectrum Master.

Bytes to Follow: 1 byte

- 1) Location to save system setup parameters:
 - 0 – 10 for Spectrum Analyzer and Transmission Modes (Option 21 Only)
 - 0 – 5 for Power Meter Mode (Option 29 Only)
 - 0 – 5 for T1/E1 Modes (Option 50 Only)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid store location
 - 238 (EEh) Time-out Error
-

Recall System Setup – Control Byte #19 (13h)

Description: Recalls system setup parameters from a specific store location. Storage locations depend on the measurement mode of the current setup. When the current mode is Spectrum Analyzer, Spectrum Analyzer setups (1-10) can be recalled. When the current mode is Power Meter, one of the 5 Power Meter setups can be recalled.. When the current mode is T1/E1, one of the T1/E1 setups can be recalled (1-5).

The Spectrum Master recalls all parameters described in Query System Status - Control Byte #29 (1Dh), (except Serial Port Echo Status) from the specified store location. The recalled setup does **not** automatically become the power-on runtime setup when exiting remote. Therefore, a call to #29 will not display the parameters in that setup.

You may want to save the recalled setup as the run-time setup by saving it to setup location 0 (which holds the power-on runtime setup). See control byte #18 (12h) for details.

Bytes to Follow: 1 byte

- 1) Location from which to recall system setup parameters:

0	= Run time setup for all measurement modes
1 – 10	= Saved setups for Spectrum Analyzer and Transmission modes (Option 21 Only)
1 – 5	= Saved setups for Power Meter mode (Option 29 Only)
1 – 5	= Saved setups for T1/E1 modes (Option 50 Only)
254	= Default setup, current mode
255	= Default setup, all modes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid store location or no saved setup
 - 227 (E3h) Frequency Mismatch Error
 - 238 (EEh) Time-out Error
-

OBSOLETE: Query System Status – Control Byte #20 (14h)

This command exists for backward compatibility with the MS2711B To access the new features use Control Byte #29 (1Dh). This command cannot be used with a frequency extension module (option 6 required) attached.

Description: Queries the Spectrum Master for current system settings.

The current state of the Spectrum Master represents the state after the last successful remote control operation. For example, change the start frequency to another valid frequency while in remote mode, then execute control byte #29. The new start frequency will be returned in bytes 4-7, even though no sweep has been performed with that frequency.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 310 bytes

- 1) Measurement Mode¹²
- 2-3) Number of Data Points (400)
- 4) Start Frequency (in Hz) (Highest byte)
- 5) Start Frequency (in Hz)
- 6) Start Frequency (in Hz)
- 7) Start Frequency (in Hz) (Lowest byte)
- 8) Stop Frequency (in Hz) (Highest byte)
- 9) Stop Frequency (in Hz)
- 10) Stop Frequency (in Hz)
- 11) Stop Frequency (in Hz) (Lowest byte)

¹² See Control Byte #3 for available measurement modes.

- 12) Center Frequency (in Hz) (Highest byte)
- 13) Center Frequency (in Hz)
- 14) Center Frequency (in Hz)
- 15) Center Frequency (in Hz) (Lowest byte)
- 16) Frequency Span (in Hz) (Highest byte)
- 17) Frequency Span (in Hz)
- 18) Frequency Span (in Hz)
- 19) Frequency Span (in Hz) (Lowest byte)
- 20) Minimum Frequency Step Size (in Hz) (Highest byte)
- 21) Minimum Frequency Step Size (in Hz)
- 22) Minimum Frequency Step Size (in Hz)
- 23) Minimum Frequency Step Size (in Hz) (Lowest byte)
- 24) Ref Level¹³ (Highest byte)
- 25) Ref Level
- 26) Ref Level
- 27) Ref Level (Lowest byte)
- 28) Scale per div¹⁴ (Highest byte)
- 29) Scale per div
- 30) Scale per div
- 31) Scale per div (Lowest byte)
- 32) Marker 1¹⁵ (Higher byte)
- 33) Marker 1 (Lower byte)
- 34) Marker 2 (Higher byte)
- 35) Marker 2 (Lower byte)
- 36) Marker 3 (Higher byte)
- 37) Marker 3 (Lower byte)
- 38) Marker 4 (Higher byte)
- 39) Marker 4 (Lower byte)
- 40) Marker 5 (Higher byte)
- 41) Marker 5 (Lower byte)
- 42) Marker 6 (Higher byte)
- 43) Marker 6 (Lower byte)
- 44) Spectrum Analyzer Single Limit¹⁶ (Highest byte)
- 45) Spectrum Analyzer Single Limit
- 46) Spectrum Analyzer Single Limit
- 47) Spectrum Analyzer Single Limit (Lowest byte)
- 48) Multiple Upper Limit 1 Start X (Frequency in Hz) (Highest byte)
- 49) Multiple Upper Limit 1 Start X (Frequency in Hz)
- 50) Multiple Upper Limit 1 Start X (Frequency in Hz)
- 51) Multiple Upper Limit 1 Start X (Frequency in Hz) (Lowest byte)
- 52) Multiple Upper Limit 1 Start Y (Power Level¹⁷) (Highest byte)
- 53) Multiple Upper Limit 1 Start Y (Power Level)
- 54) Multiple Upper Limit 1 Start Y (Power Level)
- 55) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 56) Multiple Upper Limit 1 End X (Frequency in Hz) (Highest byte)
- 57) Multiple Upper Limit 1 End X (Frequency in Hz)
- 58) Multiple Upper Limit 1 End X (Frequency in Hz)

¹³ “value” sent as (value in dBm * 1,000) + 270,000

¹⁴ “value” sent as (value * 1,000)

¹⁵ Display/Data Point

To convert from “point” to frequency:

(span / (#data points-1)) * point + start frequency

where span is stored in bytes 16-19 and #data points is stored in bytes 2-3

¹⁶ “value” sent as (value in dBm * 1,000) + 270,000

¹⁷ “value” sent as (value in dBm * 1,000) + 270,000

- 59) Multiple Upper Limit 1 End X (Frequency in Hz) (Lowest byte)
 60) Multiple Upper Limit 1 End Y (Power Level) (Highest byte)
 61) Multiple Upper Limit 1 End Y (Power Level)
 62) Multiple Upper Limit 1 End Y (Power Level)
 63) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
 64-207) Multiple Upper Limits 2-5, Multiple Lower Limits 1-5 (see bytes 48-63 for format)
 208) RBW Setting¹⁸ (Highest byte)
 209) RBW Setting
 210) RBW Setting
 211) RBW Setting (Lowest byte)
 212) VBW Setting¹⁹ (Highest byte)
 213) VBW Setting
 214) VBW Setting
 215) VBW Setting (Lowest byte)
 216) OCC BW Method (00h = % of power, 01h = dB down)
 217) OCC BW % Value (0-99) (Highest byte)
 218) OCC BW % Value (0-99)
 219) OCC BW % Value (0-99)
 220) OCC BW % Value (0-99) (Lowest byte)
 221) OCC BW dBc (0-120) (Highest byte)
 222) OCC BW dBc (0-120)
 223) OCC BW dBc (0-120)
 224) OCC BW dBc (0-120) (Lowest byte)
 225) Attenuation²⁰ (Highest byte)
 226) Attenuation
 227) Attenuation
 228) Attenuation (Lowest byte)
 229) Antenna Index (0-14)
 230-245) Antenna Name (16 bytes in ASCII)
 246) AM/FM Demod Type (00h = FM-Wide Band, 01h = FM-Narrow Band, 02h = AM, 03h = SSB/CW)
 247) AM/FM Demod Volume (00h = min, FFh = max)
 248) Reference Level Offset²¹ (Highest byte)
 249) Reference Level Offset
 250) Reference Level Offset
 251) Reference Level Offset (Lowest byte)
 252) Impedance²²
 253) Impedance Loss²³ (Highest byte)
 254) Impedance Loss
 255) Impedance Loss
 256) Impedance Loss (Lowest byte)
 257) N/A
 258) N/A
 259) N/A
 260) N/A
 261) N/A
 262) N/A
 263) N/A

¹⁸ RBW: 00h = 10 kHz, 01h = 30 kHz, 02h = 100 kHz, 03h = 1 MHz

¹⁹ VBW: 00h = 100 Hz, 01h = 300 Hz, 02h = 1 kHz, 03h = 3 kHz, 04h = 10 kHz, 05h = 30 kHz, 06h = 100 kHz, 07h = 300 kHz

²⁰ Attenuation: 00h = 0 dB, 01h = 10 dB, 02h = 20 dB, 03h = 30 dB, 04h = 40 dB, 05h = 50 dB

²¹ “value” sent as (value in dBm * 1,000) + 270,000

²² Impedance Adapter: 00h = 50 Ω, 0Ah = 75 Ω, adapter 12N50-75B, 0Ch = 75 Ω, other adapter offset

²³ “value” sent as (value in dBm * 1,000) + 270,000

- 264) N/A
- 265) Status Byte 1: (0b = Off, 1b = On)
 (LSB) bit 0 : Marker 1 On/Off
 bit 1 : Marker 2 On/Off
 bit 2 : Marker 3 On/Off
 bit 3 : Marker 4 On/Off
 bit 4 : Marker 5 On/Off
 bit 5 : Marker 6 On/Off
 bits 6-7 : Not Used
- 266) Status Byte 2: (0b = Off, 1b = On)
 (LSB) bit 0 : Not Used
 bit 1 : Marker 2 Delta On/Off
 bit 2 : Marker 3 Delta On/Off
 bit 3 : Marker 4 Delta On/Off
 bits 4-7: Not Used
- 267) Status Byte 3: (0b = Off, 1b = On)
 (LSB) bit 0 : Antenna Factors Correction On/Off
 bit 1 : AM/FM Demod Status On/Off
 bit 2 : SPA Cal Status On/Off
 bit 3-4 : Logarithmic Amplitude Units (00b = dBm 01b = dBV 10b = dBmV 11b = dBuV)
 bit 5-6 : Detection Alg (00b = positive peak 10b = negative peak, 11b = RMS averaging)
 bit 7 : LCD Back Light On/Off
- 268) Status Byte 4: (0b = Off, 1b = On)
 (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
 bit 1 : Single Limit Beep On/Off
 bit 2 : Single Limit Status On/Off
 bit 3 : Single Limit Level (0b = LOWER, 1b = UPPER)
 bit 4 : Multiple Limit Upper Segment 1 Status On/Off
 bit 5 : Multiple Limit Upper Segment 1 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Upper Segment 2 Status On/Off
 bit 7 : Multiple Limit Upper Segment 2 Limit Level UPPER / LOWER
- 269) Status Byte 5: (0b = Off/LOWER limit, 1b = On/UPPER limit)
 (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
 bit 1 : Multiple Limit Upper Segment 3 Limit Level UPPER / LOWER
 bit 2 : Multiple Limit Upper Segment 4 Status On/Off
 bit 3 : Multiple Limit Upper Segment 4 Limit Level UPPER / LOWER
 bit 4 : Multiple Limit Upper Segment 5 Status On/Off
 bit 5 : Multiple Limit Upper Segment 5 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Lower Segment 1 Status On/Off
 bit 7 : Multiple Limit Lower Segment 1 Limit Type UPPER / LOWER
- 270) Status Byte 6: (0b = Off/LOWER limit, 1b = On/UPPER limit)
 (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : Multiple Limit Lower Segment 2 Limit Level UPPER / LOWER
 bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : Multiple Limit Lower Segment 3 Limit Level UPPER / LOWER
 bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : Multiple Limit Lower Segment 4 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Limit Level UPPER / LOWER
- 271) Status Byte 7: (0b = Off , 1b = On)
 (LSB) bit 0 : Serial Port Echo Status On/Off
 bit 1 : Return Sweep Time On/Off
 bit 2 : RBW Coupling (to span) (1b = auto, 0b = manual)
 bit 3 : VBW Coupling (to RBW) (1b = auto, 0b = manual)
 bit 4 : Attenuation Coupling (to ref. Level) (1b = auto, 0b = manual)

		bit 5 : Channel Power On/Off
		bit 6 : Adjacent Channel Power On/Off
		bit 7 : Occupied Bandwidth On/Off
272)	Printer Type (see control byte #30)	
273)	Trace A/B Status	
	(LSB)	bit 0-1 : Trace A (00b = A only 01b = A - B 10b = A + B) bit 2 : Trace B On/Off (0b = Off, 1b = On) bits 3-7 : Not Used
274)	Trace B Trace ID (0 = previous A data, 1-200 = saved trace id, 255 = none)	
275)	Status Byte 8:	
	(LSB)	bits 0-6 : Number of Sweeps to Average (1-25, 1 implies averaging Off) bit 7 : Reserved
276)	Status Byte 9: (0b = Off, 1b = On)	
	(LSB)	bit 0 : Preamp Hardware Installed (1b = Yes, 0b = No) ¹² bit 1 : Preamp On/Off bit 2 : N/A bit 3 : Dynamic Attenuation On/Off bit 4 : Normalization On/Off bit 5 : Bias Tee On/Off bit 6 : Scale Type (0b = Log, 1b = Linear) bit 7 : Linear Units (0b = Watts, 1b = Volts)
277)	N/A	
278)	N/A	
279)	N/A	
280)	N/A	
281)	N/A	
282)	N/A	
283)	N/A	
284)	N/A	
285)	SSB/CW BFO Adjustment Value (0-255)	
286)	N/A	
287)	N/A	
288)	N/A	
289)	N/A	
290)	RTC Battery Voltage ²⁴ (Higher byte)	
291)	RTC Battery Voltage (Lower byte)	
292)	Motherboard PCB ID ²⁵ (Higher byte)	
293)	Motherboard PCB ID (Lower byte)	
294)	N/A	
295)	N/A	
296-310)	Not Used	

²⁴ “value” sent as (value in Volts * 10)

²⁵ “value” sent as (value in Volts * 1000)

Trigger Self-Test – Control Byte #21 (15h)

Description: Triggers a self test on the Spectrum Master.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 12 bytes

- 1) Self-treport: (0b = Fail, 1b = Pass)
(LSB) bit 0 : Phase Lock Loop
bit 1 : Integrator
bit 2 : Battery
bit 3 : Temperature
bit 4 : EEPROM read/write
bit 5 : RTC Battery
bits 6- 7 : Not Used
- 2) Self-treport: (0b = Fail, 1b = Pass)
(LSB) bit 0 : Spectrum Analyzer Lock
bits 1-7 : Not Used
- 3) Battery Voltage (Higher byte)
- 4) Battery Voltage (Lower byte)
- 5) Temperature (Higher byte)
- 6) Temperature (Lower byte)
- 7) Spectrum Analyzer Lock Fail Counter (Higher byte)
- 8) Spectrum Analyzer Lock Fail Counter (Lowest byte)

Notes:

Battery Voltage in 1/10th of a Volt (e.g. 124 = 12.4 Volts)

Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) or degree Fahrenheit (e.g. 934 = 93.4 °F), depending on the current measurement unit (Metric or English) selected

Read Fail Counter – Control Byte #22 (16h)

Description: Reads the Fail Counter. Values are integer numbers of failures.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 8 bytes

- 1) Value of SA Lock Fail Counter (Higher byte)
 - 2) Value of SA Lock Fail Counter (Lower byte)
 - 3) Reserved
 - 4) Reserved
 - 5) Value of SA Fatal Error Counter (Higher byte)
 - 6) Value of SA Fatal Error Counter (Lower byte)
-

Clear Fail Counters – Control Byte #23 (17h)

Description: Resets the Lock Fail Counter and Integrator Fail Counter and spectrum analyzer Fatal Error Counter.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
-

Query Trace Names – Control Byte #24 (18h)

Description: Returns a list of all saved traces.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 3 + (41 x number of save traces) bytes

1-2) # of saved traces

For each trace:

- 1-2) Trace Index
- 3) Measurement Mode (refer to Control Byte #3)
- 4-21) Date/Time in string format (“MM/DD/YYYYHH:MM:SS”)
- 22-25) Date/Time as Unsigned Long Integer (Seconds Since January 1, 1970)
- 26-41) Trace Name (16 bytes)

255 (FFh) Operation Complete Byte

Delete Sweep Trace – Control Byte #25 (19h)

Description: Delete single trace or all stored sweep traces in Spectrum Master.

Bytes to Follow: 1 byte

- 1) 0 - Delete all traces
- X - Delete single trace #X

Spectrum Master Returns: 1 byte

1) 255 (FFh) Operation Complete Byte

OBSOLETE - Upload SPA Sweep Trace – Control Byte #26 (1Ah)

This command exists for backward compatibility with the MS2711B. To access the new features use Control Byte #36 (24h). This command cannot be used with a frequency extension module (option 6 required) attached.

Description: Uploads a spectrum analyzer sweep trace to Spectrum Master.

For data formats, refer to the footnotes listed beside the return bytes.

Bytes to Follow: 1930 bytes

- 1-2) # of following bytes (1928)
- 3) Measurement Mode²⁶
- 4-7) Time/Date (long integer format)
- 8-17) Date in String Format (mm/dd/yyyy)
- 18-25) Time in String Format (hh:mm:ss)
- 26-41) Reference Number/Trace Name (16 bytes in ASCII)
- 42-43) # of data points (400)
- 44) Start Frequency (in Hz) (Highest byte)
- 45) Start Frequency (in Hz)
- 46) Start Frequency (in Hz)
- 47) Start Frequency (in Hz) (Lowest byte)
- 48) Stop Frequency (in Hz) (Highest byte)
- 49) Stop Frequency (in Hz)
- 50) Stop Frequency (in Hz)

²⁶ See Control Byte #3 for available measurement modes.

51)	Stop Frequency (in Hz) (Lowest byte)
52)	Center Frequency (in Hz) (Highest byte)
53)	Center Frequency (in Hz)
54)	Center Frequency (in Hz)
55)	Center Frequency (in Hz) (Lowest byte)
56)	Frequency Span (in Hz) (Highest byte)
57)	Frequency Span (in Hz)
58)	Frequency Span (in Hz)
59)	Frequency Span (in Hz) (Lowest byte)
60)	Ref Level ²⁷ (Highest byte)
61)	Ref Level
62)	Ref Level
63)	Ref Level (Lowest byte)
64)	Scale per div ²⁸ (Highest byte)
65)	Scale per div
66)	Scale per div
67)	Scale per div (Lowest byte)
68)	Marker 1 ²⁹ (Higher byte)
69)	Marker 1 (Lower byte)
70)	Marker 2 (Higher byte)
71)	Marker 2 (Lower byte)
72)	Marker 3 (Higher byte)
73)	Marker 3 (Lower byte)
74)	Marker 4 (Higher byte)
75)	Marker 4 (Lower byte)
76)	Marker 5 (Higher byte)
77)	Marker 5 (Lower byte)
78)	Marker 6 (Higher byte)
79)	Marker 6 (Lower byte)
80)	Single Limit ³⁰ (Highest byte)
81)	Single Limit
82)	Single Limit
83)	Single Limit (Lowest byte)
84)	Multiple Upper Limit 1 Start X (Frequency in Hz) (Highest byte)
85)	Multiple Upper Limit 1 Start X (Frequency in Hz)
86)	Multiple Upper Limit 1 Start X (Frequency in Hz)
87)	Multiple Upper Limit 1 Start X (Frequency in Hz) (Lowest byte)
88)	Multiple Upper Limit 1 Start Y (Power Level ³¹) (Highest byte)
89)	Multiple Upper Limit 1 Start Y (Power Level)
90)	Multiple Upper Limit 1 Start Y (Power Level)
91)	Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
92)	Multiple Upper Limit 1 End X (Frequency in Hz) (Highest byte)
93)	Multiple Upper Limit 1 End X (Frequency in Hz)
94)	Multiple Upper Limit 1 End X (Frequency in Hz)
95)	Multiple Upper Limit 1 End X (Frequency in Hz) (Lowest byte)
96)	Multiple Upper Limit 1 End Y (Power Level) (Highest byte)
97)	Multiple Upper Limit 1 End Y (Power Level)

²⁷ “value” sent as (value in dBm * 1,000) +270,000

²⁸ “value” sent as (value * 1,000)

²⁹ Display/Data Point

To convert from “point” to frequency:

(span / #data points) * point + start frequency

where span is stored in bytes 69-72 and #data points is stored in bytes 55-56

³⁰ “value” sent as (value in dBm * 1,000) +270,000

³¹ “value” sent as (value in dBm * 1,000) +270,000

98)	Multiple Upper Limit 1 End Y (Power Level)
99)	Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
100-243)	Multiple Upper Limits 2-5, Multiple Lower Limits 1-5 (see bytes 80-95 for format)
244)	RBW Setting ³² (Highest byte)
245)	RBW Setting
246)	RBW Setting
247)	RBW Setting (Lowest byte)
248)	VBW Setting ³³ (Highest byte)
249)	VBW Setting
250)	VBW Setting
251)	VBW Setting (Lowest byte)
252)	OCC BW Method (00h = % of power, 01h = dB down)
253)	OCC BW % Value (0-99) (Highest byte)
254)	OCC BW % Value (0-99)
255)	OCC BW % Value (0-99)
256)	OCC BW % Value (0-99) (Lowest byte)
257)	OCC BW dBc (0-120) (Highest byte)
258)	OCC BW dBc (0-120)
259)	OCC BW dBc (0-120)
260)	OCC BW dBc (0-120) (Lowest byte)
261)	Attenuation ³⁴ (Highest byte)
262)	Attenuation
263)	Attenuation
264)	Attenuation (Lowest byte)
265-280)	Antenna Name (16 bytes in ASCII)
281)	Reference Level Offset ³⁵ (Highest byte)
282)	Reference Level Offset
283)	Reference Level Offset
284)	Reference Level Offset (Lowest byte)
285)	Impedance ³⁶
286)	Impedance Loss ³⁷ (Highest byte)
287)	Impedance Loss
288)	Impedance Loss
289)	Impedance Loss (Lowest byte)
290)	N/A
291)	N/A
292)	N/A
293)	N/A
294)	N/A
295)	N/A
296)	N/A
297)	N/A
298)	Status Byte 1: (0b = Off, 1b = On) (LSB) bit 0 : Marker 1 On/Off bit 1 : Marker 2 On/Off bit 2 : Marker 3 On/Off bit 3 : Marker 4 On/Off bit 4 : Marker 5 On/Off bit 5 : Marker 6 On/Off

³² RBW: Valid frequencies (in Hz) are 10,000, 30,000, 100,000, 1,000000

³³ VBW: Valid frequencies (in Hz) are 100, 300, 1,000, 3,000, 10,000, 30,000, 100,000, 300,000

³⁴ "value" sent as (value * 1,000)

³⁵ "value" sent as (value in dBm * 1,000) +270,000

³⁶ Impedance Adapter: 00h = 50 Ω, 0Ah = 75 Ω, adapter 12N50-75B, 0Ch = 75 Ω, other adapter offset

³⁷ "value" sent as (value * 1,000)

- bits 6-7: Not Used
- 299) Status Byte 2: (0b = Off, 1b = On)
 (LSB) bit 0 : Marker 2 Delta On/Off
 bit 1 : Marker 3 Delta On/Off
 bit 2 : Marker 4 Delta On/Off
 bits 3-7: Not Used
- 300) Status Byte 3: (0b = Off, 1b = On)
 (LSB) bit 0 : Antenna Factor Correction On/Off
 bits 1-2 : Detection Alg (00b = pos. peak 10b= neg. peak 11b = RMS averaging)
 bits 3-4 : Amplitude Units (00b = dBm 01b = dBV 10b = dBmV 11b = dBuV)
 bit 5 : Channel Power On/Off
 bit 6 : ACPR On/Off
 bit 7 : Occupied BW On/Off
- 301) Status Byte 4: (0b = Off/LOWER limit, 1b = On/UPPER limit)
 (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
 bit 1 : Single Limit On/Off
 bit 2 : Single Limit Level (0b = LOWER, 1b = UPPER)
 bit 3 : Not Used
 bit 4 : Multiple Limit Upper Segment 1 Status On/Off
 bit 5 : Multiple Limit Upper Segment 1 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Upper Segment 2 Status On/Off
 bit 7 : Multiple Limit Upper Segment 2 Limit Level UPPER / LOWER
- 302) Status Byte 5: (0b = Off/LOWER limit, 1b =On/UPPER limit)
 (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
 bit 1 : Multiple Limit Upper Segment 3 Limit Level UPPER / LOWER
 bit 2 : Multiple Limit Upper Segment 4 Status On/Off
 bit 3 : Multiple Limit Upper Segment 4 Limit Level UPPER / LOWER
 bit 4 : Multiple Limit Upper Segment 5 Status On/Off
 bit 5 : Multiple Limit Lower Segment 5 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Lower Segment 1 Status On/Off
 bit 7 : Multiple Limit Lower Segment 1 Limit Level UPPER / LOWER
- 303) Status Byte 6: (0b = Off/LOWER limit, 1b =On/UPPER limit)
 (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : Multiple Limit Lower Segment 2 Limit Level UPPER / LOWER
 bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : Multiple Limit Lower Segment 3 Limit Level UPPER / LOWER
 bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : Multiple Limit Lower Segment 4 Limit Level UPPER / LOWER
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Limit Level UPPER / LOWER
- 304) Status Byte 7:
 (LSB) bits 0-6 : Number of Sweeps to Average
 bit 7 : Reserved
- 305) Status Byte 8: (0b = Off, 1b =On)
 (LSB) bit 0 : Preamp On/Off
 bit 1 : Normalization On/Off
 bit 2 : Bias Tee On/Off
 bit 3 : Scale Type (0b = Linear, 1b = Log)
 bit 4 : Linear Units (0b = Watts, 1b = Volts)
 bits 5-7 : Date Format (000b = MMDDYYYY, 001b = DDMMYYYY, 010b = YYYYMMDD)
- 306) N/A
- 307) N/A
- 308-330) Not Used
- 331-1930) Sweep Data (400 points * 4 bytes/point = 1600 bytes)
 4 bytes for each data point

1. dBm³⁸ (Highest byte)
2. dBm
3. dBm
4. dBm (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 2) 224 (E0h) Parameter Error: Not enough bytes transferred
 - 225 (E1h) Memory Error: Not enough memory to store data
 - 238 (EEh) Time-out Error
-

Query Sweep Memory – Control Byte #27 (1Bh)

Description: Queries Spectrum Master for percentage of memory that is available for trace storage.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) % of memory currently available (0 to 100)
-

Query System Status – Control Byte #29 (1Dh)

This command is new to the MS2711D. Use it, instead of Control Byte #20, to access the new features.

Description: Queries the Spectrum Master for current system settings. Unlike Control Byte #20, this command returns only data that is valid for the active mode, plus system settings, such as the defined printer.

The current state of the Spectrum Master represents the state after the last successful remote control operation. For example, change the start frequency to another valid frequency while in remote mode, then execute control byte #29. The new start frequency will be returned in the defined bytes, even though no sweep has been performed with that frequency.

Bytes to Follow: 0 bytes

Spectrum Master Returns:

For All Modes:

- 1) Number of Following Bytes (Higher byte)
- 2) Number of Following Bytes (Lower byte)
- 3) Measurement Mode³⁹
- 4) Printer Type⁴⁰
- 5) Current Language
(00h = English, 01h = French, 02h = German, 03h = Spanish, 04h = Chinese, 05h = Japanese)
- 6) LCD Contrast Value (0-255)
- 7) Date Format (00h = MM/DD/YYYY, 01h = DD/MM/YYYY, 02h = YYYY/MM/DD)
- 8) RTC battery⁴¹ (Higher byte)
- 9) RTC battery (Lower byte)
- 10) PC Board Revision⁴² (Higher byte)
- 11) PC Board Revision (Lower byte)
- 12-13) Digital Mother Board ID. Beginning with motherboard 64968, the hardware includes a 9-bit

³⁸ Value sent as (Value in dBm * 1000) + 270,000

³⁹ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

⁴⁰ See Control Byte #30 for supported printers.

⁴¹ Value sent as Volts * 10. For example, 2.7 V = 27.

⁴² This value is for internal use only.

digital ID port. The digital ID will be used together with the PC Board Revision (mother board ID voltage) to identify the board and “dash” number. For boards prior to 64968, bytes 12 and 13 will be 0
14-25) Not Used

For Spectrum Analyzer Mode/Transmission Mode (Option 21):

- 26) Spectrum Analyzer Mode Data Points (Higher byte)
- 27) Spectrum Analyzer Mode Data Points (Lower byte)
- 28) Spectrum Analyzer Start Frequency⁴³ (Highest byte)
- 29) Spectrum Analyzer Start Frequency
- 30) Spectrum Analyzer Start Frequency
- 31) Spectrum Analyzer Start Frequency (Lowest byte)
- 32) Spectrum Analyzer Stop Frequency⁴⁴ (Highest byte)
- 33) Spectrum Analyzer Stop Frequency
- 34) Spectrum Analyzer Stop Frequency
- 35) Spectrum Analyzer Stop Frequency (Lowest byte)
- 36) Spectrum Analyzer Center Frequency⁴⁵ (Highest byte)
- 37) Spectrum Analyzer Center Frequency
- 38) Spectrum Analyzer Center Frequency
- 39) Spectrum Analyzer Center Frequency (Lowest byte)
- 40) Spectrum Analyzer Frequency Span⁴⁶ (Highest byte)
- 41) Spectrum Analyzer Frequency Span
- 42) Spectrum Analyzer Frequency Span
- 43) Spectrum Analyzer Frequency Span (Lowest byte)
- 44) Spectrum Analyzer Minimum Frequency Step Size (Highest byte)
- 45) Spectrum Analyzer Minimum Frequency Step Size
- 46) Spectrum Analyzer Minimum Frequency Step Size
- 47) Spectrum Analyzer Minimum Frequency Step Size (Lowest byte)
- 48) Ref Level (Highest byte)⁴⁷
- 49) Ref Level
- 50) Ref Level
- 51) Ref Level (Lowest byte)
- 52) Scale per div (Highest byte)⁴⁸
- 53) Scale per div
- 54) Scale per div
- 55) Scale per div (Lowest byte)
- 56) Spectrum Analyzer Frequency Marker 1 (Higher byte)⁴⁹
- 57) Spectrum Analyzer Frequency Marker 1 (Lower byte)
- 58) Spectrum Analyzer Frequency Marker 2 (Higher byte)
- 59) Spectrum Analyzer Frequency Marker 2 (Lower byte)
- 60) Spectrum Analyzer Frequency Marker 3 (Higher byte)
- 61) Spectrum Analyzer Frequency Marker 3 (Lower byte)
- 62) Spectrum Analyzer Frequency Marker 4 (Higher byte)
- 63) Spectrum Analyzer Frequency Marker 4 (Lower byte)
- 64) Spectrum Analyzer Frequency Marker 5 (Higher byte)
- 65) Spectrum Analyzer Frequency Marker 5 (Lower byte)
- 66) Spectrum Analyzer Frequency Marker 6 (Higher byte)
- 67) Spectrum Analyzer Frequency Marker 6 (Lower byte)

⁴³ Scaled by Frequency Scale Factor (bytes 321-322)

⁴⁴ Scaled by Frequency Scale Factor (bytes 321-322)

⁴⁵ Scaled by Frequency Scale Factor (bytes 321-322)

⁴⁶ Scaled by Frequency Scale Factor (bytes 321-322)

⁴⁷ Value sent as (value in dBm * 1000) + 270,000

⁴⁸ Value sent as (value * 1000)

⁴⁹ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

- 68) Spectrum Analyzer Single Limit (Highest byte)⁵⁰
 69) Spectrum Analyzer Single Limit
 70) Spectrum Analyzer Single Limit
 71) Spectrum Analyzer Single Limit (Lowest byte)
 72) SPA Multiple Upper Limit 1 Start X⁵¹ (Highest byte)
 73) SPA Multiple Upper Limit 1 Start X
 74) SPA Multiple Upper Limit 1 Start X
 75) SPA Multiple Upper Limit 1 Start X (Lowest byte)
 76) SPA Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)⁵²
 77) SPA Multiple Upper Limit 1 Start Y (Power Level)
 78) SPA Multiple Upper Limit 1 Start Y (Power Level)
 79) SPA Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
 80) SPA Multiple Upper Limit 1 End X⁵³ (Highest byte)
 81) SPA Multiple Upper Limit 1 End X
 82) SPA Multiple Upper Limit 1 End X
 83) SPA Multiple Upper Limit 1 End X (Lowest byte)
 84) SPA Multiple Upper Limit 1 End Y (Power Level) (Highest byte)⁵⁴
 85) SPA Multiple Upper Limit 1 End Y (Power Level)
 86) SPA Multiple Upper Limit 1 End Y (Power Level)
 87) SPA Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
 88-231) SPA Multiple Upper Limits 2-5, SA Multiple Lower Limits 1-5 (see bytes 72-87 for format)
 232) RBW Setting (Highest byte)⁵⁵
 233) RBW Setting
 234) RBW Setting
 235) RBW Setting (Lowest byte)
 236) VBW Setting (Highest byte)⁵⁶
 237) VBW Setting
 238) VBW Setting
 239) VBW Setting (Lowest byte)
 240) OCC BW Method⁵⁷
 241) OCC BW % Value (Highest byte)⁵⁸
 242) OCC BW % Value
 243) OCC BW % Value
 244) OCC BW % Value (Lowest byte)
 245) OCC BW dBc (Highest byte)⁵⁹
 246) OCC BW dBc
 247) OCC BW dBc
 248) OCC BW dBc (Lowest byte)
 249) Attenuation (Highest byte)
 250) Attenuation
 251) Attenuation
 252) Attenuation (Lowest byte)
 253) Antenna Index(0-14)
 254-269) Antenna Name (16 bytes in ASCII)
 270) Status Byte 1: (0b = Off , 1b = On)

⁵⁰ Value sent as (value in dBm * 1000) + 270000

⁵¹ Scaled by Frequency Scale Factor (bytes 321-322)

⁵² Value sent as (value in dBm * 1000) + 270000

⁵³ Scaled by Frequency Scale Factor (bytes 321-322)

⁵⁴ Value sent as (value in dBm * 1000) + 270000

⁵⁵ RBW frequency sent in Hz.

⁵⁶ VBW frequency sent in Hz.

⁵⁷ 00h = % of power, 01h = dB down

⁵⁸ 0 – 99%

⁵⁹ 0 – 120 dBc

- (LSB) bit 0 : Spectrum Analyzer Mode Marker 1 On/Off
bit 1 : Spectrum Analyzer Mode Marker 2 On/Off
bit 2 : Spectrum Analyzer Mode Marker 3 On/Off
bit 3 : Spectrum Analyzer Mode Marker 4 On/Off
bit 4 : Spectrum Analyzer Mode Marker 5 On/Off
bit 5 : Spectrum Analyzer Mode Marker 6 On/Off
bits 6 - 7 : Not Used
- 271) Status Byte 2: (0b = Off, 1b = On)
(LSB) bit 0 : S21 Spa Cal Status (0 – Cal OFF, 1 – Cal ON)
bit 1 : Spectrum Analyzer Mode Marker 2 Delta On/Off
bit 2 : Spectrum Analyzer Mode Marker 3 Delta On/Off
bit 3 : Spectrum Analyzer Mode Marker 4 Delta On/Off
bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
bit 5 : Pre Amp Status On/Off
bit 6 : Dynamic Attenuation On/Off
bit 7 : Normalization On/Off
- 272) Status Byte 3: (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
(LSB) bit 0 : SPA Limit Type (0b = Single, 1b = Multiple)
bit 1 : SPA Single Limit Beep On/Off
bit 2 : SPA Single Limit Status On/Off
bit 3 : SPA Single Limit Beep Level ABOVE/BELOW
bit 4 : SPA Multiple Limit Upper Segment 1 Status On/Off
bit 5 : SPA Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW⁶⁰
bit 6 : SPA Multiple Limit Upper Segment 2 Status On/Off
bit 7 : SPA Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW
- 273) Status Byte 4 : (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
(LSB) bit 0 : SPA Multiple Limit Upper Segment 3 Status On/Off
bit 1 : SPA Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
bit 2 : SPA Multiple Limit Upper Segment 4 Status On/Off
bit 3 : SPA Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
bit 4 : SPA Multiple Limit Upper Segment 5 Status On/Off
bit 5 : SPA Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
bit 6 : SPA Multiple Limit Lower Segment 1 Status On/Off
bit 7 : SPA Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW⁶¹
- 274) Status Byte 5 : (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
(LSB) bit 0 : SPA Multiple Limit Lower Segment 2 Status On/Off
bit 1 : SPA Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
bit 2 : SPA Multiple Limit Lower Segment 3 Status On/Off
bit 3 : SPA Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
bit 4 : SPA Multiple Limit Lower Segment 4 Status On/Off
bit 5 : SPA Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
bit 6 : SPA Multiple Limit Lower Segment 5 Status On/Off
bit 7 : SPA Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
- 275) Status Byte 6: (0b = Off, 1b = On)
(LSB) bit 0 : Antenna Factors Correction On/Off
bit 1 : Bias Tee On/Off (Option 10)
bit 2 : SPA Cal Status On/Off
bits 3-4 : Amplitude Units (Log) - 00b = dBm 01b = dBV 10b = dBmV 11b = dBuV
(Linear) – 00b = Watts 01b = Volts
bits 5-6 : Detection Alg (00b = pos. peak 01b = RMS Averaging 10b = neg. peak 11b =

⁶⁰ Beep level is always 1b for upper segmented limit line

⁶¹ Beep level is always 0b for lower segmented limit line

		Sampling Mode)
		bit 7 : Units Type (0b = Log 1b = Linear)
276)	Status Byte 7: (0b = Off, 1b = On)	
	(LSB)	bit 0: Serial Port Echo Status On/Off
		bit 1: Return Sweep Time On/Off
		bit 2: RBW Coupling (1b = Auto, 0b = Manual)
		bit 3: VBW Coupling (1b = Auto, 0b = Manual)
		bit 4: Attenuation Coupling (1b = Auto, 0b = Manual)
		bit 5: Channel Power On/Off
		bit 6: Adjacent Channel Power On/Off
		bit 7: Occupied BW Measurement On/Off
277)	Reference Level Offset ⁶² (Highest byte)	
278)	Reference Level Offset	
279)	Reference Level Offset	
280)	Reference Level Offset (Lowest byte)	
281)	External Reference Frequency ⁶³	
282)	Signal Standard ⁶⁴ (Higher byte)	
283)	Signal Standard (Lower byte)	
284)	Channel Selection ⁶⁵ (Higher byte)	
285)	Channel Selection (Lower byte)	
286)	Trigger Type ⁶⁶	
287)	Interference Analysis Frequency ⁶⁷ (Highest byte)	
288)	Interference Analysis Frequency	
289)	Interference Analysis Frequency	
290)	Interference Analysis Frequency (Lowest byte)	
291)	Trigger Position (0 – 100%)	
292)	Min Sweep Time (in μ s) (Highest byte)	
293)	Min Sweep Time (in μ s)	
294)	Min Sweep Time (in μ s)	
295)	Min Sweep Time (in μ s) (Lowest byte)	
296)	Video Trigger Level ⁶⁸ (Highest byte)	
297)	Video Trigger Level	
298)	Video Trigger Level	
299)	Video Trigger Level (Lowest byte)	
300)	Status Byte 8	
	(LSB)	bit 0: Input Power Status (1b = Input Power Too High, 0b = Input Power Ok)
		bit 1: Limit Beep Output (Option 86 Only) (00h = Internal Speaker, 01h = External)
		bits 2-7: Not Used
301)	Status Byte 9	
	(LSB)	bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF)
		bit 7: Not Used
302)	Status Byte 10: (0b = Off, 1b = On)	
	(LSB)	bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B)
		bit 2: Max Hold On/Off
		bit 3: Min Hold On/Off
		bits 4-7: Not Used
303)	Impedance (00h = 50Ω, 0Ah = 75Ω Anritsu Adapter, 0Ch = 75Ω Other Adapter)	
304)	Impedance Loss ⁶⁹ (Higher byte)	

⁶² Value sent as (value in dBm * 1000) + 270,000

⁶³ 1 byte in MHz (i.e. 20 = 20MHz)

⁶⁴ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

⁶⁵ “No Channel” is sent as FFFEh

⁶⁶ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

⁶⁷ Scaled by Frequency Scale Factor (bytes 321-322)

⁶⁸ Value sent as (value in dBm * 1000) + 270,000

- 305) Impedance Loss (Lower byte)
- 306) AM/FM Demod Type⁷⁰
- 307) AM/FM Demod Status (01h = On, 00h = Off)
- 308) AM/FM Demod Volume (0 to 100)
- 309) AM/FM Demod Frequency⁷¹ (Highest byte)
- 310) AM/FM Demod Frequency
- 311) AM/FM Demod Frequency
- 312) AM/FM Demod Frequency (Lowest byte)
- 313) AM/FM Demod Time (in ms) (Highest byte)
- 314) AM/FM Demod Time (in ms)
- 315) AM/FM Demod Time (in ms)
- 316) AM/FM Demod Time (in ms) (Lowest byte)
- 317) SSB BFO Offset⁷² (Highest byte)
- 318) SSB BFO Offset
- 319) SSB BFO Offset
- 320) SSB BFO Offset (Lowest byte)
- 321) Frequency Scale Factor⁷³ (Higher byte)
- 322) Frequency Scale Factor (Lower byte)
- 323) Frequency Range Minimum⁷⁴ (Highest byte)
- 324) Frequency Range Minimum
- 325) Frequency Range Minimum
- 326) Frequency Range Minimum (Lowest byte)
- 327) Frequency Range Maximum⁷⁵ (Highest byte)
- 328) Frequency Range Maximum
- 329) Frequency Range Maximum
- 330) Frequency Range Maximum (Lowest byte)
- 331) Marker Type⁷⁶
- 332-355) Signal Standard Name, 24bytes of ASCII
- 356-400) Not Used

For Power Meter Mode (Option 29 Only):

- 26) Power Meter Start Freq⁷⁷ (Highest byte)
- 27) Power Meter Start Freq
- 28) Power Meter Start Freq
- 29) Power Meter Start Freq (Lowest byte)
- 30) Power Meter Stop Freq⁷⁸ (Highest byte)
- 31) Power Meter Stop Freq
- 32) Power Meter Stop Freq
- 33) Power Meter Stop Freq (Lowest byte)
- 34) Power Meter Center Freq⁷⁹ (Highest byte)
- 35) Power Meter Center Freq
- 36) Power Meter Center Freq
- 37) Power Meter Center Freq (Lowest byte)

⁶⁹ Value sent as (value in dB * 1000), valid values are 0 to 20 dB

⁷⁰ AM/FM Demod Type: 00h = FM-Wide Band, 01h = FM-Narrow Band, 02h = AM, 03h = SSB Lower, 04h = SSB Upper

⁷¹ Scaled by Frequency Scale Factor (bytes 321-322)

⁷² Value sent as ((value in Hz) – 10,000)

⁷³ In number of Hz

⁷⁴ Scaled by Frequency Scale Factor (bytes 321-322)

⁷⁵ Scaled by Frequency Scale Factor (bytes 321-322)

⁷⁶ 00h = Regular Marker, 01h = Noise Marker

⁷⁷ Scaled by Frequency Scale Factor (bytes 59-60)

⁷⁸ Scaled by Frequency Scale Factor (bytes 59-60)

⁷⁹ Scaled by Frequency Scale Factor (bytes 59-60)

- 38) Power Meter Span⁸⁰ (Highest byte)
- 39) Power Meter Span
- 40) Power Meter Span
- 41) Power Meter Span (Lowest byte)
- 42) Signal Standard⁸¹ (Higher byte)
- 43) Signal Standard (Lower byte)
- 44) Channel Selection⁸² (Higher byte)
- 45) Channel Selection (Lower byte)
- 46) Power Meter Offset (Highest byte)
- 47) Power Meter Offset
- 48) Power Meter Offset
- 49) Power Meter Offset (Lowest byte)
- 50) Power Meter Relative (Highest byte)⁸³
- 51) Power Meter Relative
- 52) Power Meter Relative
- 53) Power Meter Relative (Lowest byte)
- 54) Power Meter Status (00h = Off, 01h = On)
- 55) Power Meter Unit (00h = Watts, 01h = dBm)
- 56) Power Meter Relative Status (00h = Off, 01h = On)
- 57) Power Meter Offset Status (00h = Off, 01h = On)
- 58) Power Meter RMS Averaging Level (00h = Off, 01h = Low, 02h = Medium, 03h = High)
- 59) Frequency Scale Factor⁸⁴ (Higher byte)
- 60) Frequency Scale Factor (Lower byte)
- 61) Frequency Range Minimum⁸⁵ (Highest byte)
- 62) Frequency Range Minimum
- 63) Frequency Range Minimum
- 64) Frequency Range Minimum (Lowest byte)
- 65) Frequency Range Maximum⁸⁶ (Highest byte)
- 66) Frequency Range Maximum
- 67) Frequency Range Maximum
- 68) Frequency Range Maximum (Lowest byte)
- 69) Zero Status (00h = Off, 01h = On)
- 70) Zero Value⁸⁷ (Highest byte)
- 71) Zero Value
- 72) Zero Value
- 73) Zero Value (Lowest byte)
- 74-97) Signal Standard Name, 24 bytes of ASCII
- 98-120) Not Used

For T1 Mode (Option 50):

- 26) T1 Receive Input (00h: Terminate, 01h: Bridged, 02h: Monitor)
- 27) T1 Framing Mode (01h: ESF, 02h: D4SF)
- 28) T1 Line Coding (01h: B8ZS, 02h: AMI)
- 29) T1 Clock Source (00h: External, 01h: Internal)
- 30) T1 Tx Level (01h: 0 dB, 02h: -7.5 dB, 03h: -15 dB)
- 31) T1 Error Insert Type (00h: Frame Error, 01h: BPV, 02h: Bit Errors, 04h: RAI, 05h: AIS)
- 32) T1 Loop Code (00h: CSU, 01h: NIU, 02h: User 1, 03h: User 2)

⁸⁰ Scaled by Frequency Scale Factor (bytes 59-60)

⁸¹ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

⁸² “No Channel” is sent as FFFEh

⁸³ Value as ((value in dBm * 1000) + 100)

⁸⁴ In number of Hz

⁸⁵ Scaled by Frequency Scale Factor

⁸⁶ Scaled by Frequency Scale Factor

⁸⁷ Value sent as ((value in dBm * 1000) + 100)

- 33) T1 CRC Method (00h: ANSI CRC, 01h: Japanese CRC)
- 34) T1 Loop Type (00h: In Band, 01h: Data Link)
- 35) T1 Pattern (Higher byte)
- 36) T1 Pattern (Lower byte) 01h: PRBS-9, 02h: PRBS-11, 03h: PRBS-15, 04h: PRBS-20(O.151), 05h: PRBS-20(O.153), 06h: PRBS-23, 07h: QRSS, 08h: 1 in 8, 09h: 2 in 8, 0Ah: 3 in 8, 0Bh: All Ones, 0Ch: All Zeros, 0Dh: T1-DALY, 0Eh: User Defined)
- 37) T1 Pattern Invert Status (00h: Non-Inverted, 01h: Inverted)
- 38) T1 Display Type (00h: Histogram, 01h: Raw Data)
- 39) T1 Impedance
- 40 - 55) First User Defined Loop Code Down (16 bytes)
- 56 - 71) Second User Defined Loop Code Down (16 bytes)
- 72 - 87) First User Defined Loop Code Up (16 bytes)
- 88 - 103) Second User Defined Loop Code Up (16 bytes)
- 104 - 135) User Defined Pattern (32 bytes)
- 136) T1 1st User Defined Loop Up (Higher byte)
- 137) T1 1st User Defined Loop Up (Lower byte)
- 138) T1 2nd User Defined Loop Up (Higher byte)
- 139) T1 2nd User Defined Loop Up (Lower byte)
- 140) T1 1st User Defined Loop Down (Higher byte)
- 141) T1 1st User Defined Loop Down (Lower byte)
- 142) T1 2nd User Defined Loop Down (Higher byte)
- 143) T1 2nd User Defined Loop Down (Lower byte)
- 144) T1 User Defined Pattern (Highest byte)
- 145) T1 User Defined Pattern
- 146) T1 User Defined Pattern
- 147) T1 User Defined Pattern (Lowest byte)
- 148) T1 Bit Error Insert Value (1-1000) (Higher byte)
- 149) T1 Bit Error Insert Value (Lower byte)
- 150) T1 Frame Error Insert Value (1-1000) (Higher byte)
- 151) T1 Frame Error Insert Value (Lower byte)
- 152) T1 BPV Error Insert Value (1-1000) (Higher byte)
- 153) T1 BPV Error Insert Value (Lower byte)
- 154) T1 Graph Resolution⁸⁸
- 155) T1 Measurement Duration⁸⁹
- 156) T1 Voltage Measurement Scale (00h = Vpp, 01h = dBdsx)
- 157) T1 Auto Framing Mode (00h = fixed framing, >00h = auto framing mode)
- 158 – 240) Not Used

For E1 Mode (Option 50):

- 26) E1 Receive Input (00h: Terminate, 01h: Bridged, 02h: Monitor)
- 27) E1 Framing Mode (03h: PCM30, 04h: PCM30CRC, 05h: PCM31, 06h: PCM31CRC)
- 28) E1 Line Coding (02h: AMI, 03h: HDB3)
- 29) E1 Clock Source (00h: External, 01h: Internal)
- 30) E1 Tx Level
- 31) E1 Error Insert Type (00h: Frame Error, 01h: BPV, 02h: Bit Errors, 04h: RAI, 05h: AIS)
- 32) E1 Loop Code
- 33) E1 CRC Method
- 34) E1 Loop Type
- 35) E1 Pattern (Highest byte)
- 36) E1 Pattern (Lowest byte) (01h: PRBS-9, 02h: PRBS-11, 03h: PRBS-15, 04h: PRBS-20(O.151), 05h: PRBS-20(O.153), 06h: PRBS-23, 07h: QRSS, 08h: 1 in 8, 09h: 2 in 8, 0Ah: 3 in 8, 0Bh: All Ones,

⁸⁸ Graph Resolution: 00h: Auto, 01h: 1 sec, 02h: 15 sec, 03h: 30 sec, 04h: 45 sec, 05h: 1 min, 06h: 15 min, 07h: 30 min, 08h: 45 min, 09h: 60 min

⁸⁹ Measurement Duration: 00h: Manual, 01h: 3 min, 02h: 15 min, 03h: 30 min, 04h: 1 hr, 05h: 3 hrs, 06h: 6 hrs, 07h: 12 hrs, 08h: 1 day, 09h: 2 days

- 0Ch: All Zeros, 0Dh: T1-DALY, 0Eh: User Defined)
- 37) E1 Pattern Invert (00h: Non-Inverted, 01h: Inverted)
 - 38) E1 Display Type (00h: Histogram, 01h: Raw Data)
 - 39) E1 Impedance (01h: 75 Ω, 02h: 120 Ω)
 - 40 - 55) First User Defined Loop Code Down (16 bytes)
 - 56 - 71) Second User Defined Loop Code Down (16 bytes)
 - 72 - 87) First User Defined Loop Code Up (16 bytes)
 - 88 - 103) Second User Defined Loop Code Up (16 bytes)
 - 104 - 135) User Defined Pattern (32 bytes)
 - 136) E1 1st User Defined Loop Up (Highest byte)
 - 137) E1 1st User Defined Loop Up (Lowest byte)
 - 138) E1 2nd User Defined Loop Up (Highest byte)
 - 139) E1 2nd User Defined Loop Up (Lowest byte)
 - 140) E1 1st User Defined Loop Down (Highest byte)
 - 141) E1 1st User Defined Loop Down (Lowest byte)
 - 142) E1 2nd User Defined Loop Down (Highest byte)
 - 143) E1 2nd User Defined Loop Down (Lowest byte)
 - 144) E1 User Defined Pattern (Highest byte)
 - 145) E1 User Defined Pattern
 - 146) E1 User Defined Pattern
 - 147) E1 User Defined Pattern (Lowest byte)
 - 148) E1 Bit Error Insert Value (1-1000) (Higher byte)
 - 149) E1 Bit Error Insert Value (Lower byte)
 - 150) E1 Frame Error Insert Value (1-1000) (Higher byte)
 - 151) E1 Frame Error Insert Value (Lower byte)
 - 152) E1 BPV Error Insert Value (1-1000) (Higher byte)
 - 153) E1 BPV Error Insert Value (Lower byte)
 - 154) E1 Graph Resolution⁹⁰
 - 155) E1 Measurement Duration⁹¹
 - 156) E1 Voltage Measurement Scale (00h = Vpp, 01h = dBdsx)
 - 157-240) Not Used
-

⁹⁰ Graph Resolution: 00h: Auto, 01h: 1 sec, 02h: 15 sec, 03h: 30 sec, 04h 45 sec, 05h 1 min, 06h: 15 min, 07h: 30 min, 08h: 45 min, 09h: 60 min

⁹¹ Measurement Duration: 00h: Manual, 01h: 3 min, 02h: 15 min, 03h: 30 min, 04h: 1 hr, 05h: 3 hrs, 06h: 6 hrs, 07h: 12 hrs, 08h: 1 day, 09h: 2 days

Select Printer Type – Control Byte #30 (1Eh)

Description: Select Printer Type.

Bytes to Follow: 1 byte

1) Printer ID

- 0 – Epson Stylus Models
- 1 – Epson LQ Models
- 2 – Citizen PN Models
- 3 – NEC Superscript Models
- 4 – NEC Silentwriter Models
- 5 – Seiko DPU 411, 414 Models
- 6 – Canon BJC 50
- 7 – Canon BJC 80
- 8 – Canon BJC 250
- 9 – Canon BJC 4400
- 10 – HP DJ 300 Series
- 11 – HP DJ 400 Series
- 12 – HP DJ 500 Series
- 13 – HP DJ 600 Series
- 14 – HP DJ 800 Series
- 15 – HP DJ 1120
- 16 – HP LJ 6L, 6P, 4000
- 17 – Epson Esc/P Compatible
- 18 – Epson Esc/P2 Compatible
- 19 – Epson Esc/P Raster Compatible
- 20 – HP PCL3 Compatible

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
-

Recall Sweep Trace – Control Byte #33 (21h)

This command is new to the MS2711D. Use it, instead of Control Byte #17, to access the new features.

Description: Queries the Spectrum Master for sweep trace data.

NOTE: Before you can recall a sweep stored in non-volatile memory (trace numbers 1-200) you must build a trace table in the Spectrum Master's RAM. Use Control Byte #24 to build the trace table. Since the trace table exists in RAM, Control Byte #24 must be executed every time the Spectrum Master's power is cycled.

Bytes to Follow: 1 byte

- 0 = Last sweep trace before entering remote mode (sweep trace in RAM)
- 1-200 = Specific saved sweep number (stored sweeps in Flash memory)

Spectrum Master Returns:

- 1-2) # of following bytes (total length - 2)
- 3) Current Instrument Date Format⁹²
- 4) Not Used
- 5-11) Model Number (7 bytes in ASCII)
- 12-15) Software Version (4 bytes ASCII)
- 16) Measurement Mode⁹³
- 17-20) Time/Date (in Long Integer)⁹⁴)

⁹² 00h = MM/DD/YYYY, 01h = DD/MM/YYYY, 02h = YYYY/MM/DD

⁹³ Refer to Control Byte #3 “Select Measurement Mode” for detailed value.

⁹⁴ Time/Date long integer representation is in seconds since January 1, 1970

- 21-30) Date in String Format (mm/dd/yyyy)
- 31-38) Time in String Format (hh:mm:ss)
- 39-54) Reference number stamp (16 bytes in ASCII)
- 55-56) # data points (130, 259 or 517 or 401 or 100)

For Spectrum Analyzer Mode:

- 57) Start Frequency⁹⁵ (Highest byte)
- 58) Start Frequency
- 59) Start Frequency
- 60) Start Frequency (Lowest byte)
- 61) Stop Frequency⁹⁶ (Highest byte)
- 62) Stop Frequency
- 63) Stop Frequency
- 64) Stop Frequency (Lowest byte)
- 65) Center Frequency⁹⁷ (Highest byte)
- 66) Center Frequency
- 67) Center Frequency
- 68) Center Frequency (Lowest byte)
- 69) Frequency Span⁹⁸ (Highest byte)
- 70) Frequency Span
- 71) Frequency Span
- 72) Frequency Span (Lowest byte)
- 73) Minimum Frequency Step Size (Highest byte)
- 74) Minimum Frequency Step Size
- 75) Minimum Frequency Step Size
- 76) Minimum Frequency Step Size (Lowest byte)
- 77) Ref Level⁹⁹ (Highest byte)
- 78) Ref Level
- 79) Ref Level
- 80) Ref Level (Lowest byte)
- 81) Scale per div¹⁰⁰ (Highest byte)
- 82) Scale per div
- 83) Scale per div
- 84) Scale per div (Lowest byte)
- 85) Frequency Marker 1¹⁰¹ (Higher byte)
- 86) Frequency Marker 1 (Lower byte)
- 87) Frequency Marker 2 (Higher byte)
- 88) Frequency Marker 2 (Lower byte)
- 89) Frequency Marker 3 (Higher byte)
- 90) Frequency Marker 3 (Lower byte)
- 91) Frequency Marker 4 (Higher byte)
- 92) Frequency Marker 4 (Lower byte)
- 93) Frequency Marker 5 (Higher byte)
- 94) Frequency Marker 5 (Lower byte)
- 95) Frequency Marker 6 (Higher byte)
- 96) Frequency Marker 6 (Lower byte)
- 97) Single Limit¹⁰² (Highest byte)

⁹⁵ Scaled by Frequency Scale Factor (bytes 335-336)

⁹⁶ Scaled by Frequency Scale Factor (bytes 335-336)

⁹⁷ Scaled by Frequency Scale Factor (bytes 335-336)

⁹⁸ Scaled by Frequency Scale Factor (bytes 335-336)

⁹⁹ Value sent as (Value in dBm * 1000) + 270,000

¹⁰⁰ Value sent as (Value * 1000)

¹⁰¹ Value sent as data point on display. Freq = (Point * Span / (Total Data Points - 1)) + Start Freq

¹⁰² Value sent as (Value in dBm * 1000) + 270,000

- 98) Single Limit
 99) Single Limit
 100) Single Limit (Lowest byte)
 101) Multiple Upper Limit 1 Start X¹⁰³ (Highest byte)
 102) Multiple Upper Limit 1 Start X
 103) Multiple Upper Limit 1 Start X
 104) Multiple Upper Limit 1 Start X (Lowest byte)
 105) Multiple Upper Limit 1 Start Y (Power Level¹⁰⁴) (Highest byte)
 106) Multiple Upper Limit 1 Start Y (Power Level)
 107) Multiple Upper Limit 1 Start Y (Power Level)
 108) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
 109) Multiple Upper Limit 1 End X¹⁰⁵ (Highest byte)
 110) Multiple Upper Limit 1 End X
 111) Multiple Upper Limit 1 End X
 112) Multiple Upper Limit 1 End X (Lowest byte)
 113) Multiple Upper Limit 1 End Y (Power Level) (Highest byte)
 114) Multiple Upper Limit 1 End Y (Power Level)
 115) Multiple Upper Limit 1 End Y (Power Level)
 116) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
 117-260) Multiple Upper Limits 2-5, Multiple Lower Limits 1-5 (see bytes 101-116 for format)
 261) RBW Setting (Frequency in Hz) (Highest byte)
 262) RBW Setting (Frequency in Hz)
 263) RBW Setting (Frequency in Hz)
 264) RBW Setting (Frequency in Hz) (Lowest byte)
 265) VBW Setting (Frequency in Hz) (Highest byte)
 266) VBW Setting (Frequency in Hz)
 267) VBW Setting (Frequency in Hz)
 268) VBW Setting (Frequency in Hz) (Lowest byte)
 269) OCC BW Method (0b = % of power, 1b = dB down)
 270) OCC BW % Value¹⁰⁶
 271) OCC BW dBc¹⁰⁷
 272) Attenuation¹⁰⁸ (Highest byte)
 273) Attenuation
 274) Attenuation
 275) Attenuation (Lowest byte)
 276-291) Antenna Name(16 bytes in ASCII)
 292) Status Byte 1: (0b = Off , 1b = On)
 (LSB) bit 0 : Marker 1 On/Off
 bit 1 : Marker 2 On/Off
 bit 2 : Marker 3 On/Off
 bit 3 : Marker 4 On/Off
 bit 4 : Marker 5 On/Off
 bit 5 : Marker 6 On/Off
 bits 6-7: Not Used
 293) Status Byte 2: (0b = Off , 1b = On)
 (LSB) bit 0 : Not Used
 bit 1 : Marker 2 Delta On/Off
 bit 2 : Marker 3 Delta On/Off
 bit 3 : Marker 4 Delta On/Off

¹⁰³ Scaled by Frequency Scale Factor (bytes 335-336)

¹⁰⁴ Value sent as (value in dBm * 1000) + 270,000

¹⁰⁵ Scaled by Frequency Scale Factor (bytes 335-336)

¹⁰⁶ % value is 0-99

¹⁰⁷ dBc value 0 – 120 dBc

¹⁰⁸ Value sent as (value in dB * 1000)

- bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
 bit 5 : Pre Amp Status On/Off
 bit 6 : Dynamic Attenuation On/Off
 bit 7 : Normalization On/Off
- 294) Status Byte 3: (0b = Off, 1b = On)
- (LSB) bit 0 : Antenna Factor Correction On/Off
 bits 1-2 : Detection alg (00b = pos. peak 01b = RMS average 10b = neg. peak 11b = sampling mode)
 bits 3-4 : Amplitude Units (Log) - (00b = dBm 01b = dBV 10b = dBmV 11b = dBuV)
 (Linear) – (00b = Watts 01b = Volts)
 bit 5 : Channel Power On/Off
 bit 6 : Adjacent Channel Power On/Off
 bit 7 : Units Type (0b = Log 1b = Linear)
- 295) Status Byte 4¹⁰⁹
 (0b = Off/Beep if data is BELOW line, 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
 bit 1 : Not Used
 bit 2 : Single Limit On/Off
 bit 3 : Single Limit Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Upper Segment 1 Status On/Off
 bit 5 : Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW¹¹⁰
 bit 6 : Multiple Limit Upper Segment 2 Status On/Off
 bit 7 : Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW
- 296) Status Byte 5
 (0b = Off/Beep if data is below line, 1b = On/Beep if data is above line)
 (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
 bit 1 : Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
 bit 2 : Multiple Limit Upper Segment 4 Status On/Off
 bit 3 : Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Upper Segment 5 Status On/Off
 bit 5 : Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
 bit 6 : Multiple Limit Lower Segment 1 Status On/Off
 bit 7 : Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW¹¹¹
- 297) Status Byte 6
 (0b = Off/Beep if data is BELOW line, 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
- 298) Status Byte 7
 (LSB) bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF)
 bit 7: Not Used
- 299) Reference Level Offset¹¹²(Highest byte)
 300) Reference Level Offset
 301) Reference Level Offset
 302) Reference Level Offset (Lowest byte)
 303) External Reference Frequency¹¹³

¹⁰⁹ For bits 2, 1 and 0 (“X” is “don’t care): 0X0=no limit, 1X0=single limit, 0X1=multiple limit, 1X1=multiple limit.

¹¹⁰ Upper limits always trigger an error beep if data is ABOVE the limit segment, for example, this bit is always 1b.

¹¹¹ Lower limits always trigger an error beep if data is BELOW the limit segment, for example, this bit is always 0b.

¹¹² Value sent as (value in dBm * 1000) + 270,000

- 304) Signal Standard¹¹⁴ (Higher byte)
- 305) Signal Standard (Lower byte)
- 306) Channel Selection¹¹⁵ (Higher byte)
- 307) Channel Selection (Lower byte)
- 308) Interference Analysis Cellular Standard¹¹⁶
- 309) Interference Analysis Estimated Bandwidth (Highest byte)
- 310) Interference Analysis Estimated Bandwidth
- 311) Interference Analysis Estimated Bandwidth
- 312) Interference Analysis Estimated Bandwidth (Lowest byte)
- 313) Interference Analysis Frequency¹¹⁷ (Highest byte)
- 314) Interference Analysis Frequency
- 315) Interference Analysis Frequency
- 316) Interference Analysis Frequency (Lowest byte)
- 317-320) Reserved
- 321) Trigger Type¹¹⁸
- 322) Trigger Position (0 – 100%)
- 323) Min Sweep Time (in μ s) (Highest byte)
- 324) Min Sweep Time (in μ s)
- 325) Min Sweep Time (in μ s)
- 326) Min Sweep Time (in μ s) (Lowest byte)
- 327) Video Trigger Level¹¹⁹ (Highest byte)
- 328) Video Trigger Level
- 329) Video Trigger Level
- 330) Video Trigger Level (Lowest byte)
- 331) Status Byte 8 (0b = Off, 1b = On)
 - (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B)
 - bit 2: Max Hold On/Off
 - bit 3: Min Hold On/Off
 - bit 4: Transmission Calibration On/Off (Option 21 Only)
 - bit 5: Bias Tee On/Off (Option 10 Only)
 - bit 6: Occupied BW Measurement On/Off
 - bit 7: Not Used
- 332) Impedance (00h = 50 Ω , 0Ah = 75 Ω Anritsu Adapter, 0Ch = 75 Ω Other Adapter)
- 333) Impedance Loss¹²⁰ (Higher byte)
- 334) Impedance Loss (Lower byte)
- 335) Frequency Scale Factor¹²¹ (Higher byte)
- 336) Frequency Scale Factor (Lower byte)
- 337) Frequency Range Minimum¹²² (Highest byte)
- 338) Frequency Range Minimum
- 339) Frequency Range Minimum
- 340) Frequency Range Minimum (Lowest byte)
- 341) Frequency Range Maximum¹²³ (Highest byte)
- 342) Frequency Range Maximum

¹¹³ 1 byte in MHz (i.e. 20 = 20MHz)

¹¹⁴ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

¹¹⁵ “No Channel” is sent as FFFEh

¹¹⁶ 4 Standards – 00h = 1250kHz CDMA, 01h = GSM, 02h = TDMA, 03h = AMPS, 04h = Unknown, FFh = Interference Analysis Measurement OFF

¹¹⁷ Scaled by Frequency Scale Factor (bytes 335-336)

¹¹⁸ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

¹¹⁹ Value sent as (value in dBm * 1000) + 270,000

¹²⁰ Value sent as (value in dB * 1000), valid values are 0 to 20 dB

¹²¹ In number of Hz

¹²² Scaled by Frequency Scale Factor

¹²³ Scaled by Frequency Scale Factor

- 343) Frequency Range Maximum
- 344) Frequency Range Maximum (Lowest byte)
- 345) Linked Trace Number (1-200)
- 346) Status Byte 9 (0b = Off, 1b = On)
 - (LSB) bit 0: C/I Measurement On/Off
 - bits 1-3: C/I Carrier Trace/Signal Type¹²⁴
 - bits 4-7: Not Used
- 347) C/I Calculated Power¹²⁵ (Carrier or Interference – NB FHSS)¹²⁶ (Highest byte)
- 348) C/I Calculated Power (Carrier or Interference – NB FHSS)
- 349) C/I Calculated Power (Carrier or Interference – NB FHSS)
- 350) C/I Calculated Power (Carrier or Interference – NB FHSS) (Lowest byte)
- 351) C/I Calculated Power¹²⁷ (Interference – WB FHSS)¹²⁸ (Highest byte)
- 352) C/I Calculated Power (Interference – WB FHSS)
- 353) C/I Calculated Power (Interference – WB FHSS)
- 354) C/I Calculated Power (Interference – WB FHSS) (Lowest byte)
- 355) C/I Calculated Power¹²⁹ (Interference – Broadband)¹³⁰ (Highest byte)
- 356) C/I Calculated Power (Interference – Broadband)
- 357) C/I Calculated Power (Interference – Broadband)
- 358) C/I Calculated Power (Interference – Broadband) (Lowest byte)
- 359) Occupied Bandwidth Power(Highest byte)¹³¹
- 360) Occupied Bandwidth Power
- 361) Occupied Bandwidth Power
- 362) Occupied Bandwidth Power (Lowest byte)
- 363) Marker Type¹³²
 - 364-367) GPS Position – Latitude (long integer)¹³³
 - 368-371) GPS Position – Longitude (long integer)
 - 372-373) GPS Position – Altitude (short integer)
 - 374) Signal Standard Link Type¹³⁴
- 375-398) Signal Standard Name, 24 bytes in ASCII
- 399) Measure Offset Status (00h = Off, 01h = On)
- 400-431) Not Used
- 432-2035) Sweep Data (401 points * 4 bytes/point= 1604 bytes)
 - 4 bytes for each data point
 - 1. dBm¹³⁵ (Highest byte)
 - 2. dBm
 - 3. dBm
 - 4. dBm (Lowest byte)

¹²⁴ 000b = Carrier – NB FHSS, 001b = Carrier – WB FHSS, 010b = Carrier – Broadband, 111b = Interference

¹²⁵ Value sent as (value in dBm * 1000) + 270,000

¹²⁶ If Status Byte 9, bytes 1-3 equal 111b, then signal will be calculated power for the Interference – NB FHSS trace. Otherwise, these bytes represent the calculated Carrier power.

¹²⁷ Value sent as (value in dBm * 1000) + 270,000

¹²⁸ If Status Byte 9, bytes 1-3 equal 111b, then signal will be calculated power for the Interference – WB FHSS trace. Otherwise, these bytes should be ignored.

¹²⁹ Value sent as (value in dBm * 1000) + 270,000

¹³⁰ If Status Byte 9, bytes 1-3 equal 111b, then signal will be calculated power for the Interference – Broadband trace. Otherwise, these bytes should be ignored.

¹³¹ If Method is % of power then the value is db Down * 1000. If the method is db down, then the value is %

¹³² 00h = Regular Marker, 01h = Noise Marker

¹³³ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

¹³⁴ 1 – Uplink, 2 – Downlink, 3 – Both, 0 – Invalid Link

¹³⁵ Value sent as (value in dBm * 1000) + 270,000

For Power Meter Mode (Option 29 Only):

- 57) Power Monitor Mode (00h = Off, 01h = On)
- 58) Power Meter Unit (00h = dBm, 01h = Watts)
- 59) Start Frequency¹³⁶ (Highest byte)
- 60) Start Frequency
- 61) Start Frequency
- 62) Start Frequency (Lowest byte)
- 63) Stop Frequency¹³⁷ (Highest byte)
- 64) Stop Frequency
- 65) Stop Frequency
- 66) Stop Frequency (Lowest byte)
- 67) Center Frequency¹³⁸ (Highest byte)
- 68) Center Frequency
- 69) Center Frequency
- 70) Center Frequency (Lowest byte)
- 71) Frequency Span¹³⁹ (Highest byte)
- 72) Frequency Span
- 73) Frequency Span
- 74) Frequency Span (Lowest byte)
- 75) Power Offset Status (00h = Off, 01h = On)
- 76) Power Offset¹⁴⁰ (Highest byte)
- 77) Power Offset
- 78) Power Offset
- 79) Power Offset (Lowest byte)
- 80) Power Relative Status (00h = Off, 01h = On)
- 81) Power Relative Value¹⁴¹ (Highest byte)
- 82) Power Relative Value
- 83) Power Relative Value
- 84) Power Relative Value (Lowest byte)
- 85) RMS Averaging Level (00h = Off, 01h = Low, 02h = Medium, 03h = High)
- 86) Power Zero Status (00h = Off, 01h = On)
- 87) External Reference Status (00h = Off, 01h = On)
- 88) External Reference Frequency (in Hz) (Highest byte)
- 89) External Reference Frequency (in Hz)
- 90) External Reference Frequency (in Hz)
- 91) External Reference Frequency (in Hz) (Lowest byte)
- 92) Signal Standard¹⁴² (Highest byte)
- 93) Signal Standard (Lowest byte)
- 94) Channel Selection¹⁴³ (Highest byte)
- 95) Channel Selection (Lowest byte)
- 96) Frequency Scale Factor¹⁴⁴ (Higher byte)
- 97) Frequency Scale Factor (Lower byte)
- 98) Frequency Range Minimum¹⁴⁵ (Highest byte)
- 99) Frequency Range Minimum
- 100) Frequency Range Minimum

¹³⁶ Scaled by Frequency Scale Factor (bytes 96-97)

¹³⁷ Scaled by Frequency Scale Factor (bytes 96-97)

¹³⁸ Scaled by Frequency Scale Factor (bytes 96-97)

¹³⁹ Scaled by Frequency Scale Factor (bytes 96-97)

¹⁴⁰ Value sent as (value in dB * 1000), valid values are 0 to 60 dB

¹⁴¹ Value sent as (value in dBm * 1000)

¹⁴² Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

¹⁴³ “No Channel” is sent as FFFEh

¹⁴⁴ In number of Hz

¹⁴⁵ Scaled by Frequency Scale Factor

- 101) Frequency Range Minimum (Lowest byte)
- 102) Frequency Range Maximum¹⁴⁶ (Highest byte)
- 103) Frequency Range Maximum
- 104) Frequency Range Maximum
- 105) Frequency Range Maximum (Lowest byte)
- 106 – 150) Not Used
- 151) Power Meter Reading¹⁴⁷ (Highest byte)
- 152) Power Meter Reading
- 153) Power Meter Reading
- 154) Power Meter Reading (Lowest byte)
- 155) Measure Offset Status (00h = Off, 01h = On)

For Channel Scanner Mode:

- 57) Reference Level (Highest Byte)
- 58) Reference Level
- 59) Reference Level
- 60) Reference Level (Lowest Byte)
- 61) Scale Division (Highest Byte)
- 62) Scale Division
- 63) Scale Division
- 64) Scale Division (Lowest Byte)
- 65) Start Frequency (Highest Byte)
- 66) Start Frequency
- 67) Start Frequency
- 68) Start Frequency (Lowest Byte)
- 69) Span Frequency (Highest Byte)
- 70) Span Frequency
- 71) Span Frequency
- 72) Span Frequency (Lowest Byte)
- 73) Channel Step (Highest Byte)
- 74) Channel Step (Lowest Byte)
- 75) Channel Frequency Step (Highest Byte)
- 76) Channel Frequency Step
- 77) Channel Frequency Step
- 78) Channel Frequency Step (Lowest Byte)
- 79) Number of Channels Displayed
- 80) External Reference Frequency¹⁴⁸
- 81) Display Type Channels or Frequencies¹⁴⁹
- 82) Display Type Graph or Text¹⁵⁰
- 83) Signal Standard (Highest Byte)
- 84) Signal Standard
- 85) Signal Standard
- 86) Signal Standard (Lowest Byte)
- 87-90) GPS Position – Latitude (long integer)¹⁵¹
- 91-94) GPS Position – Longitude (long integer)
- 95-96) GPS Position – Altitude (short integer)
- 97) Start Channel (Highest Byte)

¹⁴⁶ Scaled by Frequency Scale Factor

¹⁴⁷ Power sent as (power in dBm * 1000). Use two's-complement method to decode negative power levels.

¹⁴⁸ Frequency in MHz, OFF if 0

¹⁴⁹ 0 – Channel, 1 - Frequency

¹⁵⁰ 0 – Graph, 1 - Text

¹⁵¹ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

98)	Start Channel
99)	Start Channel
100)	Start Channel (Lowest Byte)
101 – 124)	Signal Standard Name, 24 bytes in ASCII
125 – 152)	Reserved
153 – 272)	Channel Scanner Data ¹⁵²

For Interference Analyzer RSSI Mode

57)	Center Frequency (Highest Byte)
58)	Center Frequency
59)	Center Frequency
60)	Center Frequency (Lowest Byte)
61)	Reference Level (Highest Byte)
62)	Reference Level
63)	Reference Level
64)	Reference Level (Lowest Byte)
65)	Scale (Highest Byte)
66)	Scale
67)	Scale
68)	Scale (Lowest Byte)
69)	RBW (Highest Byte)
70)	RBW
71)	RBW
72)	RBW (Lowest Byte)
73)	VBW (Highest Byte)
74)	VBW
75)	VBW
76)	VBW (Lowest Byte)
77)	Status Byte 1 <ul style="list-style-type: none"> Bit 0 - Detection Algorithm (Lowest Bit)¹⁵³ Bit 1 - Detection Algorithm Bit 2 - Detection Algorithm (Highest Bit) Bit 3 - Not Used Bit 4 - Not Used Bit 5 - Not Used Bit 6 - Not Used
78)	Reference Level Offset (Highest Byte)
79)	Reference Level Offset
80)	Reference Level Offset
81)	Reference Level Offset (Lowest Byte)
82)	External Reference Frequency ¹⁵⁴
83)	Signal Standard (Highest Byte)
84)	Signal Standard (Lowest Byte)
85)	Channel (Highest Byte) ¹⁵⁵
86)	Channel (Lowest Byte)
87)	Min RSSI Measured (Highest Byte)
88)	Min RSSI Measured
89)	Min RSSI Measured
90)	Min RSSI Measured (Lowest Byte)
91)	Max RSSI Measured (Highest Byte)

¹⁵² 20 points, 6 bytes per point. First 2 bytes are channel numbers(Invalid channels sent as 0xFFFF) and 4 bytes are values. Value sent as (value in dBm) * 1000 + 270,000

¹⁵³ 000 – Positive Peak, 010 – RMS Averaging, 100 – Negative Peak, 110 – Sampling Mode

¹⁵⁴ Frequency in MHz, OFF if 0

¹⁵⁵ Invalid channels are sent as 0xFFFF

92)	Max RSSI Measured
93)	Max RSSI Measured
94)	Max RSSI Measured (Lowest Byte)
95)	Measure Duration (Highest Byte) ¹⁵⁶
96)	Measure Duration
97)	Measure Duration
98)	Measure Duration (Lowest Byte)
99)	Sweep Point Interval(Highest Byte) ¹⁵⁷
100)	Sweep Point Interval
101)	Sweep Point Interval
102)	Sweep Point Interval (Lowest Byte)
103 – 106)	GPS Position – Latitude (long integer) ¹⁵⁸
107 – 110)	GPS Position – Longitude (long integer)
111 – 112)	GPS Position – Altitude (short integer)
113)	Signal Standard
114 – 117)	Start GPS Position – Latitude (long integer) ¹⁵⁹
118 – 121)	Start GPS Position – Longitude (long integer)
122 – 123)	Start GPS Position – Altitude (short integer)
124)	Attenuation (Highest Byte) ¹⁶⁰
125)	Attenuation
126)	Attenuation
127)	Attenuation (Lowest Byte)
128 – 151)	Signal Standard Name, 24 bytes in ASCII
152)	Measure Offset Status (00h = Off, 01h = On)
153 – 207)	Reserved
208 – 3415)	RSSI Sweep data ¹⁶¹

For High Accuracy Power Meter Mode

57)	Center Frequency(Highest Byte) ¹⁶²
58)	Center Frequency
59)	Center Frequency
60)	Center Frequency(Lowest Byte)
61)	Power Reading(Highest Byte) ¹⁶³
62)	Power Reading(Lowest Byte)
63)	Max Hold Status (00h = Off, 01h = On)
64)	Offset Status (00h = Off, 01h = On)
65)	Offset Value(Highest Byte) ¹⁶⁴
66)	Offset Value(Lowest Byte)
67)	Measure Offset Status (00h = Off, 01h = On)
68)	Measure Offset Value(Highest Byte) ¹⁶⁵

¹⁵⁶ Measure Duration time in minutes

¹⁵⁷ Sweep Point Interval time in milliseconds

¹⁵⁸ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

¹⁵⁹ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

¹⁶⁰ Attenuation is sent as (Att in dB * 1000)

¹⁶¹ Sweep Data contains 401 display points, 8 bytes per display point. The first 4 bytes are the amplitude, the next 2 bytes are the latitude increments from the start GPS position and the following 2 bytes are the longitude increments from the Start GPS position.

¹⁶² in kHz

¹⁶³ in 2-complement and in dBm

¹⁶⁴ in 2-complement and in dB

69)	Measure Offset Value(Lowest Byte)
70)	Relative Value(Highest Byte) ¹⁶⁶
71)	Relative Value(Lowest Byte)
72)	Relative Status (00h = Off, 01h = On)
73)	Running Averages Number(Highest Byte)
74)	Running Averages Number(Lowest Byte)
75 – 76)	Signal Standard ID
77 – 100)	Signal Standard Name
101)	Zero Status (00h = Off, 01h = On)
102)	Limit Status (00h = Off, 01h = On)
103)	Upper Limit dBm(Highest Byte) ¹⁶⁷
104)	Upper Limit dBm(Lowest Byte)
105)	Lower Limit dBm(Highest Byte) ¹⁶⁸
106)	Lower Limit dBm(Lowest Byte)
107)	Limit Unit Display
108)	Error Message Status ¹⁶⁹
109 – 112)	GPS Position – Latitude (long integer) ¹⁷⁰
113 – 116)	GPS Position – Longitude (long integer)
117 – 118)	GPS Position – Altitude (short integer)
119 – 128)	UTC Time, 10 bytes in ASCII
129 – 256)	Reserved Byte

Spectrum Master Returns (For invalid sweeps/empty stored sweep locations): 11 bytes

- 1-2) Number of following bytes (9 bytes for invalid sweep recall)
- 3) Current Instrument Date Format¹⁷¹
- 4) Model # (unsigned integer, 16h for Spectrum Master model MS2711D)
- 5-11) Extended Model # (7 bytes in ASCII)

Spectrum Master Returns (Invalid sweep location): 1 byte

- 1) 224 (E0) Parameter Error: Invalid sweep location
-

¹⁶⁵ in 2-complement and in dB

¹⁶⁶ in 2-complement and in dBm

¹⁶⁷ in 2-complement

¹⁶⁸ in 2-complement

¹⁶⁹ Bit 0: set to 1 if there is power supply error in the power sensor module. Bit 1: set to 1 if there is too much RF power going into the sensor module. Bit 2: set to 1 if zeroing is done incorrectly. Bit 3: set to 1 if power sensor's operating temperature range is exceeded. Bit 4: set to 1 if temperature has drifted by more than specified degree since the last zeroing.

¹⁷⁰ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

¹⁷¹ 00h = MM/DD/YYYY, 01h = DD/MM/YYYY, 02h = YYYY/MM/DD

OBSOLETE: Set A/B Trace – Control Byte #34 (22h)

This command is included for compatibility with the MS2711B. To access the new features, use Control Byte #35.

Description: Defines traces “A” and “B”.

Trace A is always the currently measured data (with or without trace math). It is always visible.

Trace B is always stored data and may come from a saved sweep or a previous “A” trace. There is no default for trace B. Trace B can be ON (visible) or OFF.

Bytes to Follow: 3 bytes

- 1) “A” trace display (00h = A only, 01h = A-B, 02h = A+B)
- 2) “B” trace status (00h = OFF, 01h = ON)
- 3) “B” trace number
0 = save current “A” data into “B” buffer, use that as “B”
1-200 = trace number
255 = no “B” trace defined

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Incorrect “A” or “B” trace, “B” trace requested to be used in calculations or displayed, but no trace or invalid trace specified
 - 238 (EEh) Time-out Error
-

Set SPA A/B Trace – Control Byte #35 (23h)

This command is new to the MS2711D. Use it, instead of Control Byte #34, to access the new features.

Description: Defines traces “A” and “B” for Spectrum Analyzer mode.

Trace A is always the currently measured data (with or without trace math). It is always visible.

Trace B is always stored data and may come from a saved sweep or a previous “A” trace. There is no default for trace B. Trace B can be ON (visible) or OFF.

Bytes to Follow: 3 bytes

- 1) “A” trace display (00h = A only, 01h = A-B, 02h = A+B)
- 2) “B” trace status (00h = Off, 01h = On)
- 3) “B” trace number
0 = save current “A” data into “B” buffer, use that as “B”
1-200 = trace number
255 = no “B” trace defined

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Not enough bytes transferred, “B” trace requested to be used in calculations or displayed, but no trace or invalid trace specified
 - 238 (EEh) Time-out Error
-

Upload Spectrum Master Sweep Trace – Control Byte #36 (24h)

This command is new to the MS2711D. Use it, instead of Control Byte #26, to access the new features.

Description: Uploads a sweep trace to the Spectrum Master.

Bytes to Follow:

For All Modes:

- 1-2) # of following bytes
- 3) Measurement Mode¹⁷²
- 4-7) Time/Date (in Long Integer)
- 8-17) Date in String Format (MM/DD/YYYY)
- 18-25) Time in String Format (HH:MM:SS)
- 26-41) Reference number stamp (16 ASCII bytes)
- 42-43) # of data points (401)

For Spectrum Analyzer Mode:

- 44) Start Frequency¹⁷³ (Highest byte)
- 45) Start Frequency
- 46) Start Frequency
- 47) Start Frequency (Lowest byte)
- 48) Stop Frequency¹⁷⁴ (Highest byte)
- 49) Stop Frequency
- 50) Stop Frequency
- 51) Stop Frequency (Lowest byte)
- 52) Center Frequency¹⁷⁵ (Highest byte)
- 53) Center Frequency
- 54) Center Frequency
- 55) Center Frequency (Lowest byte)
- 56) Frequency Span¹⁷⁶ (Highest byte)
- 57) Frequency Span
- 58) Frequency Span
- 59) Frequency Span (Lowest byte)
- 60) Ref Level¹⁷⁷ (Highest byte)
- 61) Ref Level
- 62) Ref Level
- 63) Ref Level (Lowest byte)
- 64) Scale per div¹⁷⁸ (Highest byte)
- 65) Scale per div
- 66) Scale per div
- 67) Scale per div (Lowest byte)
- 68) Marker 1¹⁷⁹ (Higher byte)
- 69) Marker 1 (Lower byte)
- 70) Marker 2 (Higher byte)
- 71) Marker 2 (Lower byte)
- 72) Marker 3 (Higher byte)
- 73) Marker 3 (Lower byte)

¹⁷² See Control Byte #3 “Set Measurement Mode” for available measurement modes.

¹⁷³ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁷⁴ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁷⁵ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁷⁶ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁷⁷ Value sent as (value in dBm * 1000) + 270,000

¹⁷⁸ Value sent as (value * 1000)

¹⁷⁹ Marker values are sent as # of data point on display.

See Control Byte #102, “Set Spectrum Analyzer Marker” for calculation of data point.

- 74) Marker 4 (Higher byte)
- 75) Marker 4 (Lower byte)
- 76) Marker 5 (Higher byte)
- 77) Marker 5 (Lower byte)
- 78) Marker 6 (Higher byte)
- 79) Marker 6 (Lower byte)
- 80) Single Limit¹⁸⁰ (Highest byte)
- 81) Single Limit
- 82) Single Limit
- 83) Single Limit (Lowest byte)
- 84) Multiple Upper Limit 1 Start X¹⁸¹ (Highest byte)
- 85) Multiple Upper Limit 1 Start X
- 86) Multiple Upper Limit 1 Start X
- 87) Multiple Upper Limit 1 Start X (Lowest byte)
- 88) Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)
- 89) Multiple Upper Limit 1 Start Y (Power Level)
- 90) Multiple Upper Limit 1 Start Y (Power Level)
- 91) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 92) Multiple Upper Limit 1 End X¹⁸² (Highest byte)
- 93) Multiple Upper Limit 1 End X
- 94) Multiple Upper Limit 1 End X
- 95) Multiple Upper Limit 1 End X (Lowest byte)
- 96) Multiple Upper Limit 1 End Y (Power Level) (Highest byte)
- 97) Multiple Upper Limit 1 End Y (Power Level)
- 98) Multiple Upper Limit 1 End Y (Power Level)
- 99) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 100-243) Multiple Upper Limits 2-5, Multiple Lower Limits 1-5 (see bytes 84-99 for format)
- 244) RBW Setting¹⁸³ (Highest byte)
- 245) RBW Setting
- 246) RBW Setting
- 247) RBW Setting (Lowest byte)
- 248) VBW Setting¹⁸⁴ (Highest byte)
- 249) VBW Setting
- 250) VBW Setting
- 251) VBW Setting (Lowest byte)
- 252) OCC BW Method (00h = % of power, 01h = dB down)
- 253) OCC BW % Value (0-99)
- 254) OCC BW dBc (0-120)
- 255) Attenuation¹⁸⁵ (Highest byte)
- 256) Attenuation
- 257) Attenuation
- 258) Attenuation (Lowest byte)
- 259-274) Antenna Name (16 bytes in ASCII)

¹⁸⁰ All amplitude values are sent as (value in dBm * 1000) + 270,000

¹⁸¹ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁸² Scaled by Frequency Scale Factor (bytes 318-319)

¹⁸³ Valid frequencies (in Hz) are 100, 300, 1,000, 3,000, 10,000, 30,000, 100,000, 300,000, 1,000,000

¹⁸⁴ Valid frequencies (in Hz) are 100, 300, 1,000, 3,000, 10,000, 30,000, 100,000, 300,000

¹⁸⁵ Value sent as (value * 1000)

- 275) Status Byte 1: (0b = Off, 1b = On)

 - (LSB) bit 0 : Marker 1 On/Off
 - bit 1 : Marker 2 On/Off
 - bit 2 : Marker 3 On/Off
 - bit 3 : Marker 4 On/Off
 - bit 4 : Marker 5 On/Off
 - bit 5 : Marker 6 On/Off
 - bits 6-7: Not Used

276) Status Byte 2: (0b = Off, 1b = On)

 - (LSB) bit 0 : Not Used
 - bit 1 : Marker 2 Delta On/Off
 - bit 2 : Marker 3 Delta On/Off
 - bit 3 : Marker 4 Delta On/Off
 - bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
 - bit 5 : Pre Amp Status On/Off
 - bit 6 : Dynamic Attenuation On/Off
 - bit 7 : Normalization On/Off

277) Status Byte 3: (0b = Off, 1b = On)

 - (LSB) bit 0 : Antenna Factor Correction On/Off
 - bits 1-2 : Detection alg (00b = pos. peak 01b = RMS Averaging 10b= neg. peak, 11 = Sampling Mode)
 - bits 3-4 : Amplitude Units (log) (00b = dBm 01b = dBV 10b = dBmV 11b = dBuV)
(Linear) – (00b = Watts 01b = Volts)
 - bit 5: Channel Power On/Off
 - bit 6: Adjacent Channel Power Ratio On/Off
 - bit 7 : Units Type (0b = Log 1b = Linear)

278) Status Byte 4

 - (0b = Off/Beep if data is BELOW line, 1b = On/Beep if data is ABOVE line)
 - (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
 - bit 1 : Single Limit On/Off
 - bit 2 : Single Limit Beep Level (0b = beep when data is below line 1b = above)
 - bit 3 : Not Used
 - bit 4 : Multiple Limit Upper Segment 1 Status On/Off
 - bit 5 : Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW
 - bit 6 : Multiple Limit Upper Segment 2 Status On/Off
 - bit 7 : Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW

279) Status Byte 5

 - (0b = Off/Beep if data is BELOW line, 1b =On/Beep if data is ABOVE line)
 - (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
 - bit 1 : Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
 - bit 2 : Multiple Limit Upper Segment 4 Status On/Off
 - bit 3 : Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
 - bit 4 : Multiple Limit Upper Segment 5 Status On/Off
 - bit 5 : Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
 - bit 6 : Multiple Limit Lower Segment 1 Status On/Off
 - bit 7 : Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW

280) Status Byte 6

 - (0b = Off/Beep if data is BELOW line, 1b = On/Beep if data is ABOVE line)
 - (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 - bit 1 : Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 - bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 - bit 3 : Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 - bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 - bit 5 : Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 - bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 - bit 7 : Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW

- 281) Status Byte 7
 (LSB) bits 0-6: Number of Sweeps to Average (1-25, 1 implies averaging OFF)
 bit 7 : Not Used
- 282) Reference Level Offset¹⁸⁶ (Highest byte)
- 283) Reference Level Offset
- 284) Reference Level Offset
- 285) Reference Level Offset (Lowest byte)
- 286) External Reference Frequency¹⁸⁷
- 287) Signal Standard¹⁸⁸ (Higher byte)
- 288) Signal Standard (Lower byte)
- 289) Channel Selection¹⁸⁹ (Higher byte)
- 290) Channel Selection (Lower byte)
- 291) Interference Analysis Cellular Standard¹⁹⁰
- 292) Interference Analysis Estimated Bandwidth (Highest byte)
- 293) Interference Analysis Estimated Bandwidth
- 294) Interference Analysis Estimated Bandwidth
- 295) Interference Analysis Estimated Bandwidth (Lowest byte)
- 296) Interference Analysis Frequency¹⁹¹ (Highest byte)
- 297) Interference Analysis Frequency
- 298) Interference Analysis Frequency
- 299) Interference Analysis Frequency (Lowest byte)
- 300-303) Reserved
- 304) Trigger Type¹⁹²
- 305) Trigger Position (0 – 100%)
- 306) Min Sweep Time (in μ s) (Highest byte)
- 307) Min Sweep Time (in μ s)
- 308) Min Sweep Time (in μ s)
- 309) Min Sweep Time (in μ s) (Lowest byte)
- 310) Video Trigger Level¹⁹³ (Highest byte)
- 311) Video Trigger Level
- 312) Video Trigger Level
- 313) Video Trigger Level (Lowest byte)
- 314) Status Byte 8 (0b = Off, 1b = On)
 (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B)
 bit 2: Max Hold On/Off
 bit 3: Min Hold On/Off
 bit 4: Transmission Calibration Status (Option 21 Only)
 bit 5: Bias Tee On/Off (Option 10 Only)
 bit 6: Occupied BW Measurement On/Off
 bit 7: Not Used
- 315) Impedance (00h = 50Ω, 0Ah = 75Ω Anritsu Adapter, 0Ch = 75Ω Other Adapter)
- 316) Impedance Loss¹⁹⁴ (Higher byte)
- 317) Impedance Loss (Lower byte)
- 318) Frequency Scale Factor¹⁹⁵ (Higher byte)

¹⁸⁶ Value sent as (Value in dBm * 1000) + 270,000

¹⁸⁷ byte in MHz (i.e. 20 = 20MHz)

¹⁸⁸ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh.

¹⁸⁹ “No Channel” is sent as FFFEh.

¹⁹⁰ 4 Standards – 00h = 1250kHz CDMA, 01h = GSM, 02h = TDMA, 03h = AMPS, 04h = Unknown FFh = Interference Analysis Measurement OFF

¹⁹¹ Scaled by Frequency Scale Factor (bytes 318-319)

¹⁹² Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

¹⁹³ Value sent as (Value in dBm * 1000) + 270,000

¹⁹⁴ Value sent as (value in dB * 1000), valid values are 0 to 20 dB

¹⁹⁵ In number of Hz

319)	Frequency Scale Factor (Lower byte)
320)	Frequency Range Minimum ¹⁹⁶ (Highest byte)
321)	Frequency Range Minimum
322)	Frequency Range Minimum
323)	Frequency Range Minimum (Lowest byte)
324)	Frequency Range Maximum ¹⁹⁷ (Highest byte)
325)	Frequency Range Maximum
326)	Frequency Range Maximum
327)	Frequency Range Maximum (Lowest byte)
328)	Linked Trace Number (1-200)
329)	Status Byte 9 (0b = Off, 1b = On)
	(LSB) bit 0: C/I Measurement On/Off
	bits 1-3: C/I Carrier Trace/Signal Type ¹⁹⁸
	bits 4-7: Not Used
330)	C/I Calculated Power ¹⁹⁹ (Carrier or Interference – NB FHSS ²⁰⁰) (Highest byte)
331)	C/I Calculated Power (Carrier or Interference – NB FHSS)
332)	C/I Calculated Power (Carrier or Interference – NB FHSS)
333)	C/I Calculated Power (Carrier or Interference – NB FHSS) (Lowest byte)
334)	C/I Calculated Power ²⁰¹ (Interference – WB FHSS ²⁰²) (Highest byte)
335)	C/I Calculated Power (Interference – WB FHSS)
336)	C/I Calculated Power (Interference – WB FHSS)
337)	C/I Calculated Power (Interference – WB FHSS) (Lowest byte)
338)	C/I Calculated Power ²⁰³ (Interference – Broadband ²⁰⁴) (Highest byte)
339)	C/I Calculated Power (Interference – Broadband)
340)	C/I Calculated Power (Interference – Broadband)
341)	C/I Calculated Power (Interference – Broadband) (Lowest byte)
342)	Marker Type ²⁰⁵
343-346)	GPS Position – Latitude (long integer) ²⁰⁶
347-350)	GPS Position – Longitude (long integer)
351-352)	GPS Position – Altitude (short integer)
353)	Reserved
354-377)	Signal Standard Name, 24 bytes in ASCII
378)	Measure Offset Status (00h = Off, 01h = On)
379-400)	Not Used

¹⁹⁶ Scaled by Frequency Scale Factor

¹⁹⁷ Scaled by Frequency Scale Factor

¹⁹⁸ 000b = Carrier – NB FHSS, 001b = Carrier – WB FHSS, 010b = Carrier – Broadband, 111b = Interference

¹⁹⁹ Value sent as (value in dBm * 1000) + 270,000

²⁰⁰ If Status Byte 9, bytes 1-3 equal 111b, then value will be calculated power for the Interference – NB FHSS trace. Otherwise, these bytes represent the calculated Carrier power.

²⁰¹ Value sent as (value in dBm * 1000) + 270,000

²⁰² If Status Byte 9, bytes 1-3 equal 111b, then value will be calculated power for the Interference – WB FHSS trace. Otherwise, these bytes should be ignored.

²⁰³ Value sent as (value in dBm * 1000) + 270,000

²⁰⁴ If Status Byte 9, bytes 1-3 equal 111b, then value will be calculated power for the Interference – Broadband trace. Otherwise, these bytes should be ignored.

²⁰⁵ 00h = Regular Marker, 01h = Noise Marker

²⁰⁶ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

401-2004) Sweep Data (401 points * 4 bytes/point = 1604 bytes)

4 bytes for each data point

1. dBm²⁰⁷ (Highest byte)
2. dBm
3. dBm
4. dBm (Lowest byte)

For Power Meter Mode (Option 29 Only):

- 57) Power Monitor Mode (00h = Off, 01h = On)
- 58) Power Meter Unit (00h = dBm, 01h = Watts)
- 59) Start Frequency²⁰⁸ (Highest byte)
- 60) Start Frequency
- 61) Start Frequency
- 62) Start Frequency (Lowest byte)
- 63) Stop Frequency²⁰⁹ (Highest byte)
- 64) Stop Frequency
- 65) Stop Frequency
- 66) Stop Frequency (Lowest byte)
- 67) Center Frequency²¹⁰ (Highest byte)
- 68) Center Frequency
- 69) Center Frequency
- 70) Center Frequency (Lowest byte)
- 71) Frequency Span²¹¹ (Highest byte)
- 72) Frequency Span
- 73) Frequency Span
- 74) Frequency Span (Lowest byte)
- 75) Power Offset Status (00h = Off, 01h = On)
- 76) Power Offset²¹² (Highest byte)
- 77) Power Offset
- 78) Power Offset
- 79) Power Offset (Lowest byte)
- 80) Power Relative Status (00h = Off, 01h = On)
- 81) Power Relative Value²¹³ (Highest byte)
- 82) Power Relative Value
- 83) Power Relative Value
- 84) Power Relative Value (Lowest byte)
- 85) RMS Averaging Level (00h = Off, 01h = Low, 02h = Medium, 03h = High)
- 86) Power Zero Status (00h = Off, 01h = On)
- 87) External Reference Status (00h = Off, 01h = On)
- 88) External Reference Frequency (in Hz) (Highest byte)
- 89) External Reference Frequency (in Hz)
- 90) External Reference Frequency (in Hz)
- 91) External Reference Frequency (in Hz) (Lowest byte)
- 92) Signal Standard²¹⁴ (Higher byte)
- 93) Signal Standard (Lower byte)
- 94) Channel Selection²¹⁵ (Higher byte)

²⁰⁷ Value sent as (Value in dBm * 1000) + 270,000

²⁰⁸ Scaled by Frequency Scale Factor (bytes 96-97)

²⁰⁹ Scaled by Frequency Scale Factor (bytes 96-97)

²¹⁰ Scaled by Frequency Scale Factor (bytes 96-97)

²¹¹ Scaled by Frequency Scale Factor (bytes 96-97)

²¹² Value sent as (value in dB * 1000), valid values are 0 to 60 dB

²¹³ Value sent as (value in dBm * 1000)

²¹⁴ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

²¹⁵ “No Channel” is sent as FFEEh

- 95) Channel Selection (Lower byte)
- 96) Frequency Scale Factor²¹⁶ (Higher byte)
- 97) Frequency Scale Factor (Lower byte)
- 98) Frequency Range Minimum²¹⁷ (Highest byte)
- 99) Frequency Range Minimum
- 100) Frequency Range Minimum
- 101) Frequency Range Minimum (Lowest byte)
- 102) Frequency Range Maximum²¹⁸ (Highest byte)
- 103) Frequency Range Maximum
- 104) Frequency Range Maximum
- 105) Frequency Range Maximum (Lowest byte)
- 96 – 150) Not Used
- 151) Power Meter Reading²¹⁹ (Highest byte)
- 152) Power Meter Reading
- 153) Power Meter Reading
- 154) Power Meter Reading (Lowest byte)
- 155) Measure Offset Status (00h = Off, 01h = On)

For Channel Scanner Mode:

- 44) Reference Level (Highest Byte)
- 45) Reference Level
- 46) Reference Level
- 47) Reference Level (Lowest Byte)
- 48) Scale Division (Highest Byte)
- 49) Scale Division
- 50) Scale Division
- 51) Scale Division (Lowest Byte)
- 52) Start Frequency (Highest Byte)
- 53) Start Frequency
- 54) Start Frequency
- 55) Start Frequency (Lowest Byte)
- 56) Span Frequency (Highest Byte)
- 57) Span Frequency
- 58) Span Frequency
- 59) Span Frequency (Lowest Byte)
- 60) Channel Step (Highest Byte)
- 61) Channel Step (Lowest Byte)
- 62) Channel Frequency Step (Highest Byte)
- 63) Channel Frequency Step
- 64) Channel Frequency Step
- 65) Channel Frequency Step (Lowest Byte)
- 66) Number of Channels Displayed
- 67) External Reference Frequency²²⁰
- 68) Display Type Channels or Frequencies²²¹
- 69) Display Type Graph or Text²²²
- 70) Signal Standard (Highest Byte)
- 71) Signal Standard (Lowest Byte)

²¹⁶ In number of Hz

²¹⁷ Scaled by Frequency Scale Factor

²¹⁸ Scaled by Frequency Scale Factor

²¹⁹ Power sent as (power in dBm * 1000). Use two's-complement method to decode negative power levels.

²²⁰ Frequency in MHz, OFF if 0

²²¹ 0 – Channel, 1 - Frequency

²²² 0 – Graph, 1 - Text

72-75)	GPS Position – Latitude (long integer) ²²³
76-79)	GPS Position – Longitude (long integer)
80-81)	GPS Position – Altitude (short integer)
82)	Start Channel (Highest Byte)
83)	Start Channel
84)	Start Channel
85)	Start Channel (Lowest Byte)
86 – 109)	Signal Standard Name, 24bytes in ASCII
110 – 137)	Reserved
138 – 257)	Channel Scanner Data ²²⁴

For Interference Analyzer RSSI Mode

44)	Center Frequency (Highest Byte)
45)	Center Frequency
46)	Center Frequency
47)	Center Frequency (Lowest Byte)
48)	Reference Level (Highest Byte)
49)	Reference Level
50)	Reference Level
51)	Reference Level (Lowest Byte)
52)	Scale (Highest Byte)
53)	Scale
54)	Scale
55)	Scale (Lowest Byte)
56)	RBW (Highest Byte)
57)	RBW
58)	RBW
59)	RBW (Lowest Byte)
60)	VBW (Highest Byte)
61)	VBW
62)	VBW
63)	VBW (Lowest Byte)
64)	Status Byte 1 <ul style="list-style-type: none"> Bit 0 - Detection Algorithm (Lowest Bit)²²⁵ Bit 1 - Detection Algorithm Bit 2 - Detection Algorithm (Highest Bit) Bit 3 - Not Used Bit 4 - Not Used Bit 5 - Not Used Bit 6 - Not Used
65)	Reference Level Offset (Highest Byte)
66)	Reference Level Offset
67)	Reference Level Offset
68)	Reference Level Offset (Lowest Byte)
69)	External Reference Frequency ²²⁶
70)	Signal Standard (Highest Byte)
71)	Signal Standard (Lowest Byte)

²²³ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

²²⁴ 20 points, 6 bytes per point. First 2 bytes are channel numbers(Invalid channels sent as 0xFFFF) and 4 bytes are values. Value sent as (value in dBm) * 1000 + 270,000

²²⁵ 000 - Positive Peak, 010 – RMS Averaging, 100 – Negative Peak, 110 – Sampling Mode

²²⁶ Frequency in MHz, OFF if 0

72)	Channel (Highest Byte) ²²⁷
73)	Channel (Lowest Byte)
74)	Min RSSI Measured (Highest Byte)
75)	Min RSSI Measured
76)	Min RSSI Measured
77)	Min RSSI Measured (Lowest Byte)
78)	Max RSSI Measured (Highest Byte)
79)	Max RSSI Measured
80)	Max RSSI Measured
81)	Max RSSI Measured (Lowest Byte)
82)	Measure Duration (Highest Byte) ²²⁸
83)	Measure Duration
84)	Measure Duration
85)	Measure Duration (Lowest Byte)
86)	Sweep Point Interval(Highest Byte) ²²⁹
87)	Sweep Point Interval
88)	Sweep Point Interval
89)	Sweep Point Interval (Lowest Byte)
90 - 93)	GPS Position – Latitude (long integer) ²³⁰
94 - 97)	GPS Position – Longitude (long integer)
98 - 99)	GPS Position – Altitude (short integer)
100)	Signal Standard
101-104)	Start GPS Position – Latitude (long integer) ²³¹
105-108)	Start GPS Position – Longitude (long integer)
109-110)	Start GPS Position – Altitude (short integer)
111)	Attenuation (Highest Byte) ²³²
112)	Attenuation
113)	Attenuation
114)	Attenuation (Lowest Byte)
115 – 138)	Signal Standard Name, 24bytes in ASCII
139)	Measure Offset Status (00h = Off, 01h = On)
140 – 194)	Reserved
195 – 3402)	RSSI Sweep data ²³³

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Not enough bytes transferred
 - 225 (E1h) Memory Error: Not enough memory to store data
 - 238 (EEh) Time-out Error
-

²²⁷ Invalid channels are sent as 0xFFFF

²²⁸ Measure Duration time in minutes

²²⁹ Sweep Point Interval time in milliseconds

²³⁰ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

²³¹ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

²³² Attenuation is sent as (Att in dB * 1000)

²³³ Sweep Data contains 401 display points, 8 bytes per display point. The first 4 bytes are the amplitude, the next 2 bytes are the latitude increments from the start GPS position and the following 2 bytes are the longitude increments from the Start GPS position.

Get Options – Control Byte #37 (25h)

Description: Queries the option(s) installed on the Spectrum Master, returns a list as an ASCII string.

Bytes to Follow: 0 bytes

Spectrum Master Returns: Depends on the option(s) installed

Maximum string: “3/6/10A/19/21/25/27/28/29/31”

If NO options are installed: “None”

Query Power Level – Control Byte #39 (27h)

This command is available with Option 29 and/or Option 5.

Description: Return Power Level at the RF In port. Also returns power meter settings.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 30 bytes

- 1) Status Byte #1 (0b = Off, 1b = On)
(LSB) bit 0 : Unit (0b - Watt/%, 1b – dBm/dBr)
bit 2 : Relative Mode On/Off
bit 3: Offset Mode On/Off
bit 4: Zero Mode On/Off
bits 5-7: Not Used
- 2) RMS Averaging Status²³⁴
3 - 6) Relative Mode Reference Power Level in dBm
7 - 10) Offset Mode Power Level
11 - 14) Zero Mode Power Level
15 - 18) Absolute Power Level
19 - 22) Power
23 --26) Center Frequency (scaled by frequency scale factor)
27 - 30) Span Frequency (scaled by frequency scale factor)

Notes:

Power is returned as (dBm * 1000)

Relative power is returned as (dB * 1000)

Offset is returned as (dB * 1000)

²³⁴ RMS Averaging – 00h = Off, 01h = Low, 02h = Medium, 03h = High

Set Power Meter Units – Control Byte #40 (28h)

This command is available with Option 29 and/or Option 5.

Description: Set Power Meter units to watts or dBm.

Bytes to Follow: 1 byte

- 1) Units
 - 00h = Watt (% if in relative mode)
 - 01h = dBm (dB if in relative mode)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid Units
 - 238 (EEh) Time-out Error
-

Power Meter Relative Mode On/Off – Control Byte #41 (29h)

This command is available with Option 29 and/or Option 5.

Description: Enable or disable Power Meter Relative Mode.

Bytes to Follow: 1 byte

- 1) Relative Mode State
 - 00h = Off
 - 01h = On w/ trigger (use the current power level as a reference power level)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid parameter
 - 238 (EEh) Time-out Error
-

Power Meter Offset Mode On/Off – Control Byte #42 (2Ah)

This command is available with Option 29 and/or Option 5.

Description: Enable or disable Power Meter Offset Mode.

Bytes to Follow: 5 bytes

- 1) On/Off (01h = On, 00h = Off)
- 2 - 5) Offset Power Level in dB (Multiplied by 1000)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid parameter
- 238 (EEh) Time-out Error

Note:

If you turn the Offset mode off, you must still send the other bytes. Bytes 2 - 5 will be ignored.

Power Meter Zero Mode On/Off – Control Byte #43 (2Bh)

This command is available with Option 29 and/or Option 5.

Description: Enable or disable Power Meter Zeroing Mode.

Bytes to Follow: 1 byte

- 1) Zero Mode Status
 - 00h = Off
 - 01h = On with trigger (current power level is referenced as -80 dBm)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid status
 - 238 (EEh) Time-out Error
-

Power Meter RMS Averaging On/Off – Control Byte #44 (2Ch)

This command is available with Option 29 only.

Description: Disable/enable Power Meter RMS Averaging. Enabling can be set to 3 different levels.

Bytes to Follow: 1 byte

- 1) RMS Averaging State
 - 00h = Off
 - 01h = On (Low) with trigger (current power level is referenced as -80 dBm)
 - 02h = On (Medium)
 - 03h = On (High)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid state
 - 238 (EEh) Time-out Error
-

Power Meter Center Frequency and Span Setup – Control Byte #45 (2Dh)

This command is available with Option 29 only.

Description: Sets the center frequency and span frequency for the Power Meter mode.

If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 8 bytes

- 1) Center Frequency (Highest byte)
- 2) Center Frequency
- 3) Center Frequency
- 4) Center Frequency (Lowest byte)
- 5) Span (Highest byte)
- 6) Span
- 7) Span
- 8) Span (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid frequency range
 - 238 (EEh) Time-out Error
-

Trigger Sweep – Control Byte #48 (30h)

Description: Causes the Spectrum Master to perform a sweep if it is in single sweep mode.

This command works only when the Spectrum Master is NOT in remote mode. Send this command, and then wait for the “Sweep Complete Byte” to signify the end of the sweep.

Note: The "Sweep Complete Byte" is not returned unless serial echo status is turned on using command #10.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 192 (C0h) Sweep Complete Byte (at the end of the sweep)
-

Sweep Data Echo On/Off – Control Byte #49 (31h)

Description: Sets the sweep data echo mode On/Off.

Sweep Data Echo Mode behaves much like the Serial Port Echo Mode (see Control Byte #10). It automatically puts the unit into single sweep mode. At the end of each sweep cycle, the Spectrum Master sends a Sweep Complete Byte #192 (C0h) to the serial port. At this time, sweep data can be queried (see Control Byte #33) without having to enter remote mode first or exit remote mode when done. Depending on the value of the second following byte, the next sweep can be automatically triggered after the sweep data has been sent.

This mode activates once the Spectrum Master exits from the remote mode. Sweep Data Echo status can't be saved to or recalled from saved setups. Cycling power resets the Sweep Data Echo status to Off.

The Sweep Data Echo Mode allows run-time handshaking between the Spectrum Master and computer by doing the following:

- 1) Enter remote mode. Set Sweep Data Echo Mode On. Exit remote mode.
- 2) The Spectrum Master sweeps once and then sends the Sweep Complete Byte.
- 3) After you receive it: Recall sweep 0 (last sweep trace in RAM).
- 4) If using auto triggering, repeat steps 2-3. If using manual triggering, go to step 5.
- 5) Send Sweep Triggering Byte #48 (30h) and wait for the next sweep cycle.
- 6) Repeat steps 2-5.

Note: To execute commands other than #33, you must use the traditional Enter Remote, Send Commands, Exit Remote communication sequence.

Bytes to Follow: 2 bytes

- 1) Sweep Data Echo Status
 - 00h : Off
 - 01h : On
- 2) Next Sweep Trigger
 - 00h : Manual
 - 01h : Automatic

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid sweep data echo status
 - 238 (EEh) Time-out Error
-

Check Battery Status – Control Byte #50 (32h)

Description: Return Smart Battery status.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 17 bytes

- 1-2) Battery Status flags (Refer to Smart Battery Data Spec 5.1.2.1)
- 3-4) State of Charge (unsigned integer 0 to 100(%))Full)
- 5-6) Battery Voltage (unsigned integer 0 to 65535 in mV)
- 7-8) Battery Current (signed integer -32,768 to +32,7687 mA, positive = Charging)
- 9-10) Battery Average current (signed integer -32,768 to +32,7687 mA, positive = Charging)
- 11-12) Average time to empty (unsigned integer 0 to 65535 minute)
- 13-14) Battery Charge Cycle Count (unsigned integer 0 to 65535 cycles)
- 15-16) Battery Capacity at Full Charge in mA Hours (unsigned integer 0 to 65535 cycles)
- 17) Unit under battery power (1 = YES; 0 = NO)

Note:

The Smart Battery Data Spec can be found at <http://www.sbs-forum.org/specs/index.html>

Set SPA Minimum Sweep Time – Control Byte #53 (35h)

Description: Sets the minimum sweep time (in μ s) for the spectrum analyzer when the span is 0.

Valid range is 50 to 200,000,000.

Bytes to Follow: 4 bytes

- 1) Minimum Sweep Time (in μ s) (Highest byte)
- 2) Minimum Sweep Time (in μ s)
- 3) Minimum Sweep Time (in μ s)
- 4) Minimum Sweep Time (in μ s) (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid sweep time
 - 238 (EEh) Time-out Error
-

Set Trigger Position – Control Byte #54 (36h)

Description: Sets the trigger position (in percent) for the spectrum analyzer when the span is 0.

Bytes to Follow: 1 byte

- 1) Trigger Position (0 – 100%)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid trigger position
 - 238 (EEh) Time-out Error
-

Set Video Trigger Level – Control Byte #55 (37h)

Description: Sets the trigger level (-120 - +20 dBm) for the spectrum analyzer when the span is 0 and trigger mode is video.

The trigger level should be sent as (value in dBm * 1000) + 120,000.

Bytes to Follow: 4 bytes

- 1) Trigger Level (Highest byte)
- 2) Trigger Level
- 3) Trigger Level
- 4) Trigger Level (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid trigger level
 - 238 (EEh) Time-out Error
-

Automatically Save Runtime Setup – Control Byte #64 (40h)

Description: Automatically save the runtime setup when exiting remote mode.

This flag must be set once per power cycle of the Spectrum Master. It returns to its default value when the unit is turned off. The default value is (0), DO NOT automatically save the runtime setup.

Bytes to Follow: 1 byte

- 1) Save runtime setup On/Off
00h = Off (default)
01h = On

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
238 (EEh) Time Out Error
-

Download Saved Setup – Control Byte #65 (41h)

Description: Returns parameters associated with the specified setup number. Since different modes have different numbers of setup locations available, the command requires the mode be specified as well as the setup number.

Bytes to Follow: 2 bytes

- 1) Measurement Mode²³⁵
- 2) Setup Number
 - 0 = Run time setup
 - 1 – 10 = Saved setups for Spectrum Analyzer/Transmission Measurement modes
 - 1 – 5 = Saved setups for Power Meter mode (Option 29 Only)

Spectrum Master Returns:

For All Modes:

- 1) Number of Following Bytes (Higher byte)
- 2) Number of Following Bytes (Lower byte)
- 3) Measurement Mode²³⁶
- 5-20) Not Used

For Spectrum Analyzer Mode:

- 21) Spectrum Analyzer Start Frequency²³⁷ (Highest byte)
- 22) Spectrum Analyzer Start Frequency
- 23) Spectrum Analyzer Start Frequency
- 24) Spectrum Analyzer Start Frequency (Lowest byte)
- 25) Spectrum Analyzer Stop Frequency²³⁸ (Highest byte)
- 26) Spectrum Analyzer Stop Frequency
- 27) Spectrum Analyzer Stop Frequency
- 28) Spectrum Analyzer Stop Frequency (Lowest byte)
- 29) Spectrum Analyzer Center Frequency²³⁹ (Highest byte)
- 30) Spectrum Analyzer Center Frequency
- 31) Spectrum Analyzer Center Frequency
- 32) Spectrum Analyzer Center Frequency (Lowest byte)
- 33) Spectrum Analyzer Frequency Span²⁴⁰ (Highest byte)

²³⁵ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

²³⁶ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

²³⁷ Scaled by Frequency Scale Factor (bytes 301-302)

²³⁸ Scaled by Frequency Scale Factor (bytes 301-302)

²³⁹ Scaled by Frequency Scale Factor (bytes 301-302)

²⁴⁰ Scaled by Frequency Scale Factor (bytes 301-302)

- 34) Spectrum Analyzer Frequency Span
- 35) Spectrum Analyzer Frequency Span
- 36) Spectrum Analyzer Frequency Span (Lowest byte)
- 37) Ref Level (Highest byte)²⁴¹
- 38) Ref Level
- 39) Ref Level
- 40) Ref Level (Lowest byte)
- 41) Scale per div (Highest byte)²⁴²
- 42) Scale per div
- 43) Scale per div
- 44) Scale per div (Lowest byte)
- 45) Spectrum Analyzer Frequency Marker 1 (Higher byte)²⁴³
- 46) Spectrum Analyzer Frequency Marker 1 (Lower byte)
- 47) Spectrum Analyzer Frequency Marker 2 (Higher byte)
- 48) Spectrum Analyzer Frequency Marker 2 (Lower byte)
- 49) Spectrum Analyzer Frequency Marker 3 (Higher byte)
- 50) Spectrum Analyzer Frequency Marker 3 (Lower byte)
- 51) Spectrum Analyzer Frequency Marker 4 (Higher byte)
- 52) Spectrum Analyzer Frequency Marker 4 (Lower byte)
- 53) Spectrum Analyzer Frequency Marker 5 (Higher byte)
- 54) Spectrum Analyzer Frequency Marker 5 (Lower byte)
- 55) Spectrum Analyzer Frequency Marker 6 (Higher byte)
- 56) Spectrum Analyzer Frequency Marker 6 (Lower byte)
- 57) Spectrum Analyzer Single Limit (Highest byte)²⁴⁴
- 58) Spectrum Analyzer Single Limit
- 59) Spectrum Analyzer Single Limit
- 60) Spectrum Analyzer Single Limit (Lowest byte)
- 61) SPA Multiple Upper Limit 1 Start X²⁴⁵ (Highest byte)
- 62) SPA Multiple Upper Limit 1 Start X
- 63) SPA Multiple Upper Limit 1 Start X
- 64) SPA Multiple Upper Limit 1 Start X (Lowest byte)
- 65) SPA Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)²⁴⁶
- 66) SPA Multiple Upper Limit 1 Start Y (Power Level)
- 67) SPA Multiple Upper Limit 1 Start Y (Power Level)
- 68) SPA Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 69) SPA Multiple Upper Limit 1 End X²⁴⁷ (Highest byte)
- 70) SPA Multiple Upper Limit 1 End X
- 71) SPA Multiple Upper Limit 1 End X
- 72) SPA Multiple Upper Limit 1 End X (Lowest byte)
- 73) SPA Multiple Upper Limit 1 End Y (Power Level) (Highest byte)²⁴⁸
- 74) SPA Multiple Upper Limit 1 End Y (Power Level)
- 75) SPA Multiple Upper Limit 1 End Y (Power Level)
- 76) SPA Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 77-220) SPA Multiple Upper Limits 2-5, SA Multiple Lower Limits 1-5 (see bytes 61-76 for format)
- 221) RBW Setting (Highest byte)²⁴⁹

²⁴¹ Value sent as (value in dBm * 1000) + 270,000

²⁴² Value sent as (value * 1000)

²⁴³ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points - 1)) + start frequency.

²⁴⁴ Value sent as (value in dBm * 1000) + 270000

²⁴⁵ Scaled by Frequency Scale Factor (bytes 301-302)

²⁴⁶ Value sent as (value in dBm * 1000) + 270000

²⁴⁷ Scaled by Frequency Scale Factor (bytes 301-302)

²⁴⁸ Value sent as (value in dBm * 1000) + 270000

²⁴⁹ RBW frequency sent in Hz.

- 222) RBW Setting
- 223) RBW Setting
- 224) RBW Setting (Lowest byte)
- 225) VBW Setting (Highest byte)²⁵⁰
- 226) VBW Setting
- 227) VBW Setting
- 228) VBW Setting (Lowest byte)
- 229) OCC BW Method²⁵¹
- 230) OCC BW % Value²⁵²
- 231) OCC BW dBc²⁵³
- 232) Attenuation
- 233) Antenna Index (0-14)
- 234-249) Antenna Name (16 bytes in ASCII)
- 250) Status Byte 1: (0b = Off , 1b = On)
 - (LSB) bit 0 : Spectrum Analyzer Mode Marker 1 On/Off
 - bit 1 : Spectrum Analyzer Mode Marker 2 On/Off
 - bit 2 : Spectrum Analyzer Mode Marker 3 On/Off
 - bit 3 : Spectrum Analyzer Mode Marker 4 On/Off
 - bit 4 : Spectrum Analyzer Mode Marker 5 On/Off
 - bit 5 : Spectrum Analyzer Mode Marker 6 On/Off
 - bits 6 - 7 : Not Used
- 251) Status Byte 2: (0b = Off, 1b = On)
 - (LSB) bit 0 : Not Used
 - bit 1 : Spectrum Analyzer Mode Marker 2 Delta On/Off
 - bit 2 : Spectrum Analyzer Mode Marker 3 Delta On/Off
 - bit 3 : Spectrum Analyzer Mode Marker 4 Delta On/Off
 - bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
 - bit 5 : Pre Amp Status On/Off
 - bit 6 : Dynamic Attenuation On/Off
 - bit 7 : Normalization On/Off
- 252) Status Byte 3: (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
 - (LSB) bit 0 : SPA Limit Type (0b = Single, 1b = Multiple)
 - bit 1 : SPA Single Limit Beep On/Off
 - bit 2 : SPA Single Limit Status On/Off
 - bit 3 : SPA Single Limit Beep Level ABOVE/BELLOW
 - bit 4 : SPA Multiple Limit Upper Segment 1 Status On/Off
 - bit 5 : SPA Multiple Limit Upper Segment 1 Beep Level ABOVE/BELLOW²⁵⁴
 - bit 6 : SPA Multiple Limit Upper Segment 2 Status On/Off
 - bit 7 : SPA Multiple Limit Upper Segment 2 Beep Level ABOVE/BELLOW

²⁵⁰ VBW frequency sent in Hz.

²⁵¹ 00h = % of power, 01h = dB down

²⁵² 0 – 99%

²⁵³ 0 – 120 dBc

²⁵⁴ Beep level is always 1b for upper segmented limit line

- 253) Status Byte 4 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : SPA Multiple Limit Upper Segment 3 Status On/Off
 bit 1 : SPA Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
 bit 2 : SPA Multiple Limit Upper Segment 4 Status On/Off
 bit 3 : SPA Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
 bit 4 : SPA Multiple Limit Upper Segment 5 Status On/Off
 bit 5 : SPA Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
 bit 6 : SPA Multiple Limit Lower Segment 1 Status On/Off
 bit 7 : SPA Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW²⁵⁵
- 254) Status Byte 5 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : SPA Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : SPA Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 bit 2 : SPA Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : SPA Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 bit 4 : SPA Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : SPA Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 bit 6 : SPA Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : SPA Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
- 255) Status Byte 6: (0b = Off, 1b = On)
 (LSB) bit 0 : Antenna Factors Correction On/Off
 bit 1 : Bias Tee On/Off (Option 10)
 bit 2 : Amplitude Units (Linear) – 00b = Watts 01b = Volts
 bits 3-4 : Amplitude Units (Log) - 00b = dBm 01b = dBV 10b = dBmV 11b = dBuV
 bits 5-6 : Detection Alg (00b = pos. peak 01b = RMS Averaging 10b = neg. peak 11b = Sampling Mode)
 bit 7 : Units Type (0b = Log 1b = Linear)
- 256) Status Byte 7: (0b = Off, 1b = On)
 (LSB) bit 0: Interference Analysis On/Off
 bit 1: C/I Measurement On/Off
 bit 2: RBW Coupling (1b = Auto, 0b = Manual)
 bit 3: VBW Coupling (1b = Auto, 0b = Manual)
 bit 4: Attenuation Coupling (1b = Auto, 0b = Manual)
 bit 5: Channel Power On/Off
 bit 6: Adjacent Channel Power On/Off
 bit 7: Occupied BW Measurement On/Off
- 257) Reference Level Offset²⁵⁶ (Highest byte)
 258) Reference Level Offset
 259) Reference Level Offset
 260) Reference Level Offset (Lowest byte)
 261) External Reference Frequency²⁵⁷
 262) Signal Standard²⁵⁸ (Higher byte)
 263) Signal Standard (Lower byte)
 264) Channel Selection²⁵⁹ (Higher byte)
 265) Channel Selection (Lower byte)
 266) Trigger Type²⁶⁰
 267) Interference Analysis Frequency²⁶¹ (Highest byte)

²⁵⁵ Beep level is always 0b for lower segmented limit line

²⁵⁶ Value sent as (value in dBm * 1000) + 270,000

²⁵⁷ 1 byte in MHz (i.e. 20 = 20MHz)

²⁵⁸ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

²⁵⁹ “No Channel” is sent as FFFEh

²⁶⁰ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

²⁶¹ Scaled by Frequency Scale Factor (bytes 301-302)

268)	Interference Analysis Frequency
269)	Interference Analysis Frequency
270)	Interference Analysis Frequency (Lowest byte)
271)	Trigger Position (0 – 100%)
272)	Min Sweep Time (in μ s) (Highest byte)
273)	Min Sweep Time (in μ s)
274)	Min Sweep Time (in μ s)
275)	Min Sweep Time (in μ s) (Lowest byte)
276)	Video Trigger Level ²⁶² (Highest byte)
277)	Video Trigger Level
278)	Video Trigger Level
279)	Video Trigger Level (Lowest byte)
280)	Status Byte 8 (LSB) bit 0: Limit Beep Output (Option 86 Only) (00h = Internal Speaker, 01h = External) bits 2-7: Not Used
281)	Status Byte 9 (LSB) bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF) bit 7: Not Used
282)	Status Byte 10: (0b = Off, 1b = On) (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B) bit 2: Max Hold On/Off bit 3: Min Hold On/Off bit 4: View B On/Off bit 5: External Reference Frequency On/Off bits 6-7: Not Used
283)	Impedance (00h = 50 Ω , 10h = 75 Ω Anritsu Adapter, 12h = 75 Ω Other Adapter)
284)	Impedance Loss ²⁶³ (Higher byte)
285)	Impedance Loss (Lower byte)
286)	AM/FM Demod Type ²⁶⁴
287)	AM/FM Demod Status (01h = On, 00h = Off)
288)	AM/FM Demod Volume (0 to 100)
289)	AM/FM Demod Frequency ²⁶⁵ (Highest byte)
290)	AM/FM Demod Frequency
291)	AM/FM Demod Frequency
292)	AM/FM Demod Frequency (Lowest byte)
293)	AM/FM Demod Time (in ms) (Highest byte)
294)	AM/FM Demod Time (in ms)
295)	AM/FM Demod Time (in ms)
296)	AM/FM Demod Time (in ms) (Lowest byte)
297)	SSB BFO Offset ²⁶⁶ (Highest byte)
298)	SSB BFO Offset
299)	SSB BFO Offset
300)	SSB BFO Offset (Lowest byte)
301)	Frequency Scale Factor ²⁶⁷ (Higher byte)
302)	Frequency Scale Factor (Lower byte)
303)	Frequency Range Minimum ²⁶⁸ (Highest byte)
304)	Frequency Range Minimum

²⁶² Value sent as (value in dBm * 1000) + 270,000

²⁶³ Value sent as (value in dB * 1000), valid values are 0 to 20 dB

²⁶⁴ AM/FM Demod Type: 00h = FM-Wide Band, 01h = FM-Narrow Band, 02h = AM, 03h = SSB Lower, 04h = SSB Upper

²⁶⁵ Scaled by Frequency Scale Factor (bytes 301-302)

²⁶⁶ Value sent as ((value in Hz) – 10,000)

²⁶⁷ In number of Hz

²⁶⁸ Scaled by Frequency Scale Factor (bytes 301-302)

- 305) Frequency Range Minimum
- 306) Frequency Range Minimum (Lowest byte)
- 307) Frequency Range Maximum²⁶⁹ (Highest byte)
- 308) Frequency Range Maximum
- 309) Frequency Range Maximum
- 310) Frequency Range Maximum (Lowest byte)
- 311) Marker Type²⁷⁰
- 312) Channel Power Int BW²⁷¹ (Highest byte)
- 313) Channel Power Int BW
- 314) Channel Power Int BW
- 315) Channel Power Int BW (Lowest byte)
- 316) ACPR Main Channel BW²⁷² (Highest byte)
- 317) ACPR Main Channel BW
- 318) ACPR Main Channel BW
- 319) ACPR Main Channel BW (Lowest byte)
- 320) ACPR Adjacent Channel BW²⁷³ (Highest byte)
- 321) ACPR Adjacent Channel BW
- 322) ACPR Adjacent Channel BW
- 323) ACPR Adjacent Channel BW (Lowest byte)
- 324) ACPR Channel Spacing²⁷⁴ (Highest byte)
- 325) ACPR Channel Spacing
- 326) ACPR Channel Spacing
- 327) ACPR Channel Spacing (Lowest byte)
- 328) Interference Analysis Cell Std²⁷⁵
- 329) Interference Analysis Est. BW²⁷⁶ (Highest byte)
- 330) Interference Analysis Est. BW
- 331) Interference Analysis Est. BW
- 332) Interference Analysis Est. BW (Lowest byte)
- 333) Trace B Trace Id²⁷⁷
- 334-500) Not Used

For Transmission Mode (Option 21):

- 21) Start Frequency²⁷⁸ (Highest byte)
- 22) Start Frequency
- 23) Start Frequency
- 24) Start Frequency (Lowest byte)
- 25) Stop Frequency²⁷⁹ (Highest byte)
- 26) Stop Frequency
- 27) Stop Frequency
- 28) Stop Frequency (Lowest byte)
- 29) Center Frequency²⁸⁰ (Highest byte)
- 30) Center Frequency

²⁶⁹ Scaled by Frequency Scale Factor (bytes 301-302)

²⁷⁰ 00h = Regular Marker, 01h = Noise Marker

²⁷¹ Scaled by Frequency Scale Factor (bytes 301-302)

²⁷² Scaled by Frequency Scale Factor (bytes 301-302)

²⁷³ Scaled by Frequency Scale Factor (bytes 301-302)

²⁷⁴ Scaled by Frequency Scale Factor (bytes 301-302)

²⁷⁵ 4 Standards – 00h = 1250 kHz CDMA, 01h = GSM, 02h = TDMA, 03h = AMPS, 04h = Unknown, FFh = Interference Analysis Measurement OFF

²⁷⁶ Frequency in Hz

²⁷⁷ FFh indicates no trace selected

²⁷⁸ Scaled by Frequency Scale Factor (bytes 244-245)

²⁷⁹ Scaled by Frequency Scale Factor (bytes 244-245)

²⁸⁰ Scaled by Frequency Scale Factor (bytes 244-245)

- 31) Center Frequency
- 32) Center Frequency (Lowest byte)
- 33) Frequency Span²⁸¹ (Highest byte)
- 34) Frequency Span
- 35) Frequency Span
- 36) Frequency Span (Lowest byte)
- 37) Ref Level (Highest byte)²⁸²
- 38) Ref Level
- 39) Ref Level
- 40) Ref Level (Lowest byte)
- 41) Scale per div (Highest byte)²⁸³
- 42) Scale per div
- 43) Scale per div
- 44) Scale per div (Lowest byte)
- 45) Frequency Marker 1 (Higher byte)²⁸⁴
- 46) Frequency Marker 1 (Lower byte)
- 47) Frequency Marker 2 (Higher byte)
- 48) Frequency Marker 2 (Lower byte)
- 49) Frequency Marker 3 (Higher byte)
- 50) Frequency Marker 3 (Lower byte)
- 51) Frequency Marker 4 (Higher byte)
- 52) Frequency Marker 4 (Lower byte)
- 53) Frequency Marker 5 (Higher byte)
- 54) Frequency Marker 5 (Lower byte)
- 55) Frequency Marker 6 (Higher byte)
- 56) Frequency Marker 6 (Lower byte)
- 57) Single Limit (Highest byte)²⁸⁵
- 58) Single Limit
- 59) Single Limit
- 60) Single Limit (Lowest byte)
- 61) Multiple Upper Limit 1 Start X²⁸⁶ (Highest byte)
- 62) Multiple Upper Limit 1 Start X
- 63) Multiple Upper Limit 1 Start X
- 64) Multiple Upper Limit 1 Start X (Lowest byte)
- 65) Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)²⁸⁷
- 66) Multiple Upper Limit 1 Start Y (Power Level)
- 67) Multiple Upper Limit 1 Start Y (Power Level)
- 68) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 69) Multiple Upper Limit 1 End X²⁸⁸ (Highest byte)
- 70) Multiple Upper Limit 1 End X
- 71) Multiple Upper Limit 1 End X
- 72) Multiple Upper Limit 1 End X (Lowest byte)
- 73) Multiple Upper Limit 1 End Y (Power Level) (Highest byte)²⁸⁹
- 74) Multiple Upper Limit 1 End Y (Power Level)
- 75) Multiple Upper Limit 1 End Y (Power Level)

²⁸¹ Scaled by Frequency Scale Factor (bytes 244-245)

²⁸² Value sent as (value in dBm * 1000) + 270,000

²⁸³ Value sent as (value * 1000)

²⁸⁴ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

²⁸⁵ Value sent as (value in dBm * 1000) + 270000

²⁸⁶ Scaled by Frequency Scale Factor (bytes 244-245)

²⁸⁷ Value sent as (value in dBm * 1000) + 270000

²⁸⁸ Scaled by Frequency Scale Factor (bytes 244-245)

²⁸⁹ Value sent as (value in dBm * 1000) + 270000

- 76) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 77-220) Multiple Upper Limits 2-5, SA Multiple Lower Limits 1-5 (see bytes 61-76 for format)
- 221) RBW Setting (Highest byte)²⁹⁰
- 222) RBW Setting
- 223) RBW Setting
- 224) RBW Setting (Lowest byte)
- 225) VBW Setting (Highest byte)²⁹¹
- 226) VBW Setting
- 227) VBW Setting
- 228) VBW Setting (Lowest byte)
- 229) Attenuation
- 230) Status Byte 1: (0b = Off , 1b = On)
- (LSB) bit 0 : Marker 1 On/Off
bit 1 : Marker 2 On/Off
bit 2 : Marker 3 On/Off
bit 3 : Marker 4 On/Off
bit 4 : Marker 5 On/Off
bit 5 : Marker 6 On/Off
bits 6 - 7 : Not Used
- 231) Status Byte 2: (0b = Off, 1b = On)
- (LSB) bit 0 : S21 Spa Cal Status (0 – Cal OFF, 1 – Cal ON)
bit 1 : Marker 2 Delta On/Off
bit 2 : Marker 3 Delta On/Off
bit 3 : Marker 4 Delta On/Off
bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
bit 5 : Pre Amp Status On/Off
bit 6 : Dynamic Attenuation On/Off
bit 7 : Not Used
- 232) Status Byte 3: (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
- (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
bit 1 : Single Limit Beep On/Off
bit 2 : Single Limit Status On/Off
bit 3 : Single Limit Beep Level ABOVE/BELOW
bit 4 : Multiple Limit Upper Segment 1 Status On/Off
bit 5 : Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW²⁹²
bit 6 : Multiple Limit Upper Segment 2 Status On/Off
bit 7 : Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW
- 233) Status Byte 4 : (0b = Off/Beep if data is BELOW line ,
1b = On/Beep if data is ABOVE line)
- (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
bit 1 : Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
bit 2 : Multiple Limit Upper Segment 4 Status On/Off
bit 3 : Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
bit 4 : Multiple Limit Upper Segment 5 Status On/Off
bit 5 : Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
bit 6 : Multiple Limit Lower Segment 1 Status On/Off
bit 7 : Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW²⁹³

²⁹⁰ RBW frequency sent in Hz.

²⁹¹ VBW frequency sent in Hz.

²⁹² Beep level is always 1b for upper segmented limit line

²⁹³ Beep level is always 0b for lower segmented limit line

- 234) Status Byte 5 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
- 235) Status Byte 6: (0b = Off, 1b = On)
 (LSB) bit 0 : Not Used
 bit 1 : Bias Tee On/Off (Option 10)
 bit 2 : External Reference Freq On/Off
 bits 3-4 : Amplitude Units (Log) - 00b = dBm 01b = dBV 10b = dBmV 11b = dBuV
 (Linear) – 00b = Watts 01b = Volts
 bits 5-6 : Detection Alg (00b = pos. peak 01b = RMS Averaging 10b = neg. peak 11b = Sampling Mode)
 bit 7 : Units Type (0b = Log 1b = Linear)
- 236) External Reference Frequency²⁹⁴
 237) Signal Standard²⁹⁵ (Higher byte)
 238) Signal Standard (Lower byte)
 239) Channel Selection²⁹⁶ (Higher byte)
 240) Channel Selection (Lower byte)
 241) Trigger Type²⁹⁷
 242) Status Byte 7
 (LSB) bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF)
 bit 7: Not Used
- 243) Status Byte 8: (0b = Off, 1b = On)
 (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B)
 bit 2: Max Hold On/Off
 bit 3: Min Hold On/Off
 bit 4: RBW Coupling (1b = Auto, 0b = Manual)
 bit 5: VBW Coupling (1b = Auto, 0b = Manual)
 bit 6: Attenuation Coupling (1b = Auto, 0b = Manual)
 bit 7: View B On/Off
- 244) Frequency Scale Factor²⁹⁸ (Higher byte)
 245) Frequency Scale Factor (Lower byte)
 246) Frequency Range Minimum²⁹⁹ (Highest byte)
 247) Frequency Range Minimum
 248) Frequency Range Minimum
 249) Frequency Range Minimum (Lowest byte)
 250) Frequency Range Maximum³⁰⁰ (Highest byte)
 251) Frequency Range Maximum
 252) Frequency Range Maximum
 253) Frequency Range Maximum (Lowest byte)
 254) Marker Type³⁰¹

²⁹⁴ 1 byte in MHz (i.e. 20 = 20MHz)

²⁹⁵ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFEEh

²⁹⁶ “No Channel” is sent as FFEEh

²⁹⁷ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

²⁹⁸ In number of Hz

²⁹⁹ Scaled by Frequency Scale Factor (bytes 244-245)

³⁰⁰ Scaled by Frequency Scale Factor (bytes 244-245)

³⁰¹ 00h = Regular Marker, 01h = Noise Marker

- 255) Trace B Trace Id³⁰²
 256) Status Byte 9
 (LSB) bit 0: Limit Beep Output (Option 86 Only) (00h = Internal Speaker, 01h = External)
 bits 1-7: Not Used
 257-400) Not Used

For Power Meter Mode (Option 29 Only):

- 21) Power Meter Start Freq³⁰³ (Highest byte)
- 22) Power Meter Start Freq
- 23) Power Meter Start Freq
- 24) Power Meter Start Freq³⁰⁴ (Lowest byte)
- 25) Power Meter Stop Freq (Highest byte)
- 26) Power Meter Stop Freq
- 27) Power Meter Stop Freq
- 28) Power Meter Stop Freq (Lowest byte)
- 29) Power Meter Center Freq³⁰⁵ (Highest byte)
- 30) Power Meter Center Freq
- 31) Power Meter Center Freq
- 32) Power Meter Center Freq (Lowest byte)
- 33) Power Meter Span³⁰⁶ (Highest byte)
- 34) Power Meter Span
- 35) Power Meter Span
- 36) Power Meter Span (Lowest byte)
- 37) Signal Standard³⁰⁷ (Higher byte)
- 38) Signal Standard (Lower byte)
- 39) Channel Selection³⁰⁸ (Higher byte)
- 40) Channel Selection (Lower byte)
- 41) Power Meter Offset³⁰⁹ (Highest byte)
- 42) Power Meter Offset
- 43) Power Meter Offset
- 44) Power Meter Offset (Lowest byte)
- 45) Power Meter Relative (Highest byte)³¹⁰
- 46) Power Meter Relative
- 47) Power Meter Relative
- 48) Power Meter Relative (Lowest byte)
- 49) Not Used
- 50) Power Meter Unit (00h = Watts, 01h = dBm)
- 51) Power Meter Relative Status (00h = Off, 01h = On)
- 52) Power Meter Offset Status (00h = Off, 01h = On)
- 53) Power Meter RMS Averaging Level (00h = Off, 01h = Low, 02h = Medium, 03h = High)
- 54) Frequency Scale Factor³¹¹ (Higher byte)
- 55) Frequency Scale Factor (Lower byte)
- 56) Frequency Range Minimum³¹² (Highest byte)
- 57) Frequency Range Minimum

³⁰² FFh indicates no trace selected

³⁰³ Scaled by Frequency Scale Factor (bytes 54-55)

³⁰⁴ Scaled by Frequency Scale Factor (bytes 54-55)

³⁰⁵ Scaled by Frequency Scale Factor (bytes 54-55)

³⁰⁶ Scaled by Frequency Scale Factor (bytes 54-55)

³⁰⁷ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

³⁰⁸ “No Channel” is sent as FFFEh

³⁰⁹ Value sent as (value in dB * 1000)

³¹⁰ Value sent as ((value in dBm * 1000) + 100)

³¹¹ In number of Hz

³¹² Scaled by Frequency Scale Factor

- 58) Frequency Range Minimum
 - 59) Frequency Range Minimum (Lowest byte)
 - 60) Frequency Range Maximum³¹³ (Highest byte)
 - 61) Frequency Range Maximum
 - 62) Frequency Range Maximum
 - 63) Frequency Range Maximum (Lowest byte)
 - 64) Zero Status (00h = Off, 01h = On)
 - 65) Zero Value³¹⁴ (Highest byte)
 - 66) Zero Value
 - 67) Zero Value
 - 68) Zero Value (Lowest byte)
 - 69-120) Not Used
-

Upload Setup – Control Byte #66 (42h)

Description: Returns parameters associated with the specified setup number. Since different modes have different numbers of setup locations available, the command requires the mode be specified as well as the setup number.

Setup numbers as follows:

- 0 = Run time setup
- 1 – 10 = Saved setups for Spectrum Analyzer/Transmission Measurement modes
- 1 – 5 = Saved setups for Power Meter mode (Option 29 Only)

Bytes to Follow: 2 bytes

For All Modes:

- 1) Number of Following Bytes (Higher byte)
- 2) Number of Following Bytes (Lower byte)
- 3) Measurement Mode³¹⁵
- 4) Setup Number in which to store setup
- 5-20) Not Used

For Spectrum Analyzer Mode:

- 21) Spectrum Analyzer Start Frequency³¹⁶ (Highest byte)
- 22) Spectrum Analyzer Start Frequency
- 23) Spectrum Analyzer Start Frequency
- 24) Spectrum Analyzer Start Frequency (Lowest byte)
- 25) Spectrum Analyzer Stop Frequency³¹⁷ (Highest byte)
- 26) Spectrum Analyzer Stop Frequency
- 27) Spectrum Analyzer Stop Frequency
- 28) Spectrum Analyzer Stop Frequency (Lowest byte)
- 29) Spectrum Analyzer Center Frequency³¹⁸ (Highest byte)
- 30) Spectrum Analyzer Center Frequency
- 31) Spectrum Analyzer Center Frequency
- 32) Spectrum Analyzer Center Frequency (Lowest byte)
- 33) Spectrum Analyzer Frequency Span³¹⁹ (Highest byte)
- 34) Spectrum Analyzer Frequency Span
- 35) Spectrum Analyzer Frequency Span
- 36) Spectrum Analyzer Frequency Span (Lowest byte)

³¹³ Scaled by Frequency Scale Factor

³¹⁴ Value sent as ((value in dBm * 1000) + 100)

³¹⁵ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

³¹⁶ Scaled by Frequency Scale Factor (bytes 301-302)

³¹⁷ Scaled by Frequency Scale Factor (bytes 301-302)

³¹⁸ Scaled by Frequency Scale Factor (bytes 301-302)

³¹⁹ Scaled by Frequency Scale Factor (bytes 301-302)

- 37) Ref Level (Highest byte)³²⁰
- 38) Ref Level
- 39) Ref Level
- 40) Ref Level (Lowest byte)
- 41) Scale per div (Highest byte)³²¹
- 42) Scale per div
- 43) Scale per div
- 44) Scale per div (Lowest byte)
- 45) Spectrum Analyzer Frequency Marker 1 (Higher byte)³²²
- 46) Spectrum Analyzer Frequency Marker 1 (Lower byte)
- 47) Spectrum Analyzer Frequency Marker 2 (Higher byte)
- 48) Spectrum Analyzer Frequency Marker 2 (Lower byte)
- 49) Spectrum Analyzer Frequency Marker 3 (Higher byte)
- 50) Spectrum Analyzer Frequency Marker 3 (Lower byte)
- 51) Spectrum Analyzer Frequency Marker 4 (Higher byte)
- 52) Spectrum Analyzer Frequency Marker 4 (Lower byte)
- 53) Spectrum Analyzer Frequency Marker 5 (Higher byte)
- 54) Spectrum Analyzer Frequency Marker 5 (Lower byte)
- 55) Spectrum Analyzer Frequency Marker 6 (Higher byte)
- 56) Spectrum Analyzer Frequency Marker 6 (Lower byte)
- 57) Spectrum Analyzer Single Limit (Highest byte)³²³
- 58) Spectrum Analyzer Single Limit
- 59) Spectrum Analyzer Single Limit
- 60) Spectrum Analyzer Single Limit (Lowest byte)
- 61) SPA Multiple Upper Limit 1 Start X³²⁴ (Highest byte)
- 62) SPA Multiple Upper Limit 1 Start X
- 63) SPA Multiple Upper Limit 1 Start X
- 64) SPA Multiple Upper Limit 1 Start X (Lowest byte)
- 65) SPA Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)³²⁵
- 66) SPA Multiple Upper Limit 1 Start Y (Power Level)
- 67) SPA Multiple Upper Limit 1 Start Y (Power Level)
- 68) SPA Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 69) SPA Multiple Upper Limit 1 End X³²⁶ (Highest byte)
- 70) SPA Multiple Upper Limit 1 End X
- 71) SPA Multiple Upper Limit 1 End X
- 72) SPA Multiple Upper Limit 1 End X (Lowest byte)
- 73) SPA Multiple Upper Limit 1 End Y (Power Level) (Highest byte)³²⁷
- 74) SPA Multiple Upper Limit 1 End Y (Power Level)
- 75) SPA Multiple Upper Limit 1 End Y (Power Level)
- 76) SPA Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 77-220) SPA Multiple Upper Limits 2-5, SA Multiple Lower Limits 1-5 (see bytes 61-76 for format)
- 221) RBW Setting (Highest byte)³²⁸
- 222) RBW Setting
- 223) RBW Setting
- 224) RBW Setting (Lowest byte)

³²⁰ Value sent as (value in dBm * 1000) + 270,000

³²¹ Value sent as (value * 1000)

³²² Value sent as data point on the display. Equivalent frequency = (point * span / (# data points - 1)) + start frequency.

³²³ Value sent as (value in dBm * 1000) + 270000

³²⁴ Scaled by Frequency Scale Factor (bytes 301-302)

³²⁵ Value sent as (value in dBm * 1000) + 270000

³²⁶ Scaled by Frequency Scale Factor (bytes 301-302)

³²⁷ Value sent as (value in dBm * 1000) + 270000

³²⁸ RBW frequency sent in Hz.

225)	VBW Setting (Highest byte) ³²⁹
226)	VBW Setting
227)	VBW Setting
228)	VBW Setting (Lowest byte)
229)	OCC BW Method ³³⁰
230)	OCC BW % Value ³³¹
231)	OCC BW dBc ³³²
232)	Attenuation
233)	Antenna Index (0-14)
234-249)	Antenna Name (16 bytes in ASCII)
250)	Status Byte 1: (0b = Off , 1b = On)
	(LSB) bit 0 : Spectrum Analyzer Mode Marker 1 On/Off
	bit 1 : Spectrum Analyzer Mode Marker 2 On/Off
	bit 2 : Spectrum Analyzer Mode Marker 3 On/Off
	bit 3 : Spectrum Analyzer Mode Marker 4 On/Off
	bit 4 : Spectrum Analyzer Mode Marker 5 On/Off
	bit 5 : Spectrum Analyzer Mode Marker 6 On/Off
	bits 6 - 7 : Not Used
251)	Status Byte 2: (0b = Off, 1b = On)
	(LSB) bit 0 : Not Used
	bit 1 : Spectrum Analyzer Mode Marker 2 Delta On/Off
	bit 2 : Spectrum Analyzer Mode Marker 3 Delta On/Off
	bit 3 : Spectrum Analyzer Mode Marker 4 Delta On/Off
	bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
	bit 5 : Pre Amp Status On/Off
	bit 6 : Dynamic Attenuation On/Off
	bit 7 : Normalization On/Off
252)	Status Byte 3: (0b = Off/Beep if data is BELOW line , 1b = On/Beep if data is ABOVE line)
	(LSB) bit 0 : SPA Limit Type (0b = Single, 1b = Multiple)
	bit 1 : SPA Single Limit Beep On/Off
	bit 2 : SPA Single Limit Status On/Off
	bit 3 : SPA Single Limit Beep Level ABOVE/BELOW
	bit 4 : SPA Multiple Limit Upper Segment 1 Status On/Off
	bit 5 : SPA Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW ³³³
	bit 6 : SPA Multiple Limit Upper Segment 2 Status On/Off
	bit 7 : SPA Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW
253)	Status Byte 4 : (0b = Off/Beep if data is BELOW line , 1b = On/Beep if data is ABOVE line)
	(LSB) bit 0 : SPA Multiple Limit Upper Segment 3 Status On/Off
	bit 1 : SPA Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
	bit 2 : SPA Multiple Limit Upper Segment 4 Status On/Off
	bit 3 : SPA Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
	bit 4 : SPA Multiple Limit Upper Segment 5 Status On/Off
	bit 5 : SPA Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
	bit 6 : SPA Multiple Limit Lower Segment 1 Status On/Off
	bit 7 : SPA Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW ³³⁴

³²⁹ VBW frequency sent in Hz.

³³⁰ 00h = % of power, 01h = dB down

³³¹ 0 – 99%

³³² 0 – 120 dBc

³³³ Beep level is always 1b for upper segmented limit line

³³⁴ Beep level is always 0b for lower segmented limit line

- 254) Status Byte 5 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
 (LSB) bit 0 : SPA Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : SPA Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 bit 2 : SPA Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : SPA Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 bit 4 : SPA Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : SPA Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 bit 6 : SPA Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : SPA Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW
- 255) Status Byte 6: (0b = Off, 1b = On)
 (LSB) bit 0 : Antenna Factors Correction On/Off
 bit 1 : Bias Tee On/Off (Option 10)
 bit 2 : Amplitude Units (Linear) – 00b = Watts 01b = Volts
 bits 3-4 : Amplitude Units (Log) - 00b = dBm 01b = dBV 10b = dBmV 11b = dBuV
 bits 5-6 : Detection Alg (00b = pos. peak 01b = RMS Averaging 10b = neg. peak 11b = Sampling Mode)
 bit 7 : Units Type (0b = Log 1b = Linear)
- 256) Status Byte 7: (0b = Off, 1b = On)
 (LSB) bit 0: Interference Analysis On/Off
 bit 1: C/I Measurement On/Off
 bit 2: RBW Coupling (1b = Auto, 0b = Manual)
 bit 3: VBW Coupling (1b = Auto, 0b = Manual)
 bit 4: Attenuation Coupling (1b = Auto, 0b = Manual)
 bit 5: Channel Power On/Off
 bit 6: Adjacent Channel Power On/Off
 bit 7: Occupied BW Measurement On/Off
- 257) Reference Level Offset³³⁵ (Highest byte)
 258) Reference Level Offset
 259) Reference Level Offset
 260) Reference Level Offset (Lowest byte)
 261) External Reference Frequency³³⁶
 262) Signal Standard³³⁷ (Higher byte)
 263) Signal Standard (Lower byte)
 264) Channel Selection³³⁸ (Higher byte)
 265) Channel Selection (Lower byte)
 266) Trigger Type³³⁹
 267) Interference Analysis Frequency³⁴⁰ (Highest byte)
 268) Interference Analysis Frequency
 269) Interference Analysis Frequency
 270) Interference Analysis Frequency (Lowest byte)
 271) Trigger Position (0 – 100%)
 272) Min Sweep Time (in μ s) (Highest byte)
 273) Min Sweep Time (in μ s)
 274) Min Sweep Time (in μ s)
 275) Min Sweep Time (in μ s) (Lowest byte)
 276) Video Trigger Level³⁴¹ (Highest byte)
 277) Video Trigger Level

³³⁵ Value sent as (value in dBm * 1000) + 270,000

³³⁶ 1 byte in MHz (i.e. 20 = 20MHz)

³³⁷ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

³³⁸ “No Channel” is sent as FFFEh

³³⁹ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

³⁴⁰ Scaled by Frequency Scale Factor (bytes 301-302)

³⁴¹ Value sent as (value in dBm * 1000) + 270,000

278)	Video Trigger Level
279)	Video Trigger Level (Lowest byte)
280)	Status Byte 8 (LSB) bit 0: Limit Beep Output (Option 86 Only) (00h = Internal Speaker, 01h = External) bits 1-7: Not Used
281)	Status Byte 9 (LSB) bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF) bit 7: Not Used
282)	Status Byte 10: (0b = Off, 1b = On) (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B) bit 2: Max Hold On/Off bit 3: Min Hold On/Off bit 4: View B On/Off bit 5: External Reference Frequency On/Off bits 6-7: Not Used
283)	Impedance (00h = 50Ω , 10h = 75Ω Anritsu Adapter, 12h = 75Ω Other Adapter)
284)	Impedance Loss ³⁴² (Higher byte)
285)	Impedance Loss (Lower byte)
286)	AM/FM Demod Type ³⁴³
287)	AM/FM Demod Status (01h = On, 00h = Off)
288)	AM/FM Demod Volume (0 to 100)
289)	AM/FM Demod Frequency ³⁴⁴ (Highest byte)
290)	AM/FM Demod Frequency
291)	AM/FM Demod Frequency
292)	AM/FM Demod Frequency (Lowest byte)
293)	AM/FM Demod Time (in ms) (Highest byte)
294)	AM/FM Demod Time (in ms)
295)	AM/FM Demod Time (in ms)
296)	AM/FM Demod Time (in ms) (Lowest byte)
297)	SSB BFO Offset ³⁴⁵ (Highest byte)
298)	SSB BFO Offset
299)	SSB BFO Offset
300)	SSB BFO Offset (Lowest byte)
301)	Frequency Scale Factor ³⁴⁶ (Higher byte)
302)	Frequency Scale Factor (Lower byte)
303)	Frequency Range Minimum ³⁴⁷ (Highest byte)
304)	Frequency Range Minimum
305)	Frequency Range Minimum
306)	Frequency Range Minimum (Lowest byte)
307)	Frequency Range Maximum ³⁴⁸ (Highest byte)
308)	Frequency Range Maximum
309)	Frequency Range Maximum
310)	Frequency Range Maximum (Lowest byte)
311)	Marker Type ³⁴⁹
312)	Channel Power Int BW ³⁵⁰ (Highest byte)

³⁴² Value sent as (value in dB * 1000), valid values are 0 to 20 dB

³⁴³ AM/FM Demod Type: 00h = FM-Wide Band, 01h = FM-Narrow Band, 02h = AM, 03h = SSB Lower, 04h = SSB Upper

³⁴⁴ Scaled by Frequency Scale Factor (bytes 301-302)

³⁴⁵ Value sent as ((value in Hz) – 10,000)

³⁴⁶ In number of Hz

³⁴⁷ Scaled by Frequency Scale Factor (bytes 301-302)

³⁴⁸ Scaled by Frequency Scale Factor (bytes 301-302)

³⁴⁹ 00h = Regular Marker, 01h = Noise Marker

³⁵⁰ Scaled by Frequency Scale Factor (bytes 301-302)

- 313) Channel Power Int BW
- 314) Channel Power Int BW
- 315) Channel Power Int BW (Lowest byte)
- 316) ACPR Main Channel BW³⁵¹ (Highest byte)
- 317) ACPR Main Channel BW
- 318) ACPR Main Channel BW
- 319) ACPR Main Channel BW (Lowest byte)
- 320) ACPR Adjacent Channel BW³⁵² (Highest byte)
- 321) ACPR Adjacent Channel BW
- 322) ACPR Adjacent Channel BW
- 323) ACPR Adjacent Channel BW (Lowest byte)
- 324) ACPR Channel Spacing³⁵³ (Highest byte)
- 325) ACPR Channel Spacing
- 326) ACPR Channel Spacing
- 327) ACPR Channel Spacing (Lowest byte)
- 328) Interference Analysis Cell Std³⁵⁴
- 329) Interference Analysis Est. BW³⁵⁵ (Highest byte)
- 330) Interference Analysis Est. BW
- 331) Interference Analysis Est. BW
- 332) Interference Analysis Est. BW (Lowest byte)
- 333) Trace B Trace Id³⁵⁶
- 334-500) Not Used

For Transmission Mode (Option 21 Only):

- 21) Start Frequency³⁵⁷ (Highest byte)
- 22) Start Frequency
- 23) Start Frequency
- 24) Start Frequency (Lowest byte)
- 25) Stop Frequency³⁵⁸ (Highest byte)
- 26) Stop Frequency
- 27) Stop Frequency
- 28) Stop Frequency (Lowest byte)
- 29) Center Frequency³⁵⁹ (Highest byte)
- 30) Center Frequency
- 31) Center Frequency
- 32) Center Frequency (Lowest byte)
- 33) Frequency Span³⁶⁰ (Highest byte)
- 34) Frequency Span
- 35) Frequency Span
- 36) Frequency Span (Lowest byte)
- 37) Ref Level (Highest byte)³⁶¹
- 38) Ref Level
- 39) Ref Level

³⁵¹ Scaled by Frequency Scale Factor (bytes 301-302)

³⁵² Scaled by Frequency Scale Factor (bytes 301-302)

³⁵³ Scaled by Frequency Scale Factor (bytes 301-302)

³⁵⁴ 4 Standards – 00h = 1250 kHz CDMA, 01h = GSM, 02h = TDMA, 03h = AMPS, 04h = Unknown, FFh = Interference Analysis Measurement OFF

³⁵⁵ Frequency in Hz

³⁵⁶ FFh indicates to trace selected

³⁵⁷ Scaled by Frequency Scale Factor (bytes 244-245)

³⁵⁸ Scaled by Frequency Scale Factor (bytes 244-245)

³⁵⁹ Scaled by Frequency Scale Factor (bytes 244-245)

³⁶⁰ Scaled by Frequency Scale Factor (bytes 244-245)

³⁶¹ Value sent as (value in dBm * 1000) + 270,000)

- 40) Ref Level (Lowest byte)
- 41) Scale per div (Highest byte)³⁶²
- 42) Scale per div
- 43) Scale per div
- 44) Scale per div (Lowest byte)
- 45) Frequency Marker 1 (Higher byte)³⁶³
- 46) Frequency Marker 1 (Lower byte)
- 47) Frequency Marker 2 (Higher byte)
- 48) Frequency Marker 2 (Lower byte)
- 49) Frequency Marker 3 (Higher byte)
- 50) Frequency Marker 3 (Lower byte)
- 51) Frequency Marker 4 (Higher byte)
- 52) Frequency Marker 4 (Lower byte)
- 53) Frequency Marker 5 (Higher byte)
- 54) Frequency Marker 5 (Lower byte)
- 55) Frequency Marker 6 (Higher byte)
- 56) Frequency Marker 6 (Lower byte)
- 57) Single Limit (Highest byte)³⁶⁴
- 58) Single Limit
- 59) Single Limit
- 60) Single Limit (Lowest byte)
- 61) Multiple Upper Limit 1 Start X³⁶⁵ (Highest byte)
- 62) Multiple Upper Limit 1 Start X
- 63) Multiple Upper Limit 1 Start X
- 64) Multiple Upper Limit 1 Start X (Lowest byte)
- 65) Multiple Upper Limit 1 Start Y (Power Level) (Highest byte)³⁶⁶
- 66) Multiple Upper Limit 1 Start Y (Power Level)
- 67) Multiple Upper Limit 1 Start Y (Power Level)
- 68) Multiple Upper Limit 1 Start Y (Power Level) (Lowest byte)
- 69) Multiple Upper Limit 1 End X³⁶⁷ (Highest byte)
- 70) Multiple Upper Limit 1 End X
- 71) Multiple Upper Limit 1 End X
- 72) Multiple Upper Limit 1 End X (Lowest byte)
- 73) Multiple Upper Limit 1 End Y (Power Level) (Highest byte)³⁶⁸
- 74) Multiple Upper Limit 1 End Y (Power Level)
- 75) Multiple Upper Limit 1 End Y (Power Level)
- 76) Multiple Upper Limit 1 End Y (Power Level) (Lowest byte)
- 77-220) Multiple Upper Limits 2-5, SA Multiple Lower Limits 1-5 (see bytes 67-82 for format)
- 221) RBW Setting (Highest byte)³⁶⁹
- 222) RBW Setting
- 223) RBW Setting
- 224) RBW Setting (Lowest byte)
- 225) VBW Setting (Highest byte)³⁷⁰
- 226) VBW Setting
- 227) VBW Setting

³⁶² Value sent as (value * 1000)

³⁶³ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

³⁶⁴ Value sent as (value in dBm * 1000) + 270000

³⁶⁵ Scaled by Frequency Scale Factor (bytes 244-245)

³⁶⁶ Value sent as (value in dBm * 1000) + 270000

³⁶⁷ Scaled by Frequency Scale Factor (bytes 244-245)

³⁶⁸ Value sent as (value in dBm * 1000) + 270000

³⁶⁹ RBW frequency sent in Hz.

³⁷⁰ VBW frequency sent in Hz.

- 228) VBW Setting (Lowest byte)
- 229) Attenuation
- 230) Status Byte 1: (0b = Off , 1b = On)
- (LSB) bit 0 : Marker 1 On/Off
 bit 1 : Marker 2 On/Off
 bit 2 : Marker 3 On/Off
 bit 3 : Marker 4 On/Off
 bit 4 : Marker 5 On/Off
 bit 5 : Marker 6 On/Off
 bits 6 - 7 : Not Used
- 231) Status Byte 2: (0b = Off, 1b = On)
- (LSB) bit 0 : S21 Spa Cal Status (0 – Cal OFF, 1 – Cal ON)
 bit 1 : Marker 2 Delta On/Off
 bit 2 : Marker 3 Delta On/Off
 bit 3 : Marker 4 Delta On/Off
 bit 4 : Pre Amp Mode (0b = Manual, 1b = Auto)
 bit 5 : Pre Amp Status On/Off
 bit 6 : Dynamic Attenuation On/Off
 bit 7 : Not Used
- 232) Status Byte 3: (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
- (LSB) bit 0 : Limit Type (0b = Single, 1b = Multiple)
 bit 1 : Single Limit Beep On/Off
 bit 2 : Single Limit Status On/Off
 bit 3 : Single Limit Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Upper Segment 1 Status On/Off
 bit 5 : Multiple Limit Upper Segment 1 Beep Level ABOVE/BELOW³⁷¹
 bit 6 : Multiple Limit Upper Segment 2 Status On/Off
 bit 7 : Multiple Limit Upper Segment 2 Beep Level ABOVE/BELOW
- 233) Status Byte 4 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
- (LSB) bit 0 : Multiple Limit Upper Segment 3 Status On/Off
 bit 1 : Multiple Limit Upper Segment 3 Beep Level ABOVE/BELOW
 bit 2 : Multiple Limit Upper Segment 4 Status On/Off
 bit 3 : Multiple Limit Upper Segment 4 Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Upper Segment 5 Status On/Off
 bit 5 : Multiple Limit Upper Segment 5 Beep Level ABOVE/BELOW
 bit 6 : Multiple Limit Lower Segment 1 Status On/Off
 bit 7 : Multiple Limit Lower Segment 1 Beep Level ABOVE/BELOW³⁷²
- 234) Status Byte 5 : (0b = Off/Beep if data is BELOW line ,
 1b = On/Beep if data is ABOVE line)
- (LSB) bit 0 : Multiple Limit Lower Segment 2 Status On/Off
 bit 1 : Multiple Limit Lower Segment 2 Beep Level ABOVE/BELOW
 bit 2 : Multiple Limit Lower Segment 3 Status On/Off
 bit 3 : Multiple Limit Lower Segment 3 Beep Level ABOVE/BELOW
 bit 4 : Multiple Limit Lower Segment 4 Status On/Off
 bit 5 : Multiple Limit Lower Segment 4 Beep Level ABOVE/BELOW
 bit 6 : Multiple Limit Lower Segment 5 Status On/Off
 bit 7 : Multiple Limit Lower Segment 5 Beep Level ABOVE/BELOW

³⁷¹ Beep level is always 1b for upper segmented limit line

³⁷² Beep level is always 0b for lower segmented limit line

- 235) Status Byte 6: (0b = Off, 1b = On)
 (LSB) bit 0 : External Reference Frequency On/Off
 bit 1 : Bias Tee On/Off (Option 10)
 bit 2 : Amplitude Units (Linear) – 00b = Watts 01b = Volts
 bits 3-4 : Amplitude Units (Log) - 00b = dBm 01b = dBV 10b = dBmV 11b = dBuV
 bits 5-6 : Detection Alg (00b = pos. peak 01b = RMS Averaging 10b = neg. peak 11b = Sampling Mode)
 bit 7 : Units Type (0b = Log 1b = Linear)
- 236) External Reference Frequency³⁷³
 237) Signal Standard³⁷⁴ (Higher byte)
 238) Signal Standard (Lower byte)
 239) Channel Selection³⁷⁵ (Higher byte)
 240) Channel Selection (Lower byte)
 241) Trigger Type³⁷⁶
 242) Status Byte 7
 (LSB) bits 0-6: Number of sweeps to average (1-25, 1 implies averaging OFF)
 bit 7: Not Used
- 243) Status Byte 8: (0b = Off, 1b = On)
 (LSB) bits 0-1: Trace Math Operation (00b = A only, 01b = A-B, 10b = A+B)
 bit 2: Max Hold On/Off
 bit 3: Min Hold On/Off
 bit 4: RBW Coupling (1b = Auto, 0b = Manual)
 bit 5: VBW Coupling (1b = Auto, 0b = Manual)
 bit 6: Attenuation Coupling (1b = Auto, 0b = Manual)
 bit 7: View B On/Off
- 244) Frequency Scale Factor³⁷⁷ (Higher byte)
 245) Frequency Scale Factor (Lower byte)
 246) Frequency Range Minimum³⁷⁸ (Highest byte)
 247) Frequency Range Minimum
 248) Frequency Range Minimum
 249) Frequency Range Minimum (Lowest byte)
 250) Frequency Range Maximum³⁷⁹ (Highest byte)
 251) Frequency Range Maximum
 252) Frequency Range Maximum
 253) Frequency Range Maximum (Lowest byte)
 254) Marker Type³⁸⁰
 255) Trace B Trace Id³⁸¹
 256) Status Byte 9
 (LSB) bit 0: Limit Beep Output (Option 86 Only) (00h = Internal Speaker, 01h = External)
 bits 1-7: Not Used
- 257-400) Not Used

³⁷³ 1 byte in MHz (i.e. 20 = 20MHz)

³⁷⁴ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFFEh

³⁷⁵ “No Channel” is sent as FFFEh

³⁷⁶ Trigger Type – 00h = Single, 01h = Free Run, 02h = Video, 03h = External

³⁷⁷ In number of Hz

³⁷⁸ Scaled by Frequency Scale Factor (bytes 244-245)

³⁷⁹ Scaled by Frequency Scale Factor (bytes 244-245)

³⁸⁰ 00h = Regular Marker, 01h = Noise Marker

³⁸¹ FFh indicates no trace selected

For Power Meter Mode (Option 29 Only):

- 21) Power Meter Start Freq³⁸² (Highest byte)
- 22) Power Meter Start Freq
- 23) Power Meter Start Freq
- 24) Power Meter Start Freq³⁸³ (Lowest byte)
- 25) Power Meter Stop Freq (Highest byte)
- 26) Power Meter Stop Freq
- 27) Power Meter Stop Freq
- 28) Power Meter Stop Freq (Lowest byte)
- 29) Power Meter Center Freq³⁸⁴ (Highest byte)
- 30) Power Meter Center Freq
- 31) Power Meter Center Freq
- 32) Power Meter Center Freq (Lowest byte)
- 33) Power Meter Span³⁸⁵ (Highest byte)
- 34) Power Meter Span
- 35) Power Meter Span
- 36) Power Meter Span (Lowest byte)
- 37) Signal Standard³⁸⁶ (Higher byte)
- 38) Signal Standard (Lower byte)
- 39) Channel Selection³⁸⁷ (Higher byte)
- 40) Channel Selection (Lower byte)
- 41) Power Meter Offset³⁸⁸ (Highest byte)
- 42) Power Meter Offset
- 43) Power Meter Offset
- 44) Power Meter Offset (Lowest byte)
- 45) Power Meter Relative (Highest byte)³⁸⁹
- 46) Power Meter Relative
- 47) Power Meter Relative
- 48) Power Meter Relative (Lowest byte)
- 49) Not Used
- 50) Power Meter Unit (00h = Watts, 01h = dBm)
- 51) Power Meter Relative Status (00h = Off, 01h = On)
- 52) Power Meter Offset Status (00h = Off, 01h = On)
- 53) Power Meter RMS Averaging Level (00h = Off, 01h = Low, 02h = Medium, 03h = High)
- 54) Frequency Scale Factor³⁹⁰ (Higher byte)
- 55) Frequency Scale Factor (Lower byte)
- 56) Frequency Range Minimum³⁹¹ (Highest byte)
- 57) Frequency Range Minimum
- 58) Frequency Range Minimum
- 59) Frequency Range Minimum (Lowest byte)
- 60) Frequency Range Maximum³⁹² (Highest byte)
- 61) Frequency Range Maximum
- 62) Frequency Range Maximum
- 63) Frequency Range Maximum (Lowest byte)

³⁸² Scaled by Frequency Scale Factor (bytes 54-55)

³⁸³ Scaled by Frequency Scale Factor (bytes 54-55)

³⁸⁴ Scaled by Frequency Scale Factor (bytes 54-55)

³⁸⁵ Scaled by Frequency Scale Factor (bytes 54-55)

³⁸⁶ Index into Standard List (use control byte #89 to retrieve the ASCII string name). “No Standard” sent as FFEEh

³⁸⁷ “No Channel” is sent as FFEEh

³⁸⁸ Value sent as (value in dB * 1000)

³⁸⁹ Value sent as ((value in dBm * 1000) + 100)

³⁹⁰ In number of Hz

³⁹¹ Scaled by Frequency Scale Factor

³⁹² Scaled by Frequency Scale Factor

- 64) Zero Status (00h = Off, 01h = On)
- 65) Zero Value³⁹³ (Highest byte)
- 66) Zero Value
- 67) Zero Value
- 68) Zero Value (Lowest byte)
- 69-120) Not Used

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Read Parameter Limits – Control Byte #67 (43h)

Description: Returns limits (minimum and maximum values) associated with each parameter defined for the specified measurement mode.

“Frequency Parameters” are start, stop, and center frequencies, multiple limit “x” parameters and AM/FM demod frequency parameters.

Send this command twice. The first time, send it for the frequency parameter limits then send it for all other parameter limits.

Bytes to Follow: 2 bytes

- 1) Measurement Mode³⁹⁴
- 2) Limits to Read (00h = Frequency Parameter Limits, 01h = Available RBWs, 02h = Available VBWs, FFh = All Other Parameter Limits)

Spectrum Master Returns:

For All Modes:

- 1) Number of Following Bytes (Higher byte)
- 2) Number of Following Bytes (Lower byte)
- 3) Measurement Mode³⁹⁵
- 5-20) Not Used

For All Modes, Frequency Parameter Limits:

- 21) Number of Valid Frequency Ranges

For each range:

- 1) Range Scale Factor³⁹⁶ (Higher byte)
- 2) Range Scale Factor (Lower byte)
- 3) Range Start Frequency³⁹⁷ (Highest byte)
- 4) Range Start Frequency
- 5) Range Start Frequency
- 6) Range Start Frequency (Lowest byte)
- 7) Range Stop Frequency³⁹⁸ (Highest byte)
- 8) Range Stop Frequency
- 9) Range Stop Frequency
- 10) Range Stop Frequency (Lowest byte)

³⁹³ Value sent as ((value in dBm * 1000) + 100)

³⁹⁴ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

³⁹⁵ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

³⁹⁶ Scale Factor in number of Hz

³⁹⁷ Scaled by Span Scale Factor

³⁹⁸ Scaled by Span Scale Factor

For Spectrum Analyzer, Transmission Measurement Modes, Available RBWS:

- 21) Number of Valid RBWs

For each RBW:

- 1) RBW Frequency (in Hz) (Highest byte)
- 2) RBW Frequency (in Hz)
- 3) RBW Frequency (in Hz)
- 4) RBW Frequency (in Hz) (Lowest byte)

For Spectrum Analyzer, Transmission Measurement Modes, Available VBWS:

- 21) Number of Valid VBWs

For each VBW:

- 1) VBW Frequency (in Hz) (Highest byte)
- 2) VBW Frequency (in Hz)
- 3) VBW Frequency (in Hz)
- 4) VBW Frequency (in Hz) (Lowest byte)

For Spectrum Analyzer Mode, All Other Parameter Limits:

- 21) Frequency Scale Factor Minimum³⁹⁹ (Higher byte)
- 22) Frequency Scale Factor Minimum (Lower byte)
- 23) Frequency Scale Factor Maximum⁴⁰⁰ (Higher byte)
- 24) Frequency Scale Factor Maximum (Lower byte)
- 25) Span Minimum⁴⁰¹ (Highest byte)
- 26) Span Minimum
- 27) Span Minimum
- 28) Span Minimum (Lowest byte)
- 29) Span Maximum⁴⁰² (Highest byte)
- 30) Span Maximum
- 31) Span Maximum
- 32) Span Maximum (Lowest byte)
- 33) Reference Level Minimum⁴⁰³ (Highest byte)
- 34) Reference Level Minimum
- 35) Reference Level Minimum
- 36) Reference Level Minimum (Lowest byte)
- 37) Reference Level Maximum⁴⁰⁴ (Highest byte)
- 38) Reference Level Maximum
- 39) Reference Level Maximum
- 40) Reference Level Maximum (Lowest byte)
- 41) Scale Minimum⁴⁰⁵ (Highest byte)
- 42) Scale Minimum
- 43) Scale Minimum
- 44) Scale Minimum (Lowest byte)
- 45) Scale Maximum⁴⁰⁶ (Highest byte)
- 46) Scale Maximum
- 47) Scale Maximum

³⁹⁹ Scale Factor in number of Hz

⁴⁰⁰ Scale Factor in number of Hz

⁴⁰¹ Scaled by Span Scale Factor

⁴⁰² Scaled by Span Scale Factor

⁴⁰³ Value sent as (value * 1000) + 270,000

⁴⁰⁴ Value sent as (value * 1000) + 270,000

⁴⁰⁵ Value sent as (value * 1000)

⁴⁰⁶ Value sent as (value * 1000)

- 48) Scale Maximum (Lowest byte)
- 49) Marker Minimum⁴⁰⁷ (Higher byte)
- 50) Marker Minimum (Lower byte)
- 51) Marker Maximum⁴⁰⁸ (Higher byte)
- 52) Marker Maximum (Lower byte)
- 53) Limit Y Minimum⁴⁰⁹ (Highest byte)
- 54) Limit Y Minimum
- 55) Limit Y Minimum
- 56) Limit Y Minimum (Lowest byte)
- 57) Limit Y Maximum⁴¹⁰ (Highest byte)
- 58) Limit Y Maximum
- 59) Limit Y Maximum
- 60) Limit Y Maximum (Lowest byte)
- 61) OBW Method Minimum
- 62) OBW Method Maximum
- 63) OBW % of Power Minimum
- 64) OBW % of Power Maximum
- 65) OBW dBc Minimum
- 66) OBW dBc Maximum
- 67) Attenuation Minimum
- 68) Attenuation Maximum
- 69) Amplitude Units Minimum
- 70) Amplitude Units Maximum
- 71) Detection Algorithm Minimum
- 72) Detection Algorithm Maximum
- 73) RL Offset Minimum⁴¹¹ (Highest byte)
- 74) RL Offset Minimum
- 75) RL Offset Minimum
- 76) RL Offset Minimum (Lowest byte)
- 77) RL Offset Maximum⁴¹² (Highest byte)
- 78) RL Offset Maximum
- 79) RL Offset Maximum
- 80) RL Offset Maximum (Lowest byte)
- 81) External Reference Frequency Minimum⁴¹³ (Highest byte)
- 82) External Reference Frequency Minimum
- 83) External Reference Frequency Minimum
- 84) External Reference Frequency Minimum (Lowest byte)
- 85) External Reference Frequency Maximum⁴¹⁴ (Highest byte)
- 86) External Reference Frequency Maximum
- 87) External Reference Frequency Maximum
- 88) External Reference Frequency Maximum (Lowest byte)
- 89) Trigger Type Minimum
- 90) Trigger Type Maximum
- 91) Minimum Sweep Type (in μ s) Minimum (Highest byte)
- 92) Minimum Sweep Type (in μ s) Minimum

⁴⁰⁷ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

⁴⁰⁸ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

⁴⁰⁹ Value sent as (value * 1000) + 270,000

⁴¹⁰ Value sent as (value * 1000) + 270,000

⁴¹¹ Value sent as (value * 1000) + 270,000

⁴¹² Value sent as (value * 1000) + 270,000

⁴¹³ Reference frequency in Hz

⁴¹⁴ Reference frequency in Hz

- 93) Minimum Sweep Type (in μ s) Minimum
- 94) Minimum Sweep Type (in μ s) Minimum (Lowest byte)
- 95) Minimum Sweep Type (in μ s) Maximum (Highest byte)
- 96) Minimum Sweep Type (in μ s) Maximum
- 97) Minimum Sweep Type (in μ s) Maximum
- 98) Minimum Sweep Type (in μ s) Maximum (Lowest byte)
- 99) Video Trigger Level Minimum⁴¹⁵ (Highest byte)
- 100) Video Trigger Level Minimum
- 101) Video Trigger Level Minimum
- 102) Video Trigger Level Minimum (Lowest byte)
- 103) Video Trigger Level Maximum⁴¹⁶ (Highest byte)
- 104) Video Trigger Level Maximum
- 105) Video Trigger Level Maximum
- 106) Video Trigger Level Maximum (Lowest byte)
- 107) Sweep Average Minimum
- 108) Sweep Average Maximum
- 109) Trace Math Minimum
- 110) Trace Math Maximum
- 111) Impedance Loss Minimum⁴¹⁷ (Highest byte)
- 112) Impedance Loss Minimum
- 113) Impedance Loss Minimum
- 114) Impedance Loss Minimum (Lowest byte)
- 115) Impedance Loss Maximum⁴¹⁸ (Highest byte)
- 116) Impedance Loss Maximum
- 117) Impedance Loss Maximum
- 118) Impedance Loss Maximum (Lowest byte)
- 119) Demod Type Minimum
- 120) Demod Type Maximum
- 121) Demod Volume Minimum
- 122) Demod Volume Maximum
- 123) Demod Time Minimum (in ms) (Highest byte)
- 124) Demod Time Minimum (in ms)
- 125) Demod Time Minimum (in ms)
- 126) Demod Time Minimum (in ms) (Lowest byte)
- 127) Demod Time Maximum (in ms) (Highest byte)
- 128) Demod Time Maximum (in ms)
- 129) Demod Time Maximum (in ms)
- 130) Demod Time Maximum (in ms) (Lowest byte)
- 131) SSB BFO Offset Minimum⁴¹⁹ (Highest byte)
- 132) SSB BFO Offset Minimum
- 133) SSB BFO Offset Minimum
- 134) SSB BFO Offset Minimum (Lowest byte)
- 135) SSB BFO Offset Maximum⁴²⁰ (Highest byte)
- 136) SSB BFO Offset Maximum
- 137) SSB BFO Offset Maximum
- 138) SSB BFO Offset Maximum (Lowest byte)
- 139) ACPR Main Channel BW Minimum (in Hz) (Highest byte)
- 140) ACPR Main Channel BW Minimum (in Hz)
- 141) ACPR Main Channel BW Minimum (in Hz)

⁴¹⁵ Value sent as (value * 1000) + 270,000

⁴¹⁶ Value sent as (value * 1000) + 270,000

⁴¹⁷ Value sent as (value in dB * 1000)

⁴¹⁸ Value sent as (value in dB * 1000)

⁴¹⁹ Value sent as ((value in Hz) – 10,000)

⁴²⁰ Value sent as ((value in Hz) – 10,000)

- 142) ACPR Main Channel BW Minimum (in Hz) (Lowest byte)
- 143) ACPR Main Channel BW Maximum (in Hz) (Highest byte)
- 144) ACPR Main Channel BW Maximum (in Hz)
- 145) ACPR Main Channel BW Maximum (in Hz)
- 146) ACPR Main Channel BW Maximum (in Hz) (Lowest byte)
- 147) ACPR Adjacent Channel BW Minimum (in Hz) (Highest byte)
- 148) ACPR Adjacent Channel BW Minimum (in Hz)
- 149) ACPR Adjacent Channel BW Minimum (in Hz)
- 150) ACPR Adjacent Channel BW Minimum (in Hz) (Lowest byte)
- 151) ACPR Adjacent Channel BW Maximum (in Hz) (Highest byte)
- 152) ACPR Adjacent Channel BW Maximum (in Hz)
- 153) ACPR Adjacent Channel BW Maximum (in Hz)
- 154) ACPR Adjacent Channel BW Maximum (in Hz) (Lowest byte)
- 155) ACPR Channel Spacing Minimum (in Hz) (Highest byte)
- 156) ACPR Channel Spacing Minimum (in Hz)
- 157) ACPR Channel Spacing Minimum (in Hz)
- 158) ACPR Channel Spacing Minimum (in Hz) (Lowest byte)
- 159) ACPR Channel Spacing Maximum (in Hz) (Highest byte)
- 160) ACPR Channel Spacing Maximum (in Hz)
- 161) ACPR Channel Spacing Maximum (in Hz)
- 162) ACPR Channel Spacing Maximum (in Hz) (Lowest byte)
- 163) Channel Power Integration BW Minimum (in Hz) (Highest byte)
- 164) Channel Power Integration BW Minimum (in Hz)
- 165) Channel Power Integration BW Minimum (in Hz)
- 166) Channel Power Integration BW Minimum (in Hz) (Lowest byte)
- 167) Channel Power Integration BW Maximum (in Hz) (Highest byte)
- 168) Channel Power Integration BW Maximum (in Hz)
- 169) Channel Power Integration BW Maximum (in Hz)
- 170) Channel Power Integration BW Maximum (in Hz) (Lowest byte)
- 171-300) Not Used

For Transmission Measurement Mode (Option 21 Only), All Other Parameter Limits:

- 21) Span Scale Factor Minimum⁴²¹ (Higher byte)
- 22) Span Scale Factor Minimum (Lower byte)
- 23) Span Scale Factor Maximum⁴²² (Higher byte)
- 24) Span Scale Factor Maximum (Lower byte)
- 25) Span Minimum⁴²³ (Highest byte)
- 26) Span Minimum
- 27) Span Minimum
- 28) Span Minimum (Lowest byte)
- 29) Span Maximum⁴²⁴ (Highest byte)
- 30) Span Maximum
- 31) Span Maximum
- 32) Span Maximum (Lowest byte)
- 33) Reference Level Minimum⁴²⁵ (Highest byte)
- 34) Reference Level Minimum
- 35) Reference Level Minimum
- 36) Reference Level Minimum (Lowest byte)
- 37) Reference Level Maximum⁴²⁶ (Highest byte)

⁴²¹ Scale Factor in number of Hz

⁴²² Scale Factor in number of Hz

⁴²³ Scaled by Span Scale Factor

⁴²⁴ Scaled by Span Scale Factor

⁴²⁵ Value sent as (value * 1000) + 270,000

⁴²⁶ Value sent as (value * 1000) + 270,000

- 38) Reference Level Maximum
- 39) Reference Level Maximum
- 40) Reference Level Maximum (Lowest byte)
- 41) Scale Minimum⁴²⁷ (Highest byte)
- 42) Scale Minimum
- 43) Scale Minimum
- 44) Scale Minimum (Lowest byte)
- 45) Scale Maximum⁴²⁸ (Highest byte)
- 46) Scale Maximum
- 47) Scale Maximum
- 48) Scale Maximum (Lowest byte)
- 49) Marker Minimum⁴²⁹ (Higher byte)
- 50) Marker Minimum (Lower byte)
- 51) Marker Maximum⁴³⁰ (Higher byte)
- 52) Marker Maximum (Lower byte)
- 53) Limit Y Minimum⁴³¹ (Highest byte)
- 54) Limit Y Minimum
- 55) Limit Y Minimum
- 56) Limit Y Minimum (Lowest byte)
- 57) Limit Y Maximum⁴³² (Highest byte)
- 58) Limit Y Maximum
- 59) Limit Y Maximum
- 60) Limit Y Maximum (Lowest byte)
- 61) Attenuation Minimum
- 62) Attenuation Maximum
- 63) Amplitude Units Minimum
- 64) Amplitude Units Maximum
- 65) Detection Algorithm Minimum
- 66) Detection Algorithm Maximum
- 67) External Reference Frequency Minimum⁴³³ (Highest byte)
- 68) External Reference Frequency Minimum
- 69) External Reference Frequency Minimum
- 70) External Reference Frequency Minimum (Lowest byte)
- 71) External Reference Frequency Maximum⁴³⁴ (Highest byte)
- 72) External Reference Frequency Maximum
- 73) External Reference Frequency Maximum
- 74) External Reference Frequency Maximum (Lowest byte)
- 75) Trigger Type Minimum
- 76) Trigger Type Maximum
- 77) Sweep Average Minimum
- 78) Sweep Average Maximum
- 79) Trace Math Minimum
- 80) Trace Math Maximum
- 81-200) Not Used

⁴²⁷ Value sent as (value * 1000)

⁴²⁸ Value sent as (value * 1000)

⁴²⁹ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

⁴³⁰ Value sent as data point on the display. Equivalent frequency = (point * span / (# data points – 1)) + start frequency.

⁴³¹ Value sent as (value * 1000) + 270,000

⁴³² Value sent as (value * 1000) + 270,000

⁴³³ Reference frequency in MHz

⁴³⁴ Reference frequency in MHz

For Power Meter Mode (Option 29 Only), All Other Parameter Limits:

- 21) Span Scale Factor Minimum⁴³⁵ (Higher byte)
 - 22) Span Scale Factor Minimum (Lower byte)
 - 23) Span Scale Factor Maximum⁴³⁶ (Higher byte)
 - 24) Span Scale Factor Maximum (Lower byte)
 - 25) Span Minimum⁴³⁷ (Highest byte)
 - 26) Span Minimum
 - 27) Span Minimum
 - 28) Span Minimum (Lowest byte)
 - 29) Span Maximum⁴³⁸ (Highest byte)
 - 30) Span Maximum
 - 31) Span Maximum
 - 32) Span Maximum (Lowest byte)
 - 33) Power Meter Offset Minimum (Highest byte)
 - 34) Power Meter Offset Minimum
 - 35) Power Meter Offset Minimum
 - 36) Power Meter Offset Minimum (Lowest byte)
 - 37) Power Meter Offset Maximum (Highest byte)
 - 38) Power Meter Offset Maximum
 - 39) Power Meter Offset Maximum
 - 40) Power Meter Offset Maximum (Lowest byte)
 - 41) Power Meter Relative Minimum⁴³⁹ (Highest byte)
 - 42) Power Meter Relative Minimum
 - 43) Power Meter Relative Minimum
 - 44) Power Meter Relative Minimum (Lowest byte)
 - 45) Power Meter Relative Maximum⁴⁴⁰ (Highest byte)
 - 46) Power Meter Relative Maximum
 - 47) Power Meter Relative Maximum
 - 48) Power Meter Relative Maximum (Lowest byte)
 - 49-150) Not Used
-

⁴³⁵ Scale Factor in number of Hz

⁴³⁶ Scale Factor in number of Hz

⁴³⁷ Scaled by Span Scale Factor

⁴³⁸ Scaled by Span Scale Factor

⁴³⁹ Value sent as ((value in dBm + 100) * 1000)

⁴⁴⁰ Value sent as ((value in dBm + 100) * 1000)

Query Saved Setups – Control Byte #68 (44h)

Description: Returns a list of setups saved for the specified measurement mode. Modes that are stored in the same table (i.e. Spectrum Analyzer and Transmission Measurement modes) will be returned by this command when either mode is specified.

Frequency Scale Factor is defined as the number of Hz.

Start and Stop frequencies are sent scaled by the Frequency Scale Factor.

Bytes to Follow: 1 byte

- 1) Measurement Mode⁴⁴¹

Spectrum Master Returns:

For All Modes:

- 1) Number of Following Bytes (Higher byte)
- 2) Number of Following Bytes (Lower byte)
- 3) Number of Setups

For Each Setup:

- 1) Setup Number
- 2) Attributes
bit 0: Read Only Status (00h = Write-able, 01h = Read Only)
bits 1-7: Not Used
- 3) Measurement Mode⁴⁴²
- 4) Cal Status (Transmission Mode Setup Only, 00h = Off, 01h = On)
- 5) Frequency Scale Factor (Higher byte)
- 6) Frequency Scale Factor (Lower byte)
- 7) Start Frequency (Highest byte)
- 8) Start Frequency
- 9) Start Frequency
- 10) Start Frequency (Lowest byte)
- 11) Stop Frequency (Highest byte)
- 12) Stop Frequency
- 13) Stop Frequency
- 14) Stop Frequency (Lowest byte)
- 15-20) Not Used

⁴⁴¹ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

⁴⁴² Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

Enter Remote Mode – Control Byte #69 (45h)

Description: Enter remote mode at the end of a sweep then send model number and firmware version to the computer.

The computer sends Enter Remote mode byte #69 (45h) to the Spectrum Master and waits for response.

Since the Spectrum Master polls its serial port buffer at the end of each sweep, the computer must wait until the Spectrum Master sends the return bytes before sending a new control byte. Otherwise, the new control byte overwrites the old one (saying enter remote) and the Spectrum Master does not respond as expected.

Once in remote mode, the Spectrum Master stops sweeping. A Remote Mode Indicator appears on the LCD.

The Spectrum Master sends its model and software version numbers to the computer. The Spectrum Master is now able to take multiple control bytes. It waits for the next control byte.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 13 bytes

- 1-2) Model # (unsigned integer, 16h for Spectrum Master MS2711D)
 - 3-9) Extended Model # (7 bytes in ASCII)
 - 10-13) Software Version - 4 bytes (ASCII)
-

Enter Remote Mode Immediately – Control Byte #70 (46h)

Description: Enter remote mode in the middle of a sweep, then send the model number and firmware version to the computer.

The computer sends Enter Remote Mode Immediately byte #70 (46h) to the Spectrum Master and waits for a response. This control byte causes the unit to enter remote mode immediately. Note that this could result in incomplete sweep data. Use control byte #69 if complete data is required.

Once in remote mode, the Spectrum Master stops sweeping. A Remote Mode Indicator appears on the LCD.

The Spectrum Master sends its model and software version numbers to the computer. The Spectrum Master is now able to take multiple control bytes. It waits for the next control byte.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 13 bytes

- 1-2) Model # (unsigned integer, 16h for Spectrum Master MS2711D)
 - 3-9) Extended Model # (7 bytes in ASCII)
 - 10-13) Software Version (4 bytes in ASCII)
-

Write Protect Setup – Control Byte #71 (47h)

Description: Makes a saved setup either read-only or writeable.

Setup numbers as follows:

- 255 = All Setups in the Specified Mode
- 1 – 10 = Saved setups for Spectrum Analyzer/Transmission Measurement modes
- 1 – 5 = Saved setups for Power Meter mode (Option 29 Only)

Bytes to Follow: 3 bytes

- 1) Measurement Mode⁴⁴³
- 2) Setup Number
- 3) Write-Protect Status (00h = Allow Writes (default), 01h = Lock Setup (i.e. “read only”))

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Clear Setup Memory Location – Control Byte #72 (48h)

Description: Clears a setup memory location such that it appears as “<EMPTY>” in the Recall Setup list.

Setup numbers as follows:

- 255 = All Setups in the Specified Mode
- 1 – 10 = Saved setups for Spectrum Analyzer/Transmission Measurement modes
- 1 – 5 = Saved setups for Power Meter mode (Option 29 Only)

Bytes to Follow: 2 byte

- 1) Measurement Mode⁴⁴⁴
- 2) Setup Number

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

⁴⁴³ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

⁴⁴⁴ Refer to Control Byte #3 “Select Measurement Mode” for valid measurement modes.

Write Antenna – Control Byte #82 (52h)

Description: Receives an antenna to the Spectrum Master via the serial port.

An antenna is described with an index into the list (1-10) and an ASCII name that appears in the list on the Spectrum Master. Each antenna can have up to 60 antenna factors. Each antenna factor has an associated frequency and value. These are specified one at a time.

Frequencies are sent in Hz scaled by the Scale Factor.

The value of the antenna factor should be sent as (value * 100).

Bytes to Follow: 26 – 380, depending on the number of antenna factors

- 1) Antenna List Index (1-10)
- 2-17) Antenna Name (in ASCII)
- 18) Number of Antenna Factors (max = 60)
- 19-20) Frequency Scale Factor (in Hz)

For each antenna factor:

- 1) Frequency (scaled by Scale Factor) (Highest byte)
- 2) Frequency (scaled by Scale Factor)
- 3) Frequency (scaled by Scale Factor)
- 4) Frequency (scaled by Scale Factor) (Lowest byte)
- 5) Antenna Factor (Higher byte)
- 6) Antenna Factor (Lower byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Recall Antenna – Control Byte #83 (53h)

Description: Sends an antenna from the Spectrum Master via the serial port.

An antenna is described with an index into the list (1-10) and an ASCII name that appears in the list on the Spectrum Master. Each antenna can have up to 60 antenna factors. The number of antenna factors will be sent before the actual values are sent. Each antenna factor has an associated frequency and value. These are specified one at a time.

Frequencies are sent in Hz scaled by the Scale Factor.

The value of the antenna factor should be sent as (value * 100).

Bytes to Follow: 1 byte

- 1) Antenna List index (1-10)

Spectrum Master Returns: (28-382 bytes, depending on the number of antenna factors)

- 1) Maximum Antenna Number (10)
- 2-17) Antenna Name (in ASCII)
- 18) Number of Antenna Factors (max = 60)
- 19-20) Frequency Scale Factor (in Hz)
- 21-22) Number of Following Bytes

For each antenna factor:

- 1) Frequency (scaled by Scale Factor) (Highest byte)
- 2) Frequency (scaled by Scale Factor)
- 3) Frequency (scaled by Scale Factor)
- 4) Frequency (scaled by Scale Factor) (Lowest byte)
- 5) Antenna Factor (Higher byte)
- 6) Antenna Factor (Lower byte)

Set Field Strength Measurement – Control Byte #84 (54h)

Description: Sets the state of the measurement (ON or OFF) and the antenna index for the field strength measurement. Antennas 1-10 are custom antennas. Antennas 11-15 are the standard antennas. The standard antennas are as follows:

11. Anritsu #2000-1030 (MAXRAD MPA1750) – 1710-1880 MHz
12. Anritsu #2000-1031 (MAXRAD MPA1850) – 1850-1990 MHz
13. Anritsu #2000-1032 (MAXRAD MPA2450) – 2400-2483.5 MHz
14. Anritsu #2000-1200 (Centurion EXCSM806) – 806-899 MHz
15. Anritsu #2000-1035 (Centurion EXE-902-SM) – 896-941 MHz

If the FCN4760 frequency converter module is attached, the standard antenna is:

11. Anritsu #2000-1361 – 5725-5825 MHz

Note that if the field strength measurement is turned ON, all other measurements (channel power, adjacent channel power) are turned OFF.

Bytes to Follow: 2 bytes

- 1) Field Strength Measurement State (On/Off)
- 2) Antenna List index (1-15)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid state or index
- 238 (EEh) Time Out Error

Set Channel Power – Control Byte #85 (55h)

Description: Sets the state of the measurement (ON or OFF), and the setup parameters to perform the channel power measurement.

Send a 0 (zero) following the command to set the channel power measurement in the current setup.

Send a 1 (one) to set the channel power associated with the trace that was most recently uploaded by command #36, Upload Sweep Trace.

If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Note that if the channel power measurement is turned ON, all other measurements (field strength, adjacent channel power) are turned OFF.

Bytes to Follow: 14 bytes

- 1) Channel Power Location (0 = current setup, 1 = last uploaded trace)
- 2) Channel Power Measurement State (On/Off)
- 3-6) Center Frequency
- 7-10) Integration Bandwidth
- 11-14) Span Frequency

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Read Channel Power – Control Byte #86 (56h)

Description: Read the current channel power or the channel power of a stored trace.

Send a 0 (zero) following the command to read the current channel power measurement (i.e. the one that is updated as the unit is sweeping).

Send 1-200 to read the channel power associated with a stored trace (use Query Trace Names, #24, to obtain trace numbers).

If option 6 is installed and the frequency converter module is attached, the frequencies will be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 1 byte

- 1) Channel Power Location (0 = current measured value, 1-200 = value in stored trace)

Spectrum Master Returns: 21 bytes

- 1) Channel Power On/Off
 - 2-5) Channel Center Frequency
 - 6-9) Integration Bandwidth
 - 10-13) Channel Span Frequency
 - 14-17) Channel Power (= (power in dBm * 100) + 270000)
 - 18-21) Channel Power Density (= (density in dBm/Hz * 100) + 270000)
-

Set Adjacent Channel Power Ratio (ACPR) – Control Byte #87 (57h)

Description: Sets the state of the measurement (ON or OFF), the center frequency, the main channel bandwidth, the adjacent channel bandwidth and the channel spacing.

Send a 0 (zero) following the command to set the channel power measurement in the current setup.

Send a 1 (one) to set the adjacent channel power associated with the trace that was most recently uploaded by command #36, Upload Sweep Trace.

If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Note that if the ACPR measurement is turned ON, all other measurements (field strength, channel power) are turned OFF.

Bytes to Follow: 18 bytes

- 1) Adjacent Channel Power Location (0 = current setup, 1 = last uploaded trace)
- 2) Adjacent Channel Power Measurement State (On/Off)
- 3-6) Center Frequency
- 7-10) Main Channel Bandwidth
- 11-14) Adjacent Channel Bandwidth
- 15-18) Channel Spacing

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Read Adjacent Channel Power (ACPR) – Control Byte #88 (58h)

Description: Read the current adjacent channel power or the adjacent channel power of a stored trace.

Send a 0 (zero) following the command to read the current adjacent channel power measurement (i.e. the one that is updated as the unit is sweeping).

Send 1-200 to read the channel power associated with a stored trace (use Query Trace Names, #24, to obtain trace numbers).

If option 6 is installed and the frequency extension module is attached, the frequencies will be scaled by the scale factor of the module. If the module is not attached, frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 1 byte

- 1) Adjacent Channel Power Ratio Location (0 = current measured value, 1-200 = value in stored trace)

Spectrum Master Returns: 29 bytes

- 1) ACPR On/Off
 - 2-5) Main Channel Center Frequency
 - 6-9) Main Channel Bandwidth
 - 10-13) Adjacent Channel Bandwidth
 - 14-17) Channel Spacing
 - 18-21) Main Channel Power (= (power in dBm * 100) + 270000)
 - 22-25) Lower Adjacent Channel Power (= (power in dBm * 100) + 270000)
 - 26-29) Upper Adjacent Channel Power (= (power in dBm * 100) + 270000)
-

Select Signal Standard – Control Byte #89 (59h)

Description: Select signal standard and the link direction.

Bytes to Follow: 2 bytes

- 1) Signal Standard Index (0-199) – See “Signal Standards” for details.
- 2) Link direction (1= Uplink, 2=Downlink, 3=Uplink and downlink)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Invalid signal standard
238 (EEh) Time Out Error
-

Select Channel in Current Standard – Control Byte #90 (5Ah)

Description: Selects a channel within the range of the currently selected signal standard. Use this command only in Spectrum Analyzer mode, Power Meter mode, CDMA mode and GSM mode.

See the section “Signal Standards” for a list of valid channels for the selected channel.

Bytes to Follow: 2 bytes

- 1) Channel (Higher byte)
- 2) Channel (Lower byte)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Invalid channel
238 (EEh) Time Out Error
-

Read Current Settings of Signal Standard – Control Byte #91 (5Bh)

Description: Read the current settings of signal standard. This command can be used in any measurement mode.

Bytes to follow: none

Cell Master Returns: 44 bytes

- 1) Signal Standard ID (0-199) – See “Signal Standards” for detail.
 - 2) Link direction (0= N/A, 1= Uplink, 2=Downlink, 3=Uplink and downlink)
 - 3 – 4) Channel Number (high - low)
 - 5 – 30) Standard Name in ASCII
-

Upload User's Signal Standard Table – Control Byte #92 (5Ch)

Description: Upload user's signal standard table to Cell Master.

Bytes to Follow: Variable (64 bytes * n) where n is the total record of the signal standards.

Each record has the following data field.

- 1–2): Record type (Record type = 0xffff means last record)
- 3- 42): Signal Standard Name (maximum 40 bytes)
- 43-46): Center Frequency of first channel of uplink(in Hz)
- 47-50): Center Frequency of first channel of downlink (in Hz)
- 51-52): Start Channel Number
- 53-54): Stop Channel Number
- 55-58): Channel Band width (in Hz)
- 59-62): Channel Spacing (in Hz)
- 63-64): Channel Step.
- 65-128): 2nd record.
- 129-192): 3rd record.

Etc.

Cell Master Returns: 1 byte

255 (FFh) Operation Complete Byte

224 (E0h) Parameter Error: Invalid signal standard

238 (EEh) Time Out Error

Measure OCC BW % of Power – Control Byte #96 (60h)

Description: Measure OCC BW with % of Power method.

If option 6 is installed and the frequency extension module is attached, the OBW frequencies will be scaled by the scale factor of the module. If the module is not attached, the OBW frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 4 bytes

- 1) % of Power (Highest byte)
- 2) % of Power
- 3) % of Power
- 4) % of Power (Lowest byte) (in 100th of %, 9123 = 91.23%)

Spectrum Master Returns: 16 bytes

- 1-4) Occupied Bandwidth (in Hz)
 - 5-8) Measure dB down (dB * 100,000)
 - 9-12) Low Frequency OCC BW
 - 13-16) High Frequency OCC BW
-

Measure OCC BW dB Down – Control Byte #97 (61h)

Description: Measure OCC BW with dB down method.

If option 6 is installed and the frequency converter module is attached, the OBW frequencies will be scaled by the scale factor of the module. If the module is not attached, the OBW frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 4 bytes

1-4) dB down (in 100th of dB, 1234 = 12.34dB)

Spectrum Master Returns: 16 bytes

- 1-4) Occupied Bandwidth (in Hz)
 - 5-8) Measure % of Power (% of power * 100)
 - 9-12) Low Frequency OCC BW
 - 13-16) High Frequency OCC BW
-

Set Bias Tee Function - Control Byte #98 (62h)

This command is available only with Option 10.

Description: Set the Bias Tee function On/Off. If the Bias Tee is turned on, the Spectrum Master returns the results of Bias Tee.

Bytes to Follow: 1 byte

- 00h – Turns the Bias Tee Off
- 01h – Turns the Bias Tee On

Spectrum Master Returns:

If bias tee is turned Off (1 byte)

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid state
- 238 (EEh) Time-out Error

If bias tee is turned On (10 bytes)

- 1) Bias Tee Board Indicator (00h = No Hardware Installed, 01h = Hardware Installed)
- 2) Bias Tee Current (Highest byte)
- 3) Bias Tee Current
- 4) Bias Tee Current
- 5) Bias Tee Current (Lowest byte)
- 6) 10 * Bias Tee Voltage (Highest byte) : voltage value is in volt/10
- 7) 10 * Bias Tee Voltage
- 8) 10 * Bias Tee Voltage
- 9) 10 * Bias Tee Voltage (Lowest byte)
- 10) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error
- 238 (EEh) Time-out Error

Note: Due to the hardware delay, the Spectrum Master does not return the results of the Bias Tee until approximately 3 seconds after the Bias Tee is turned on.

Set Spectrum Analyzer Start/Stop Frequency – Control Byte #99 (63h)

Description: Sets the spectrum analyzer start and stop frequencies.

If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 8 bytes

- 1) Start Frequency (Highest byte)
- 2) Start Frequency
- 3) Start Frequency
- 4) Start Frequency (Lowest byte)
- 5) Stop Frequency (Highest byte)
- 6) Stop Frequency
- 7) Stop Frequency
- 8) Stop Frequency (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid frequency range
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Center Freq./Span – Control Byte #100 (64h)

Description: Sets the spectrum analyzer center frequency and span.

If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 8 bytes

- 1) Center Frequency (Highest byte)
- 2) Center Frequency
- 3) Center Frequency
- 4) Center Frequency (Lowest byte)
- 5) Frequency Span (Highest byte)
- 6) Frequency Span
- 7) Frequency Span
- 8) Frequency Span (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid frequency range
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Scale – Control Byte #101 (65h)

Description: Sets the reference level and the number of dB represented by each graph division.

Ref Level will be the “top” scale of the graph, and there are total of 10 division, so bottom scale can be determined by : Ref level + 10 x dB/div.

Bytes to Follow: 8 bytes

- 1) Ref Level (Highest byte)
- 2) Ref Level
- 3) Ref Level
- 4) Ref Level (lowebyte)
- 5) dB/div (Highest byte)
- 6) dB/div
- 7) dB/div
- 8) dB/div (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid scale
- 238 (EEh) Time Out Error

Notes:

Ref Level is sent as the (Ref Level * 1000) + 270,000 (0 dBm = 270,000, 20 dBm = 290000, -120 dBm = 150,000)
Scale should be sent as (dBm * 1000) (e.g. -12.34 dBm = -12340)

Set Spectrum Analyzer Marker – Control Byte #102 (66h)

Description: Sets an individual Spectrum Analyzer marker.

Bytes to Follow: 5 bytes

- 1) Marker Number (01h = marker 1, 02h = marker 2, 03h = marker 3, 04h = marker 4, 05h = marker 5, 06h = marker 6)
- 2) Marker Line On/Off (01h = On, 00h = Off)
- 3) Marker Delta Status On/Off (01h = On, 00h = Off)
- 4) Marker Value (Higher byte)
- 5) Marker Value (Lower byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid marker number, status or position
- 238 (EEh) Time Out Error

Note:

Marker Value is between 0 and 400, inclusive: Point = (400 * (marker freq - start freq)) / span

Set Spectrum Analyzer Single Limit – Control Byte #103 (67h)

Description: Sets the position and On/Off Status of the Limit Line.

Bytes to Follow: 6 bytes

- 1) Limit Number (1 for Spectrum Master)
- 2) Limit Line On/Off (01h = On, 00h = Off)
- 3) Beep at Limit On/Off (01h = On, 00h = Off)
- 4) Limit Value (Highest byte)
- 5) Limit Value
- 6) Limit Value
- 7) Limit Value (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid limit number, status or value
- 238 (EEh) Time Out Error

Note:

Limit Value is sent as the (Limit Value * 1000) + 270,000 (0 dBm=270,000, 20 dBm=290000, -120 dBm=150,000)

Set Spectrum Analyzer Peak Hold – Control Byte #105 (69h)

Description: Sets the max hold and min hold settings on the Spectrum Analyzer.

Bytes to Follow: 1 byte

- 1) Peak Hold State
 - 00h – Peak Hold Off
 - 01h – Max Hold On
 - 02h – Min Hold On

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid state
 - 238 (EEh) Time Out Error
-

OBSOLETE: Set SPA Resolution Bandwidth – Control Byte #106 (6Ah)

This command exists for backward compatibility with the MS2711B. To access new video bandwidths use Control Byte #141 (8Dh).

Description: Sets the resolution BW frequency for the Spectrum Analyzer.

Bytes to Follow: 1 byte

- 1) Resolution Bandwidth Index
 - 00h – 10 kHz BW
 - 01h – 30 kHz BW
 - 02h – 100 kHz BW
 - 03h – 1 MHz BW

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid RBW Index
 - 238 (EEh) Time Out Error
-

OBSOLETE: Set Spectrum Analyzer Video Bandwidth – Control Byte #107 (6Bh)

This command exists for backward compatibility with the MS2711B. To access new video bandwidths use Control Byte #142 (8Eh).

Description: Sets the video BW frequency for the Spectrum Analyzer.

Bytes to Follow: 1 byte

- 1) Video Bandwidth Index
 - 00h – 100 Hz BW
 - 01h – 300 Hz BW
 - 02h – 1 kHz BW
 - 03h – 3 kHz BW
 - 04h – 10 kHz BW
 - 05h – 30 kHz BW
 - 06h – 100 kHz BW
 - 07h – 300 kHz BW

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid VBW Index
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Sweep Mode – Control Byte #108 (6Ch)

Description: Enables or disables the Single Sweep Mode during Spectrum Analyzer mode of operation.

Single Sweep Mode activates once the Spectrum Master exits from the remote mode.

For Single Sweep Mode during Spectrum Master VNA modes of operation see control byte #11 (0Bh).

Bytes to Follow: 1 byte

- 1) Sweep Mode
 - 00h – Single Sweep
 - 01h – Continuous Sweep
 - 02h – Video Trigger (span must be 0)
 - 03h – External Trigger (span must be 0)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid Mode
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Marker to Peak – Control Byte #109 (6Dh)

Description: Sets the specified marker to the peak value of the sweep.

Bytes to Follow: 1 byte

- 1) Marker Number (1-6)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid Marker Number
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Marker to Center – Control Byte #110 (6Eh)

Description: Sets the center frequency equal to the frequency of the specified marker.

Bytes to Follow: 1 byte

- 1) Marker Number (1-4)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid Marker Number
 - 238 (EEh) Time Out Error
-

OBSOLETE: Set Spectrum Analyzer Attenuation – Control Byte #111 (6Fh)

This command exists for backward compatibility with the MS2711B. For future development, use Control Byte #143 (8Fh).

Description: Sets the attenuation for the Spectrum Master Spectrum Analyzer mode. Send a value of 255 (FFh) to enable dynamic attenuation.

Automatic control couples the attenuation to the reference level. Note that setting the attenuation using this command automatically sets the attenuation coupling to “MANUAL”, thereby allowing it to be defined independently of the reference level.

Bytes to Follow: 1 byte

- 1) 00h – 0 dB
- 01h – 10 dB
- 02h – 20 dB
- 03h – 30 dB
- 04h – 40 dB
- 05h – 50 dB
- FFh – Dynamic Attenuation

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid Attenuation Value
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Multiple Limit – Control Byte #113 (71h)

Description: Sets the position and On/Off Status of a limit segment.

Multiple limits are defined by multiple limit segments, each with a different finite slope. The single limit is a single, horizontal line that can be defined to act as an upper limit or as a lower limit. See control byte #103 for information about the single limit.

The limit types are mutually exclusive. That is, you cannot have both single and multiple limits at the same time. Note that setting a limit segment ON automatically makes the limit type “MULTIPLE”.

One segment is defined each time this command is sent to the Spectrum Master. The first two bytes of the command specify which segment is being defined. There are 5 upper limits and 5 lower limits available in Spectrum Analyzer mode. Byte 1 selects the segment number. Byte 2 specifies whether it is an upper limit or a lower limit. Byte 3 turns the segment ON or OFF. Byte 4 specifies whether the error beep sounds when the bound set by the segment is exceeded by the measured data.

The segment location is defined by its endpoints. The “Start” endpoint must appear to the left of the “End” endpoint on the graph. That is, Start X < End X. If Start X = End X then Start Y must equal End Y. Vertical segments are not allowed.

Bytes to Follow: 20 bytes

- 1) Segment number (1-5)
- 2) Segment type (00h = LOWER limit, 01h = UPPER limit)
- 3) Limit Line On/Off (01h = On, 00h = Off)
- 4) Limit Beep On/Off (01h = On, 00h = Off)
- 5) Limit Value Start X ⁴⁴⁵(Highest byte)
- 6) Limit Value Start X
- 7) Limit Value Start X
- 8) Limit Value Start X (Lowest byte)
- 9) Limit Value Start Y ⁴⁴⁶(Highest byte)
- 10) Limit Value Start Y
- 11) Limit Value Start Y
- 12) Limit Value Start Y (Lowest byte)
- 13) Limit Value End X ⁴⁴⁷(Highest byte)
- 14) Limit Value End X
- 15) Limit Value End X
- 16) Limit Value End X (Lowest byte)
- 17) Limit Value End Y ⁴⁴⁸(Highest byte)
- 18) Limit Value End Y
- 19) Limit Value End Y
- 20) Limit Value End Y (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid limit segment, status or value
 - 238 (EEh) Time Out Error
-

⁴⁴⁵ If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

⁴⁴⁶ (Value in dBm * 1000) + 270,000

⁴⁴⁷ If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

⁴⁴⁸ (Value in dBm * 1000) + 270,000

Set Return Spectrum Analyzer Sweep Time – Control Byte #114 (72h)

Description: If this is enabled, the duration of the current sweep (in milliseconds) will be returned as 4 bytes via the serial port at the end of the sweep. If Serial Echo Status is enabled, the 4 bytes will be returned AFTER the sweep complete byte.

Bytes to Follow: 1 byte

- 1) Return SPA Sweep Time flag state
00h = Don't Return Sweep Time
01h = Return Sweep Time

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Invalid state
238 (EEh) Time Out Error
-

Set Reference Level Offset – Control Byte #115 (73h)

Description: Set the value of the reference level offset.

The reference level offset allows the user to view the result of trace math (A+B, A-B) even if it is greater than +20 dBm or less than -120 dBm. The offset is a constant that is subtracted from the reference level.

Note that the valid range is -100 to +100 dB.

Send the value as (value in dB * 1000) + 270,000.

For example, to compensate for a 30 dB attenuator, the reference level offset should be -30 dB. That value would be sent over the serial port as (-30 * 1000) + 270,000 = 240,000.

Bytes to Follow: 4 bytes

- 1) Reference Level Offset (Highest byte)
- 2) Reference Level Offset
- 3) Reference Level Offset
- 4) Reference Level Offset (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error
238 (EEh) Time Out Error
-

Set Spectrum Analyzer Impedance – Control Byte #116 (74h)

Description: Set the impedance and the loss value due to an adapter.

The MS2711D can automatically compensate for the effects of impedance adapters. The impedance of the MS2711D is 50Ω , so there is no need for an adapter in this case. The loss for the Anritsu 75Ω adapter 12N50-75B is known by the MS2711D.

This control byte also allows for the specification of the impedance and the loss due to an adapter the system does not know. In either case, 5 bytes must be sent to the unit. If the impedance is 50Ω or one of the known adapters is specified, bytes 2-5 are ignored. If an unknown adapter is specified, the unit uses bytes 2-5 to correct for the adapter.

Bytes to Follow: 5 bytes

- 1) Impedance Adapter⁴⁴⁹
- 2) Impedance Loss⁴⁵⁰ (Highest byte)
- 3) Impedance Loss
- 4) Impedance Loss
- 5) Impedance Loss (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

⁴⁴⁹ Impedance Adapter: 00h = 50Ω 0Ah = 75Ω , adapter 12N50-75B 0Ch = 75Ω , other adapter offset

⁴⁵⁰ Send the loss value as value in dB* 1,000

Read Marker Value – Control Byte #117 (75h)

Description: Returns the frequency location of the specified marker, and the value at that location.

If option 6 is installed and the frequency converter module is attached, the frequency will be scaled by the scale factor of the module. If the module is not attached, the frequency is sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

Bytes to Follow: 1 byte

- 1) Marker number (1-6)

Spectrum Master Returns: 8 bytes (1 byte if an error occurs)

- 1) Frequency (Highest byte)
- 2) Frequency
- 3) Frequency
- 4) Frequency (Lowest byte)
- 5) Value at Marker (Highest byte)
- 6) Value at Marker
- 7) Value at Marker
- 8) Value at Marker (Lowest byte)

OR

- 1) 224 (E0h) Parameter Error: Invalid marker number
- 238 (EEh) Time-out Error

Note:

Marker value sent as (value in dBm * 1,000) +270,000

If markers are set to be noise markers, convert the returned dBm value to dBm/Hz using this formula (only if detection method is RMS Average):

$$\text{marker (in dBm/Hz)} = \text{marker value (in dBm)} - 10 * \log_{10}(\text{RBW}) - 0.13$$

Set Sweep Averaging – Control Byte #118 (76h)

Description: Sets the number of sweeps to average. The maximum number is 25. Sending a 1 turns averaging off.

Bytes to Follow: 1 byte

- 1) Number of sweeps to average (1-25, 1 turns averaging OFF)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Normalize SPA – Control Byte #130 (82h)

Description: Performs Normalization of SPA graph.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 238 (EEh) Time Out Error
-

Set SPA Normalization Status – Control Byte #131 (83h)

Description: Turns Normalization ON/OFF

Bytes to Follow: 1 bytes

- 1) Normalization Status (00h = Off, 01h = On)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error
- 238 (EEh) Time Out Error

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Check Spectrum Analyzer External Reference – Control Byte #134 (86h)

Description: Checks to see if Spectrum Analyzer external reference is present. If it is, it then checks to see if it is at the correct frequency for PLL locking.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

On Success:

- 1) 00h – Reference present and at the correct frequency (PLL functioning)
- 01h – Reference is not present
- 02h – Reference is present, but internal PLL and external frequency do not match up.

OR

On Error:

- 1) 224 (E0h) Parameter Error – Not in External reference mode
 - 238 (EEh) Time-out Error.
-

Set SA Preamp State (On/Off/Auto) – Control Byte #136 (88h)

Description: Sets the state of Spectrum Analyzer preamplifier.

Setting the preamp state to ON or OFF sets the preamp coupling to manual. That is, the preamplifier state is controlled independently of all other parameters.

Setting the preamp state to AUTO couples the preamp state to the reference level and the attenuation. If the attenuation is automatically coupled to the reference level, the preamp will turn on when the reference level is set less than -26 dBm. If the attenuation is manually coupled to the reference level, the preamp will turn on when the value of (attenuation – reference level) >= 51.

Bytes to Follow: 1 byte

- 1) Mode (00h = Off, 01h = On, 02h = Auto)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid state
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Units – Control Byte #140 (8Ch)

Description: Sets the scale type (logarithmic or linear) and the units.

Linear units can be:

- 01h = Volts
- 02h = Watts

Logarithmic units can be:

- 03h = dBm
- 04h = dBV
- 05h = dBmV
- 06h = dB μ V

Bytes to Follow: 2 bytes

- 1) Scale Type (00h = Linear, 01h = Logarithmic)
- 2) Units

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time-out Error
-

Set Spectrum Analyzer Resolution Bandwidth – Control Byte #141 (8Dh)

This command is new to the MS2711D. Use it instead of Control Byte #106 to access the new RBWs.

Description: Sets the resolution BW frequency for the Spectrum Analyzer.

Bytes to Follow: 4 bytes

- 1) Resolution Bandwidth (frequency in Hz) (Highest byte)
- 2) Resolution Bandwidth (frequency in Hz)
- 3) Resolution Bandwidth (frequency in Hz)
- 4) Resolution Bandwidth (frequency in Hz) (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid RBW
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Video Bandwidth – Control Byte #142 (8Eh)

This command is new to the MS2711D. Use it instead of Control Byte #107 to access the new VBWs.

Description: Sets the video BW frequency for the Spectrum Analyzer.

Bytes to Follow: 4 bytes

- 1) Video Bandwidth (frequency in Hz) (Highest byte)
- 2) Video Bandwidth (frequency in Hz)
- 3) Video Bandwidth (frequency in Hz)
- 4) Video Bandwidth (frequency in Hz) (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid VBW
 - 238 (EEh) Time Out Error
-

Set Spectrum Analyzer Attenuation – Control Byte #143 (8Fh)

This command is new to the MS2711D. Use it instead of Control Byte #111 to access the new attenuations.

Description: Sets the attenuation of the Spectrum Analyzer. Send a value of 255 (FFh) to enable dynamic attenuation.

Automatic control couples the attenuation to the reference level. Note that setting the attenuation using this command automatically sets the attenuation coupling to “MANUAL”, thereby allowing it to be defined independently of the reference level.

Bytes to Follow: 1 byte

- 1) Attenuation (0 – 51)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid attenuation
 - 238 (EEh) Time Out Error
-

Set AM/FM Demodulation – Control Byte #145 (91h)

Description: Sets the AM/FM/SSB Demodulation state. This command is also used to set the type of Modulation, volume, Demodulation Frequency, BFO Adjust (SSB only) and the Demodulation time. On turning demodulation ON, after exiting remote, at the end of every sweep, demodulation is performed at the Demodulation frequency for a period of time specified in the Demod Time.

Bytes to Follow: 16 bytes

- 1) Set AM/FM/SSB Demod On/Off⁴⁵¹
- 2) Demodulation Type⁴⁵²
- 3) Speaker Volume (Higher byte)⁴⁵³
- 4) Speaker Volume (Lower byte)
- 5) Demodulation Time⁴⁵⁴ (Highest byte)
- 6) Demodulation Time
- 7) Demodulation Time
- 8) Demodulation Time (Lowest byte)
- 9) Demodulation Frequency⁴⁵⁵ (Highest byte)
- 10) Demodulation Frequency
- 11) Demodulation Frequency
- 12) Demodulation Frequency (Lowest byte)
- 13) SSB BFO Adjust⁴⁵⁶ (Highest byte)
- 14) SSB BFO Adjust
- 15) SSB BFO Adjust
- 16) SSB BFO Adjust (Lowest byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error
 - 238 (EEh) Time Out Error
-

Set Baud Rate – Control Byte #197 (C5h)

Description: Set baud rate for this session. An invalid setting returns the baud rate to 9600.

Bytes to Follow: 1 byte

- 1) Baud Rate Index
 - 00h = 9600 baud
 - 01h = 19200 baud
 - 02h = 38400 baud
 - 03h = 56000 baud
 - 04h = 115200 baud

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid baud rate index
 - 238 (EEh) Time Out Error
-

⁴⁵¹ 00h = Off, 01h = On

⁴⁵² 00h = FM Wideband, 01h = FM Narrowband, 02h = AM, 03h = SSB Lower, 04h = SSB Upper

⁴⁵³ Speaker Volume is from 0 to 100 in steps of 10

⁴⁵⁴ Demodulation time in milliseconds from 100 millisec to 500 seconds

⁴⁵⁵ If option 6 is installed and the frequency converter module is attached, the frequencies should be scaled by the scale factor of the module. If the module is not attached, the frequencies are sent in Hz. Use Control Word A203 to determine whether a module is attached and the appropriate scale factor.

⁴⁵⁶ BFO Valid Values are -10 kHz to +10 kHz. Send value as BFO(in Hz) + 10,000. For Example -10 kHz would be sent as 0, 0 would be sent as 10000 and +10 kHz would be 20000

Set Language – Control Byte #198 (C6h)

Description: Set the Spectrum Master display language.

Bytes to Follow: 1 byte

- 1) Language Index
 - 00h = English
 - 01h = French
 - 02h = German
 - 03h = Spanish
 - 04h = Chinese
 - 05h = Japanese

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 2) 224 (E0h) Parameter Error: Invalid language index
 - 3) 238 (EEh) Time Out Error
-

Query Time – Control Byte #208 (D0h)

Description: Queries the Spectrum Master for the current time in ASCII format.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 8 bytes (HH:MM:SS)

- 1) Hour (Higher byte)
 - 2) Hour (Lower byte)
 - 3) :
 - 4) Minute (Higher byte)
 - 5) Minute (Lower byte)
 - 6) :
 - 7) Second (Higher byte)
 - 8) Second (Lower byte)
-

Read Main Serial Number – Control Byte #221 (DDh)

Description: Returns the Main (External) Serial Number as four bytes. This command remains for backward compatibility.

A better command to use would be “Read ASCII Serial Number” #225 (E1h) which returns the serial number in ASCII format.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 4 bytes

- 1) Main Serial Number (Highest byte)
 - 2) Main Serial Number
 - 3) Main Serial Number
 - 4) Main Serial Number (Lowest byte)
-

Read ASCII Serial Number – Control Byte #225 (E1h)

Description: Reads and returns the Spectrum Master serial number as 8 ASCII bytes.

Bytes to Follow: 1 byte

- 1) Serial number storage location
 - 01h = Main (External) Serial Number
 - 02h = Secondary (Motherboard) Serial Number
 - 03h = T1/E1 Serial Number

Spectrum Master Returns: 8 bytes

1-8) Serial Number (in ASCII)

GPS Power – Control Byte #237 (EDh)

Description: Turn On/Off power of GPS module.

Bytes to Follow: 1 bytes

Power Switch (1=ON, others=OFF)

Cell Master Returns: 1 byte

255 (FFh) Operation Complete Byte

224 (E0h) Parameter Error

238 (EEh) Time Out Error

Read GPS Position – Control Byte #238 (EEh)

Description: Read current GPS position data: Latitude, longitude, and Altitude.

Bytes to Follow: 0 byte

Cell Master Returns: 13 byte (if Ok)

- 1-2) Number of satellites in use (< 3 if not locked)
- 3-6) GPS Position – Latitude (long integer)⁴⁵⁷ (= -1 if not valid)
- 7-10) GPS Position – Longitude (long integer) (= -1 if not valid)
- 11-12) GPS Position – Altitude (short integer) (= -30000 if not valid)
- 13) 255 (FFh) Operation Complete Byte

Error code : 1 byte

224 (E0h) Parameter Error

238 (EEh) Time Out Error

⁴⁵⁷ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

Automatic Cal Disable – Control Byte #240 (F1h)

Description: Disable automatic calibration.

Bytes to Follow: 1 byte

- 1: to disable automatic calibration
- 0: to enable automatic calibration

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error : Invalid Parameter
 - 238 (Eeh) Time-out Error
-

Recall Sweep Trace – Control Byte #243 (F3h)

Description: This command is similar to another recall sweep trace with control byte #33 (21h). The only thing difference between this command and command #33 (21h) is that this command requires 2 bytes to follow whereas command #33 (21h) requires 1 byte. This makes it possible to recall traces whose indices are bigger than 255, which is not possible with command #33 (21h).

Bytes to Follow: 2 bytes

- 0 = Last sweep trace before entering remote mode (sweep trace in RAM)
- 1- 300 = Specific saved sweep number (stored sweeps in Flash memory)

- 1) Trace index (Lower byte)
- 2) Trace index (Higher byte)

Cell Master Returns:

Exactly like command #33 (21h), so please refer to that section.

Exit Remote Mode – Control Byte #255 (FFh)

Description: Spectrum Master exits remote mode.

The computer sends the Exit Remote command #255 (FFh) to the Spectrum Master. Spectrum Master returns a confirm flag (FFh). The Spectrum Master resumes sweeping, either continuously or singly.

You may also press the “ESCAPE” key on the Spectrum Master key pad to exit from remote mode (given that the serial communication is still in sync). In this case, the Spectrum Master does not return a confirm byte to the serial port.

When exiting remote mode, system parameters changed during remote mode are used immediately.

System parameters changed during remote mode are not written to the non-volatile EEPROM.

You may want to save the change to the run-time setup (saved setup location 0, which holds the power-on setup) or one of the saved setups for the current measurement mode. See control byte #18 (12h) for details.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete
-

Select SPA/Power Meter Signal Standard – Control Word (A103h)

Description: Selects a Signal Standard. Use this command for both Spectrum Analyzer and Power Meter modes.

Bytes to Follow: 1 byte

- 1) Signal Standard – See the section “Signal Standards” for a list of standards and their indices.

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid signal standard
 - 238 (EEh) Time Out Error
-

Select SPA/Power Meter Channel – Control Word (A104h)

Description: Selects a channel within the range of the currently selected signal standard. Use this command for both Spectrum Analyzer and Power Meter modes.

See the section “Signal Standards” for a list of valid channels for the selected channel.

Bytes to Follow: 2 bytes

- 1) Channel (Higher byte)
- 2) Channel (Lower byte)

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid channel
 - 238 (EEh) Time Out Error
-

Read External Module Name – Control Word (A201h)

This command is available only with option 6.

Description: Returns the name of the attached frequency converter module (option 6).

For example, module name “FCN4760” will be received as:

c,46,43,4e,34,37,36,30,0,0,0,0,ff

Bytes to Follow: 0 bytes

Spectrum Master Returns: 14 bytes (success) OR 1 byte (failure)

- 1) Length of Name (12)
- 2-13) Module Name
- 14) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

OR

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Read External Module Serial Number – Control Word (A202h)

This command is available only with option 6.

Description: Sets the serial number of the attached frequency converter module (option 6).

For example, serial number 12345678 will be received as:

8,1,2,3,4,5,6,7,8,ff

Bytes to Follow: 0 bytes

Spectrum Master Returns: 10 bytes

- 1) Length of Serial Number (8)
- 2-9) Serial Number
- 10) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

OR

- 1) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

Read External Module Frequency Range – Control Word (A203h)

This command is available only with option 6.

Description: Sets the frequency range of the attached frequency converter module (option 6). Frequency values are scaled by the scale factor value.

For example, the frequency range of the FCN4760 is as follows:

Scale factor: 10

Input Start Frequency: 4700 MHz (scaled, this number is 470 MHz)

Input End Frequency: 6000 MHz (scaled, this number is 600 MHz)

Output Start Frequency: 450 MHz (scaled, this number is 45 MHz)

Output End Frequency: 1750 MHz (scaled, this number is 600 MHz)

So the response will look like:

12,0,a,1c,3,a1,80,23,c3,46,0,2,ae,a5,40,a,6e,49,c0,ff

Bytes to Follow: 0 bytes

Spectrum Master Returns: 20 bytes (success) OR 1 byte (failure)

- 1) Length of Frequency Data (18)
- 2-3) Scale Factor (in Hz)
- 4-7) Input Start Frequency (scaled by Scale Factor)
- 8-11) Input End Frequency (scaled by Scale Factor)
- 12-15) Output Start Frequency (scaled by Scale Factor)
- 16-19) Output End Frequency (scaled by Scale Factor)
- 20) 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

OR

- 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

Read Module Fail Counter – Control Word (A204h)

This command is available only with option 6.

Description: Returns the value of the module lock fail counter.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 3 bytes (success) OR 1 byte (failure)

- 1) Fail Counter (Higher byte)
- 2) Fail Counter (Lower byte)
- 3) 255 (FFh) Operation Complete Byte

OR

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Clear Module Fail Counter – Control Word (A205h)

This command is available only with option 6.

Description: Sets the module lock fail counter to 0.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Perform Transmission Mode Calibration – Control Word (A301h)

This command is available only with option 21.

Description: Perform Transmission Mode Calibration.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Invalid channel
 - 238 (EEh) Time Out Error
-

Turn OFF Transmission Mode Calibration – Control Word (A302h)

This command is available only with option 21.

Description: Turn OFF Transmission Mode Calibration

Bytes to Follow: 0 bytes

Spectrum Master Returns: 1 byte

- 1) 255 (FFh) Operation Complete Byte
 - 238 (EEh) Time Out Error
-

Get Signal Standard Name – Control Word (A501h)

Description: Get the ASCII signal standard corresponding to a specified Index. This command can be used in any measurement mode.

Bytes to follow: 2

- 1) Index (Highest Byte)
- 2) Index (Lowest Byte)

Cell Master Returns: 20 bytes

1 – 20) Standard Name in ASCII

Perform Noise Diode Cal – Control Word (A505h)

Description: Performs noise diode calibration on SPA board

Bytes to follow: 0

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Invalid channel
238 (EEh) Time Out Error
-

Set Bias T Voltage – Control Word (A506h)

Description: Motherboards beginning with 64968 have a programmable Bias T. This command sets the Bias T voltage between 12 and 24 volts.

Bytes to follow: 1

- 1) Bias T Voltage

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Invalid channel
238 (EEh) Time Out Error
-

Select Function in IA Measurement Mode – Control Word (A700h)

This command is available only with option 25.

Description: Selects measurement function in Interference Analysis mode.

Bytes to Follow: 1 bytes

- 1) Function ID (0:Spectrum; 1:Spectrogram; 2: Signal strength; 3: RSSI; 4: Signal ID)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
224 (E0h) Parameter Error: Module not attached
238 (EEh) Time Out Error
-

Spectrogram: Set Sweep Interval – Control Word (A721h)

This command is available only with option 25.

Description: Sets the sweep interval in spectrogram mode.

Bytes to Follow: 2 bytes

- a. Sweep interval in seconds (MSB)
- b. Sweep interval in seconds (LSB)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Spectrogram: Set Time Span of Measurement – Control Word (A722h)

This command is available only with option 25.

Description: Sets the time span of spectrogram measurement. Maximum time span is 72 hours (4320 minutes) when “Auto Save” is turned on. Minimum time span is 0 which means the fastest sweep time of current setting is used.

Bytes to Follow: 2 bytes

- 1) Time span in minutes (MSB)
- 2) Time span in minutes (LSB)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Spectrogram: Turn On/Off Auto Save – Control Word (A723h)

This command is available only with option 25.

Description: Turns On or Off Auto Save switch of spectrogram mode. When Auto Save is turned on, the first 5 screens of records are saved automatically into 5 memory slots. Once all 5 memory slots have been occupied, Auto Save is going to be turned off.

Bytes to Follow: 1 bytes

On/Off Switch (0:Off; 1:On)

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Spectrogram: Get Trace Name – Control Word (A724h)

This command is available only with option 25.

Description: Get the Trace names saved in five spectrogram memory slots. The name is composed with Time & Date when the trace is saved. If the memory slot is empty, the date field is set with "--/--/----" and the time field is set with "--:--:-".

Bytes to Follow: 0 bytes

Cell Master Returns:

- When control word is received correctly: 101 bytes
 - 1-2) Index of trace (from 0 to 4)
 - 3-12) Date of save in ASCII string, format: "--/--/----"
 - 13-20) Time of save in ASCII string, format: "--:--:--"
 - 21-100) Repeat the information of 1) to 20) four times
 - 101) FFh
 - When error occurs: 1 byte
 - 224 (E0h) Parameter Error: Module not attached
 - 238 (EEh) Time Out Error
-

Spectrogram: Recall Trace – Control Word (A725h)

This command is available only with option 25.

Description: Recall a spectrogram trace by sending the trace index (0-4) of the memory slots.

Bytes to Follow: 1 bytes

- 1) Index of memory slots (0-4)

Cell Master Returns:

When control word is received correctly: 32448 bytes

- | | |
|-----------|---|
| 1-10) | Date of save (ASCII, format: "--/-/----") |
| 11-18) | Time of save (ASCII , format: "--:--:--") |
| 19-22) | Center Frequency (Integer – MSB to LSB) |
| 23-26) | Span (Integer – MSB to LSB) |
| 27-30) | RBW (Integer – MSB to LSB) |
| 31-34) | VBW (Integer – MSB to LSB) |
| 35-38) | Reference level (Integer – MSB to LSB) |
| 39-42) | Scale (Integer – MSB to LSB) |
| 43-46) | Time Span (Integer – MSB to LSB) |
| 47-48) | Sweep Interval (Integer – MSB to LSB) |
| 49-52) | GPS Position – Latitude (long integer) ⁴⁵⁸ |
| 53-56) | GPS Position – Longitude (long integer) |
| 57-58) | GPS Position – Altitude (short integer) |
| 59-32458) | 80 records of spectrogram data. Each record has the following format:
1-401) Color indices of 401 sweep data points, The formula of color index is as following:
Color Index = (Ref Level - SaMeasData) * 255 / (Division * 10)
402-405) Time Stamp of the record being generated. |

Status byte: 1 byte

- 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Module not attached
- 238 (EEh) Time Out Error

Set Frequency in CW GENERATOR MODE - CODE WORD(A801h)

Description: This command tells the Cell Master to set its frequency to the specified value.

Bytes to Follow: 4 bytes

- 1) Frequency value Byte[3]
- 2) Frequency value Byte[2]
- 3) Frequency value Byte[1]
- 4) Frequency value Byte[0]

Cell Master Returns: 1 byte

- 255 (FFh) Operation Complete Byte
- 224 (E0h) Parameter Error: Invalid frequency point
- 238 (EEh) Time Out Error

⁴⁵⁸ Signed long integer is used to represent latitude and longitude. Positive latitude means North hemisphere, negative latitude means South hemisphere; Positive longitude means East hemisphere, negative longitude means West hemisphere. Degree = int(abs(value)/1,000,000); Minute = (float)(abs(value)%1,000,000)/10,000

Retrieve Power in CW GENERATOR MODE - CODE WORD(A802h)

Description: This command tells the Cell Master to return the power value in dBm. The value returned will be an unsigned. The values under -0 dB will be made positive and multiplied by 1000. e.g. -73.12dBm will become 731200

Bytes to Follow: 0 bytes

Cell Master Returns: 5 bytes

- 1.) highest byte of the power value
- 2.) second byte of the power value
- 3.) third byte of the power value
- 4.) lowest byte of the power value
- 5.) 255 (FFh) Operation Complete Byte

Remote Self Test – Control Word (AA15h)

Description: Trigger the equivalent of a “key press” selftest.

Note: The response bytes will not all be returned immediately. The first 12 will be returned, then there will be a slight delay before the next 14 are returned, then a final delay while the T1/E1 selftest is performed and the final 12 bytes are returned.

Bytes to Follow: 0 bytes

Spectrum Master Returns:

No Options or Option 10 and/or Option 29: 22 bytes

- 1) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Higher byte)
- 2) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Lower byte)
- 3) Memory Check (01h: Pass, 00h: Fail)
- 4) RTC Voltage Check (01h: Pass, 00h: Fail)
- 5) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Higher byte)
- 6) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Lower byte)
- 7) Reserved
- 8) Reserved
- 9) Reserved
- 10) Reserved
- 11) SPA LO Failure Counter (Higher byte)
- 12) SPA LO Failure Counter (Lower byte)
- 13) H/W Config - Mother Board ID
- 14) H/W Config - SPA Board ID
- 15) H/W Config - T1E1 Board ID
- 16) H/W Config - PLD1 ID
- 17) H/W Config - PLD2 ID
- 18) H/W Config – T1E1 COLD ID
- 19) SPA LO Test - Status (01h: Pass, 00h: Fail, FFh: SPA board not installed)
- 20) SPA LO Test - Failed data point #
- 21) SPA LO Test - Failed LO #
- 22) End of Data (FFh)

Option 21 (+ Option 10 and/or Option 29): 28 bytes

- 1) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Higher byte)
- 2) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Lower byte)
- 3) Memory Check (01h: Pass, 00h: Fail)
- 4) RTC Voltage Check (01h: Pass, 00h: Fail)
- 5) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Higher byte)
- 6) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Lower byte)

- 7) VNA Lock Failure Counter (Higher byte)
- 8) VNA Lock Failure Counter (Lower byte)
- 9) VNA Integrator Failure Counter (Higher byte)
- 10) VNA Integrator Failure Counter (Lower byte)
- 11) SPA LO Failure Counter (Higher byte)
- 12) SPA LO Failure Counter (Lower byte)
- 13) H/W Config - Mother Board ID
- 14) H/W Config - SPA Board ID
- 15) H/W Config - T1E1 Board ID
- 16) H/W Config - PLD1 ID
- 17) H/W Config - PLD2 ID
- 18) H/W Config – T1E1 COLD ID
- 19) VNA PLL Lock Failure Test - Status (01h: Pass, 00h: Fail)
- 20) VNA PLL Lock Failure Test - Failed data point # (Ignore this byte if the Lock Fail Test Status was Pass)
- 21) VNA PLL Lock Failure Test - Failed PLL # (Ignore this byte if the Lock Fail Test Status was Pass)
- 22) VNA Integration Test - Status (01h: Pass, 00h: Fail)
- 23) VNA Integration Test - Failed data point # (Ignore this byte if the Integration Test Status was Pass)
- 24) VNA Integration Test - Reserved
- 25) SPA LO Test - Status (01h: Pass, 00h: Fail, FFh: SPA board not installed)
- 26) SPA LO Test - Failed data point #
- 27) SPA LO Test - Failed LO #
- 28) End of Data (FFh)

Option 6 (+ Option 10 and/or Option 29): 27 bytes

- 1) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Higher byte)
- 2) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Lower byte)
- 3) Memory Check (01h: Pass, 00h: Fail)
- 4) RTC Voltage Check (01h: Pass, 00h: Fail)
- 5) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Higher byte)
- 6) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Lower byte)
- 7) Reserved
- 8) Reserved
- 9) Reserved
- 10) Reserved
- 11) SPA LO Failure Counter (Higher byte)
- 12) SPA LO Failure Counter (Lower byte)
- 13) H/W Config - Mother Board ID
- 14) H/W Config - SPA Board ID
- 15) H/W Config - T1E1 Board ID
- 16) H/W Config - PLD1 ID
- 17) H/W Config - PLD2 ID
- 18) H/W Config – T1E1 COLD ID
- 19) SPA LO Test - Status (01h: Pass, 00h: Fail, FFh: SPA board not installed)
- 20) SPA LO Test - Failed data point #
- 21) SPA LO Test - Failed LO #
- 22) Module PLD Version
- 23) Module Attached
- 24) Module Lock (01h = Locked, 00h = Not Locked)
- 25) Module Lock Fail Counter (Higher byte)
- 26) Module Lock Fail Counter (Lower byte)
- 27) End of Data (FFh)

Option 6 + Option 21 (+ Option 10 and/or Option 29): 33 bytes

- 1) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Higher byte)
- 2) Temperature in 1/10th of degree Celsius (e.g. 362 = 36.2 °C) (Lower byte)

- 3) Memory Check (01h: Pass, 00h: Fail)
 - 4) RTC Voltage Check (01h: Pass, 00h: Fail)
 - 5) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Higher byte)
 - 6) Power Voltage in 1/10ths of a Volt (e.g. 124 = 12.4 Volts) (Lower byte)
 - 7) VNA Lock Failure Counter (Higher byte)
 - 8) VNA Lock Failure Counter (Lower byte)
 - 9) VNA Integrator Failure Counter (Higher byte)
 - 10) VNA Integrator Failure Counter (Lower byte)
 - 11) SPA LO Failure Counter (Higher byte)
 - 12) SPA LO Failure Counter (Lower byte)
 - 13) H/W Config - Mother Board ID
 - 14) H/W Config - SPA Board ID
 - 15) H/W Config - T1E1 Board ID
 - 16) H/W Config - PLD1 ID
 - 17) H/W Config - PLD2 ID
 - 18) H/W Config – T1E1 COLD ID
 - 19) VNA PLL Lock Failure Test - Status (01h: Pass, 00h: Fail)
 - 20) VNA PLL Lock Failure Test - Failed data point # (Ignore this byte if the Lock Fail Test Status was Pass)
 - 21) VNA PLL Lock Failure Test - Failed PLL # (Ignore this byte if the Lock Fail Test Status was Pass)
 - 22) VNA Integration Test - Status (01h: Pass, 00h: Fail)
 - 23) VNA Integration Test - Failed data point # (Ignore this byte if the Integration Test Status was Pass)
 - 24) VNA Integration Test - Reserved
 - 25) SPA LO Test - Status (01h: Pass, 00h: Fail, FFh: SPA board not installed)
 - 26) SPA LO Test - Failed data point #
 - 27) SPA LO Test - Failed LO #
 - 28) Module PLD Version
 - 29) Module Attached
 - 30) Module Lock (01h = Locked, 00h = Not Locked)
 - 31) Module Lock Fail Counter (Higher byte)
 - 32) Module Lock Fail Counter (Lower byte)
 - 33) End of Data (FFh)
-

Trigger Sweep – Control Word (AA30h)

Description: Causes the Spectrum Master to perform a sweep if it is in single sweep mode.

This command works only when the Spectrum Master is NOT in remote mode. Send this command, receive the “Operation Complete Byte” and then wait for the “Sweep Complete Byte” to signify the end of the sweep.

Note: The "Sweep Complete Byte" is not returned unless serial echo status is turned on using command #10.

Bytes to Follow: 0 bytes

Spectrum Master Returns: 2 bytes

- 1) 255 (FFh) Operation Complete Byte (when the command is received)
 - 2) 192 (C0h) Sweep Complete Byte (at the end of the sweep)
-

Parameter Definitions

Parameter	# of bytes	Step	Example / Description
Frequency	4 bytes unsigned	1 Hz	1000.3 MHz = 1000300000
Scale (RL, CL)	2 bytes unsigned	1 / 1000 dB	51.3 dB = 51300
(SWR)	2 bytes unsigned	1 / 1000 (ratio)	65.53 = 65530
Limit (RL, CL)	2 bytes unsigned	1 / 1000 dB	51.3 dB = 51300
(SWR)	2 bytes unsigned	1 / 1000 (ratio)	65.53 = 65530
Markers (Frequency & distance marker)	2 bytes unsigned	1 sweep point	Marker Values are given in relative position of the graph. The lowest value is 0, while the highest is (# of data points - 1).
Distan-e	4 bytes unsigned	1/100,000 m/ft	12.34 m = 1234000
Relative Propagation Velocity	4 bytes unsigned	1 / 100,000	0.837 = 83700
Cable Loss	4 bytes unsigned	1 / 100,000 dB	-0.345 dB/m = 34500
Gamma	4 bytes signed	1 / 10,000 (ratio)	Gamma value is the ratio of magnitude of reflected signal over the magnitude of incident signal.
Phase	4 bytes signed	1 / 10 degree	Phase value is the difference in phase between the incident and reflected signal.
Power: dBm/dB	4 bytes signed	1 / 1000 dBm 1 / 1000 dB	51.3 dBm = 51300 10.4 dB = 10400
Lock Fail Counter	2 bytes unsigned	1 error count	234 fails = 234
Integrator Fail Counter	2 bytes unsigned	1 error count	123 fails = 123

Programming Examples

This section contains several sample functions written in C, (and one in Visual Basic) that can be used as references when programming the Anritsu Handheld Products. These include functions to set up the com Port, enter and exit remote mode, and set the reference level of the spectrum analyzer.

Examples in C:

```
*****  
/*     unsigned char EnterRemote(BYTE *ResponseBytes)          */  
/*     Description: This function implements control byte #69, Enter   */  
/*                  Remote Mode. If successful, the unit will be in    */  
/*                  remote mode, waiting to accept additional serial    */  
/*                  commands.                                         */  
/*     Inputs :      ResponseBytes = pointer to an array of bytes at   */  
/*                      least 13 elements long (13 bytes are expected in   */  
/*                      response to the Enter Remote command).           */  
/*     Returns:      SUCCESS if the unit is in remote mode            */  
/*                  FAILURE if the command fails                     */  
/*                  Response bytes are returned in the variable       */  
/*                  ResponseBytes.                                */  
*****  
unsigned char EnterRemote(BYTE *ResponseBytes)  
{  
    BYTE *SendEnterRemoteCharPointer;    // Data to send  
    BYTE SerialCommand;  
  
    SendEnterRemoteCharPointer = &SerialCommand;  
    SerialCommand = 69; // 69 is the Enter Remote Mode serial command  
  
    // Write 1 byte of data from SendEnterRemoteCharPointer to the  
    // COM Port  
    WriteToPort(SendEnterRemoteCharPointer, 1);  
  
    // Read the data returned by the SiteMaster - expecting 13 bytes,  
    // give the unit 30 seconds to respond before timing out.  
    if(!ReadfromPort(13, ResponseBytes, 30))  
    {  
        return FAILURE;  
    }  
    else  
    {  
        return SUCCESS;  
    }  
} /* EnterRemote */  
  
*****  
/*     unsigned char SetSPAScale(unsigned long ReferenceLevel,      */  
/*                               unsigned long dBScale, BYTE *ResponseBytes) */  
/*     Description: This function implements control byte #101, Set    */  
/*                  Spectrum Analyzer Scale. It sets the spectrum       */  
/*                  analyzer reference level and scale (dB/div).       */  
/*     Inputs :      RefLevel = reference level value                */  
/*                  dBScale = scale value                            */  
*****
```

```

/*
 *          NOTE: This function assumes the values have      */
 *          already been checked to fall in the valid range   */
 *          and scaled according to the formulas in the      */
 *          Programming Manual.                            */
 *          ResponseBytes = pointer to an array of bytes at   */
 *          least 1 element long (1 byte is expected in       */
 *          response to the Set Spectrum Analyzer Scale      */
 *          command).                                         */
 *          Returns:    SUCCESS if the values are set        */
 *                      FAILURE if the command fails         */
 *          Response bytes are returned in the variable     */
 *          ResponseBytes.                                */
 ****
unsigned char SetSPAScale(unsigned long RefLevel,
                           unsigned long dBScale, BYTE *ResponseBytes)
{
    BYTE *SendScalePointer;           // Data to send
    BYTE SendBytes[9];
    BYTE SerialCommand;

    // Serial Command to Set Scale on the SPA.
    SerialCommand = 101;

    // Data pointer.
    SendScalePointer = &SendByte[0];

    // First byte to send is the serial command, #101.
    SendBytes[0] = SerialCommand;

    // Convert the reference level and scale into 8 bytes
    // (4 bytes each) for the SPA. Put the bytes in the
    // SendBytes variable, starting with byte 1 (leave byte 0
    // as the command byte).
    Get8Bytes(RefLevel, Scale, &SendBytes[1]);

    // Write 9 bytes of data in SendScalePointer to the port.
    WriteToPort(SendScalePointer, 9);

    // Expecting 1 byte back (give the unit 5 seconds to respond):
    // 0xFF = success
    // 0xE0 = parameter failure (invalid value)
    // 0xEE = time-out (insufficient # of bytes received by SPA)
    if(!ReadFromPort(1, ResponseBytes, 5))
    {
        return FAILURE;
    }
    else
    {
        if(*ResponseBytes != 0xFF)
        {
            return FAILURE;
        }
        else
        {
            return SUCCESS;
        }
    }
}

```

```

        }
    } /* SetSPAScale */

/************************************************************************/
/*      unsigned char ExitRemote(BYTE *ResponseBytes)                  */
/*      Description: This function implements control byte #255, Exit   */
/*                      Remote Mode.  If successful, the unit will leave   */
/*                      remote mode and resume sweeping.                 */
/*      Inputs :      ResponseBytes = pointer to an array of bytes at   */
/*                      least 1 element long (1 byte is expected in       */
/*                      response to the Exit Remote command).           */
/*      Returns:      SUCCESS if the unit exits remote mode             */
/*                      FAILURE if the command fails                   */
/*                      Response bytes are returned in the variable     */
/*                      ResponseBytes.                                */
/************************************************************************/
unsigned char ExitRemote(BYTE *ResponseBytes)
{
    BYTE *SendExitRemoteCharPointer;      // Data to send
    BYTE SerialCommand;

    SendExitRemoteCharPointer = &SerialCommand;
    SerialCommand = 255; // 255 is the Exit Remote Serial Command

    // Write 1 byte of data from SendExitRemoteCharPointer to the
    // COM Port
    WriteToPort(SendExitRemoteCharPointer, 1);

    // Expecting 1 byte back (give the unit 5 seconds to respond):
    // 0xFF = success
    if(!ReadFromPort(1, ResponseBytes, 1))
    {
        return FAILURE;
    }
    else
    {
        if(*ResponseBytes != 0xFF)
        {
            return FAILURE;
        }
        else
        {
            return SUCCESS;
        }
    }
} /* ExitRemote */

/************************************************************************/
/*      void Get8Bytes(unsigned long parm1, unsigned long parm2,          */
/*                      BYTE* ByteData )                                     */
/*      Description: This function converts the 2 four byte values to   */
/*                      8 bytes for transmission to the SiteMaster.  parm1   */
/*                      occupies the first four bytes, parm2 occupies the  */
/*                      second 4 bytes.                                    */
/*      Inputs:      parm1 - 4 byte unsigned long integer                */
/************************************************************************/

```

```

/*
 *      Returns:      parm2 - 4 byte unsigned long integer          */
/*           SUCCESS if the unit is in remote mode                  */
/*           FAILURE if the command fails                         */
/*           The resulting bytes are returned in the             */
/*           memory location pointed to by ByteData. This       */
/*           location must have at least 8 empty bytes.          */
/***************************************************************/
void Get8Bytes(unsigned long parm1, unsigned long parm2,
               BYTE* ByteData)
{
    // MSB of 1st parameter
    *ByteData = (BYTE)((parm1 & 0xFF000000)>>24);
    *(ByteData+1) = (BYTE)((parm1 & 0x00FF0000)>>16);
    *(ByteData+2) = (BYTE)((parm1 & 0x0000FF00)>>8);
    // LSB of 1st parameter
    *(ByteData+3) = (BYTE)(parm1 & 0x000000FF);

    // MSB of 2nd parameter
    *(ByteData+4)= (BYTE)((parm2 & 0xFF000000)>>24);
    *(ByteData+5)= (BYTE)((parm2 & 0x00FF0000)>>16);
    *(ByteData+6)= (BYTE)((parm2 & 0x0000FF00)>>8);
    // LSB of 2nd parameter
    *(ByteData+7) = (BYTE)(parm2 & 0x000000FF);
} /* Get8Bytes */

/***************************************************************/
BOOL OpenCommunications(int ComPort, int ComBaud)
/*
 * Description : This function is to Open the communication port   */
/* and set the port settings                                     */
/* Inputs :      int - ComPor- entered as a command line argument */
/*                int - ComBau- The Baud rate for Communication   */
/* Returns:        SUCCESS - If the-Communication link was established*/
/*                FAIL - IF the-e was an error opening the COM Port */
/***************************************************************/
BOOL OpenCommunications(int ComPort, int ComBaud)
{
    DCB     CommSettings; // Structure with COM Port settings
    LPCTSTR ComPortNumber; // Pointer to the COM port number
    BOOL PortReady;        // Return val after setting the COM Port
    COMMTIMEOUTS timeout; // Structure with Time out values

    switch (ComPort)
    {
        case '1':
            ComPortNumber = "COM1";
            break;
        case '2':
            ComPortNumber = "COM2";
            break;
        case '3':
            ComPortNumber = "COM3";
            break;
        case '4':
            ComPortNumber = "COM4";
            break;
    }
}

```

```

        default:
            CloseHandle(ComHandle);
            fclose(fp);
            exit(0);
        break;
    }

/* Creating a File to Open a COM Port*/
ComHandle = CreateFile( ComPortNumber,
                        GENERIC_READ | GENERIC_WRITE,
                        0, // exclusive access
                        NULL, // no security
                        OPEN_EXISTING,
                        0, // no overlapped I/O
                        NULL); // null template

/* Set up the COM Ports Input and Output Buffer
Syntax -
BOOL-SetupComm(
HANDLE hFile,      // handle to communications device
DWORD dwInQueue,  // size of input buffer
DWORD dwOutQueue // size of output buffer
);
*/
PortReady = SetupComm(ComHandle, 5000, 5000);

/* Open the existing COM Settings
Syntax -
BOOL-GetCommState(
    HANDLE hFile,      // handle to communications device
    LPDCB lpDCB        // pointer to device-control block
                           // structure
);
*/
PortReady = GetCommState(ComHandle, &CommSettings);

/*Check to see if it was successful*/
if(!PortReady)
{
    CloseHandle(ComHandle);
    fclose(fp);
    exit(0);
}

/* This is used to update the CommSettings Structure Variables*/
// Setting the Baud Rate
switch (ComBaud)
{
    case '1':
        CommSettings.BaudRate = CBR_9600; // rate - 9600
    break;
    case '2':
        CommSettings.BaudRate = CBR_19200; // rate - 19200
    break;
    case '3':
        CommSettings.BaudRate = CBR_38400; // rate - 38400
    break;
}

```

```

        case '4':
            CommSettings.BaudRate = CBR_56000; // rate - 56000
        break;
        case '5':
            CommSettings.BaudRate = CBR_115200; // rate - 115200-
        break;
        default:
            CommSettings.BaudRate = CBR_9600; //Default - 9600
        break;
    }

    // disable null stripping
    CommSettings.fNull = FALSE;
    // RTS flow control
    CommSettings.fRtsControl = RTS_CONTROL_ENABLE;
    // XON/XOFF in flow control
    CommSettings.fInX = FALSE;
    // XON/XOFF out flow control
    CommSettings.fOutX = FALSE;
    // DTR flow control type
    CommSettings.fDtrControl = DTR_CONTROL_ENABLE;
    // number of bits/byte, 4-8
    CommSettings.ByteSize = 8;
    // 0-4=no,odd,even,mark,space
    CommSettings.Parity = NOPARITY;
    // 0,1,2 = 1, 1.5, 2
    CommSettings.StopBits = ONESTOPBIT;

    /* Setting the COM State with the changed parameters
    Syntax -
        BO-L SetCommState(
            HANDLE hFile, // handle to communications device
            LPDCB lpDCB // pointer to device-control block structure
        );
    */

    PortReady = SetCommState (ComHandle, &CommSettings);

    /* Setting the parameters for the timeouts.
    NOTE: Without Timeout Settings, Reading the COM Port will not work
properly*/

    // This gives the Timeout value for each bytes received
    timeout.ReadIntervalTimeout = MAXDWORD;
    timeout.ReadTotalTimeoutConstant = 0;
    timeout.ReadTotalTimeoutMultiplier = 0;

    /* Sets the communication timeouts
    Syntax -
        BOO- SetCommTimeouts(
            HANDLE hFile, // handle to comm dev omm.
            LPCOMMTIMEOUTS lpCommTimeouts /* pointer to comm tim omm.t structure */
        );
    */
    SetCommTimeouts(ComHandle, &timeout);

    if(PortReady)

```

```
{\n        return SUCCESS;\n    }\n    else\n    {\n        CloseHandle(ComHandle);\n        fclose(fp);\n        return FAIL;\n        exit(0);\n    }\n}
```

Example in Visual Basic

```
Private Sub cmdSetBaudRateSM_Click()
    Dim ChangeBaudSerialCmd As Integer
    Dim BaudRate As Integer
    Dim strInputBuf As Variant
    Dim PreviousSettings As String

    PreviousSettings = commCtrl.Settings

    'Check that we're in remote and have selected a baud rate
    If CheckInitialConditions(True, False, True) = False Then
        GoTo SetsMBAud_err_handler
    End If

    ChangeBaudSerialCmd = 197           'Setting Baud rate Serial Command
    BaudRate = GetBaudSerialCmd         'Get the Serial cmd for the specific
                                       'baud rate
    commCtrl.Output = Chr$(ChangeBaudSerialCmd) + Chr$(BaudRate)  'Sending
                                                               'the data

    Delay (300)

    'Change the Baud setting for the application also
    If BaudRate = 0 Then
        commCtrl.Settings = "9600,n,8,1"
    ElseIf BaudRate = 1 Then
        commCtrl.Settings = "19200,n,8,1"
    ElseIf BaudRate = 2 Then
        commCtrl.Settings = "38400,n,8,1"
    ElseIf BaudRate = 3 Then
        commCtrl.Settings = "56000,n,8,1"
    ElseIf BaudRate = 4 Then
        commCtrl.Settings = "115200,n,8,1"
    Else
        'Box will fail, set back to 9600.
        commCtrl.Settings = "9600,n,8,1"
    End If

    Delay (1000)
    strInputBuf = CStr(commCtrl.Input)
    strInputBuf = Mid(strInputBuf, 1, 1)
    If strInputBuf = "" Then
        MsgBox "Invalid Baud Rate - NO STRING"
        GoTo SetsMBAud_err_handler
    End If

    If Asc(strInputBuf) = 255 Then
        MsgBox "Set Baud Rate Succesfully"
    ElseIf Asc(strInputBuf) = 238 Then
        MsgBox "SiteMaster Timed out"
        GoTo SetsMBAud_err_handler

    ElseIf Asc(strInputBuf) = 224 Then
        MsgBox "Invalid Baud Rate - ERR 22-"
        GoTo SetsMBAud_err_handler
    End If
End Sub
```

```
Else
    MsgBox "Invalid Baud Rate - ERR " + CStr(Asc(strInputBuf))
    GoTo SetSMBaud_err_handler

End If

Exit Sub
SetSMBaud_err_handler:
    commCtrl.Settings = PreviousSettings
End Sub
```

Revision History

Rev.	Date	Description	By
1.0	9/1/03	Initial revision based on Spectrum Master Internal Programming Manual.doc rev 1.36.	Jennifer Burns
1.01	9/3/03	Completed conversion to MS2711D document. Left commands for option 21, 29 and 50. Still needs AM/FM demod and normalization commands.	Jennifer Burns
1.02	9/3/03	Added Control Word A216 – Set module cal points.	Jennifer Burns
1.03	9/3/03	Copied normalization and AM/FM demod commands from Cell Master doc.	Jennifer Burns
1.04	9/9/03	Added Transmission mode to Set Model/Option command.	Jennifer Burns
1.05	9/12/03	Changed S21 to Transmission	Calvin Carter
1.06	9/16/03	Fixed case of TRANSMISSION throughout the document.	Jennifer Burns
1.07	9/18/03	Added frequency range min and max to #29, #33 and #36.	Jennifer Burns
1.08	9/19/03	Added commands to read and clear module fail counter.	Jennifer Burns
1.09	9/24/03	Changes to signal standards. Added Programming Examples. Miscellaneous changes based on S33xD PM editing.	Jennifer Burns
1.10	10/1/03	Added info to commands that require box to be unlocked before executing.	Jennifer Burns
1.11	10/8/03	Added error for module/SPA cal freq mismatch to #104.	Jennifer Burns
1.12	10/9/03	Changed 139 to catch more attenuator failures earlier in production	Derek Truesdale
1.13	10/10/03	Added C/I data to #33 and #36.	Jennifer Burns
1.14	10/17/03	Remote Selftest 0xAA15 is no longer an internal command. Added option 6 results.	Jennifer Burns

1.15	10/21/03	Added ADC_OTR_ERROR to #104, changed error numbering for MODULE_FREQUENCY_ERROR.	Jennifer Burns
1.16	10/21/03	Added OBW status bit to #29, #33 and #36.	Jennifer Burns
1.17	10/27/03	Added reference to frequency scale factor and option 6 in appropriate commands. Added scale factor and frequency range min/max to power meter versions of #33, #36 and #29.	Jennifer Burns
1.18	10/27/03	Added "End of Data" to Option 6 Remote Selftest (AA15h).	Jennifer Burns
1.19	10/31/03	Added frequency scale factor to command #82 and #83 (write/recall antenna). Added freq mismatch error to #19. Added FEh as possible argument to #19 to recall the current mode's defaults only.	Jennifer Burns
1.20	11/11/03	Added command #116 – Set SPA Impedance.	Jennifer Burns
1.21	11/11/03	Fixed numbering in AA15h (Remote Selftest).	Jennifer Burns
1.22	11/13/03	Added operation complete byte to A204.	Jennifer Burns
1.23	11/20/03	Added OCC BW measured power in SPA for Control Byte #33	Vaidya Venugopal
1.24	11/24/03	Added Set One External Module Cal Point (A217).	Jennifer Burns
1.25	12/3/03	Cleanup during review for tech pubs.	Jennifer Burns
1.26	12/10/03	Added T1/E1 COLD ID to command AA15.	Jennifer Burns
1.27	1/15/04	Added Download Setup, Upload Setup, Read Parameter Limits and Query Saved Setups.	Jennifer Burns
1.28	1/16/04	Added ability to get a list of available RBWs and VBWs from Read Parameter Limits (#67). Added channel power and ACPR parameters to SPA portion of #67.	Jennifer Burns

1.29	1/20/04	Fixed numbering on #65. Added marker type to #29, 33, 36, 65 and 66. Added zero status and zero value to #29, 65 and 66. Added channel power, ACPR and IA parameters to #65 and 66. Added external reference frequency status, view b status, IA status and C/I status to #65 and 66.	Jennifer Burns
1.30	1/22/04	Both log and linear units must be sent and received, regardless of which scale type is selected (#65 and 66). This required moving the bit that represents the external reference frequency status. Added trace B id to #65 and 66. Added format of PM offset to #65 and 66.	Jennifer Burns
1.31	1/23/04	Added Write Protect Setup #71.	Jennifer Burns
1.32	2/4/04	Merge changes from S33xD PM: Add T1E1 Read Frequency, Set T1E1 Frequency Cal, and Read T1E1 Frequency Cal commands (A017, A018, A019). Added <u>Set Filter Linearity Cal Values</u> – Control Word (BB01h). Add T1/E1 Vpp, dBdsx and Frequency to commands 33 and 36. Added Appendix A : BSL Serial Commands. Added DMB to SPA signal standard table. Added #72 Clear Setup Memory Locations	Jennifer Burns
1.33	2/23/04	Added noise marker conversion to #117.	Jennifer Burns
1.34	2/26/04	Fixed numbering on AA15h.	Jennifer Burns
1.35	3/5/04	Modifications to #67 (Parameter Limits).	Jennifer Burns
1.36	3/11/04	Removed default setup from #65. Need unproven. Added attributes, cal status to #68 (query saved setups).	Jennifer Burns
1.37	3/18/04	Added error to ref level cal #104	Derek Truesdale
1.38	3/19/04	Fixed impedance adapter indices in #29, #33 and #36.	Jennifer Burns

1.39	4/9/04	Added limit beep output to #29, #65 and #66. This is a “special” option so these bytes should not ship in standard documentation.	Jennifer Burns
1.40	4/21/04	Added #49 – Sweep Data Echo.	Jennifer Burns
1.41	6/23/04	In #86 and #88, channel power values are scaled by 100, not 1000.	Jennifer Burns
1.42	7/16/04	Fixed secret key in #219 (debug mode).	Jennifer Burns
1.44	1/22/07	Updated various areas of the document in attempt to bring it up to date from the last release.	Maximilian Maung
1.45	5/8/2007	Added new measurement modes in command byte #3 (03h).	Maximilian Maung
1.45	5/8/2007	Added new data information fields in command byte #29 (1Dh).	Maximilian Maung
1.45	5/8/2007	Updated command byte #33 (21h) and #36 (24h) to reflect new data fields available for each measurement mode.	Maximilian Maung
1.45	5/8/2007	Updated command byte #37 (25h) for the maximum string the instrument can return.	Maximilian Maung
1.45	5/8/2007	Added new command byte #89 (59h), #90 (5Ah), #91 (5Bh), #92 (5Ch), #237 (EDh), #238 (EEh), #240 (F1h), and #243 (F3h).	Maximilian Maung
1.45	5/8/2007	Added new command word A501h, A505h, A506h, A700h, A721h, A722h, A723h, A724h, A725h, A801h, and A802h.	Maximilian Maung