



THE AMERICAN
ASSOCIATION
FOR LABORATORY
ACCREDITATION

ACCREDITED LABORATORY

A2LA has accredited

**ANRITSU COMPANY MORGAN HILL
CALIBRATION SERVICES**
Morgan Hill, CA

for technical competence in the field of
Calibration

This laboratory is accredited in accordance with the recognized International Standard **ISO/IEC 17025:2005** *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation also demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 18 June 2005*).

Presented this 5th day of April 2006



President
For the Accreditation Council
Certificate Number 2160.01
Valid to April 30, 2008

For the calibrations to which this accreditation applies,
please refer to the laboratory's Calibration Scope of Accreditation.



American Association for Laboratory Accreditation

SCOPE OF ACCREDITATION TO ISO 17025:2005 & ANSI/NCSL Z540-1-1994

ANRITSU COMPANY MORGAN HILL CALIBRATION SERVICES

490 Jarvis Drive
Morgan Hill, CA 95037
Yeou-Song (Brian) Lee Phone: 408 201 1976

CALIBRATION

Valid To: April 30, 2008

Certificate Number: 2160.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – RF/Microwave

Parameter/Equipment	Range	Best Uncertainty ² (±)	Comments
S Parameters – Magnitude and Phase for S11, S12, S21, S22³			37XXX and 360 with 3663/3653, 3666/3650-1, 3667/3651-1, 3668/3652-1
Airline: S11/S22 S12/S21	10 MHz to 40 GHz	M/0.025 (lin); P/20 M/0.18 dB; P/20	
Beatty Airline: S11/S22 S12/S21		M/0.08 (lin); P/20 M/0.18 dB; P/20	MS 462X with 3663R/3753R, 3666R/3750R 3667R/3751R
20 dB Attenuation: S11/S22 S12/S21		M/0.025 (lin); P/20 M/0.18 dB; P/20	M represents magnitude P represents phase
40/50 dB Attenuation: S11/S22 S12/S21 S12/S21	10 MHz	M/0.025 (lin); P/20 M/0.23 dB; P/20 M/0.55 dB	

(A2LA Cert. No. 2160.01) 04/05/2006

Page 1 of 5



Parameter/Equipment	Range	Best Uncertainty ² (±)	Comments
50 Ohm Airline Characteristic Impedance	(2 to 18) GHz (18 to 40) GHz	0.1 % 0.2 %	18A50, 18N50, 18NF50, 19K50, 19KF50
VSWR	40 MHz to 40 GHz	2.5 dB	54169A, 18X50, 19X50, SM/PL, 26X50- X, 28X50-X, 29X50-X, 97X50-X, 560-97X50- X, 560-98X50
Power Sensors – Type N Connector Type K Connector	100 kHz to 40 GHz Calibration Factor: At 50 MHz At 100 kHz to 18 GHz At 50 MHz At 100 kHz to 40 GHz	0.56 % 0.92 % 1.4 % 3.6 %	MA2400X 100 kHz to 18 GHz MA2400X 100 kHz to 40 GHz
Power Level – Type N and K-Type Connector (20 to -110) dBm 30 MHz to 40 GHz 0 dBm (20 to -60) dBm (except 0 dBm)	 30 MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (13 to 18) GHz (19 to 32) GHz (33 to 40) GHz 30 MHz (50 to 150) MHz (0.15 to 2) GHz (2 to 12) GHz (13 to 18) GHz (19 to 32) GHz (33 to 40) GHz	 0.87 dB 0.22 dB 0.17 dB 0.27 dB 0.3 dB 0.4 dB 0.54 dB 0.88 dB 0.24 dB 0.19 dB 0.28 dB 0.32 dB 0.42 dB 0.56 dB	Direct power measurement (for type N and K-type connector), MA 247XA/B with ML 2437/8A

Peter Mlynar

Parameter/Equipment	Range	Best Uncertainty ² (±)	Comments
Frequency Modulation – Measure (cont) Rate: 20 Hz to 200 kHz, ≤400 kHz peak	(10 to 1300) MHz	5.8 % + 1 digit	8902 Measuring receiver 0 to 3.999: 1 digit = 1 Hz
Amplitude Modulation – Measure Rate: 50 Hz to 10 kHz, 5% to 99% Rate: 20 Hz to 10 kHz, to 99% Rate: 50 Hz to 50 kHz, 5% to 99% Rate: 20 Hz to 100 kHz, to 99%	(0.15 to 10) MHz (10 to 1300) MHz	2.3 % + 1 digit 3.5 % + 1 digit 1.2 % + 1 digit 3.5 % + 1 digit	8902 measuring receiver: 10 % to 99.9 %: 1 digit = 0.1 % 0 % to 9.99 %: 1 digit = 0.01%
Attenuation – Coaxial Type K (2.92 mm)	(0 to 40) dB; (0.05 to 2.0 GHz) (0 to 40) dB; (2.0 to 20.0 GHz) (0 to 40) dB; (20.0 to 40.0 GHz) (40 to 60) dB; (0.05 to 2.0 GHz) (40 to 60) dB; (2.0 to 20.0 GHz) (40 to 60) dB; (20.0 to 40.0 GHz)	0.18 dB 0.18 dB 0.18 dB 0.18 dB 0.18 dB	VNA

Peter Meyer

Parameter/Equipment	Range	Best Uncertainty ² (±)	Comments
Attenuation – (cont)			Measuring receiver
Coaxial Type-N (10 to 1300) MHz	0 dB 10 dB 20 dB 30 dB 40 dB 50 dB 60 dB 70 dB 80 dB 90 dB 100 dB 110 dB	0.026 dB 0.026 dB 0.034 dB 0.045 dB 0.055 dB 0.065 dB 0.074 dB 0.088 dB 0.096 dB 0.098 dB 0.110 dB 0.150 dB	Note: mismatch uncertainty not included
(1300 to 2000) MHz	0 dB 10 dB 20 dB 30 dB 40 dB 50 dB 60 dB 70 dB 80 dB 90 dB 100 dB 110 dB	0.035 dB 0.035 dB 0.053 dB 0.074 dB 0.096 dB 0.130 dB 0.150 dB 0.170 dB 0.190 dB 0.230 dB 0.250 dB 0.270 dB	

¹ This laboratory offers commercial and on-site calibration services.

² “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

³ On-site calibration service is available for this calibration. The uncertainties achievable on a customer's site can normally be expected to be larger than the Best Measurement Capabilities (BMC) that the accredited laboratory has been assigned as Best Uncertainty on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the BMC.