



NOT to be distributed outside of FM Approvals and its affiliates except by CUSTOMER.

APPROVAL REPORT

**Ex-Pt 720
TEMPERATURE MEASURING UNIT
FOR USE IN HAZARDOUS (CLASSIFIED)
LOCATIONS**


Prepared for:

**ecom instruments
Industriestraße 2
D-97959 Assamstadt**

Project ID: 3015067

Class: 3610

Date of Approval: October 25, 2002

Authorized by: 
G. A. Smith, Group Manager

FM Approvals
1151 Boston-Providence Turnpike
PO Box 9102
Norwood, MA 02062



FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

Ex-Pt 720 Temperature Measuring Unit

IS / I / 1 / ABCD / T4, Ta = 50 °C
I / 0 / AEx ia IIC T4, Ta = 50 °C

Special Condition of use;

1. Must be used with probes 0628-1232, 0628-1932, 0628-2232

Equipment Ratings:

Intrinsically safe for Class I, Division 1, Groups A, B, C, and D hazardous (classified) locations temperature class T4 at a 50°C ambient and intrinsically safe for Class I Zone 0 Group IIC hazardous (classified) locations temperature class T4 at a 50°C ambient.

Approved for:

ecom instruments
Industriestraße 2, D-97959 Assamstadt, Germany



This certifies that the equipment described has been found to comply with the following FM Approval Standards and other documents:

Class 3600	1998
Class 3610	1999
Class 3810	1989

Original Project ID: 3015067

FM Approval Granted: October 25, 2002

Subsequent Revision Reports / Date FM Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Global Technologies LLC

A handwritten signature in black ink, appearing to read "N. Ludlam".

Nicholas P. Ludlam
Technical Team Manager

November 20, 2002
Date