

Translation

Physikalisch-Technische Bundesanstalt  
(Federal Institute of Physics and Metrology)

1st SUPPLEMENT

to the Certificate of Conformity PTB No. Ex-90.C.2011

of the company ECOM Rolf Nied GmbH  
D-6971 Assamstadt

The field calibrator type ExMAG 20 is extended by the types ExMAG 21 and ExMAG 24. The new types differ from the original type due to their internal and external structure (ExMAG 24).

The maximum ambient temperature is 45° C for the new types.

All other data remain unchanged.

Test records

signed on

1.	Description (29 pages)	20 December 1991
2.	Drawing no. EXMAG21S4	21 April 1992
	EXMAG21S5	21 April 1992
	EXMAG21S1	21 April 1992
	EXMAG21S2	21 April 1992
	AM-MAG21 5	12 March 1992
	AM-MAG21 1	12 March 1992
	AM-MAG21 2	12 March 1992
	AM-MAG21-4	26 February 1992
	EXMAG24-4	23 April 1992
	EXMAG24 01	23 April 1992
	EXMAG24 02	23 April 1992
	EXMAG24 5	23 April 1992
	EXMAG 24 15	20 December 1991
	AKULA 6	09 January 1992
	AKULA 4	13 January 1992
	AKULA 1	13 January 1992
	AKULA 7	13 January 1992

By Order

Braunschweig, 1st September 1992

(Signature)

(Stamp)

Dr.-Ing. Schebsdat  
Regierungsdirektor

Page 1/1

Certified English Translation

The translation of the official document issued in German language is TRUE, CORRECT and COMPLETE to the best of my knowledge.

Sersheim, 8th December 1992

*Claudia Salamon*  
(Signature)

Claudia Salamon, Certified Translator (BDÜ)  
Alleestraße 17/1, 7126 Sersheim, FRG

