

Manual Supplement

Manual Title:	1594A/1595A Users	Supplement Issue:	1
Part Number:		Issue Date:	10/13
Print Date:		Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 66213

On page 9, under **General Specification** add:

Safety	IEC 61010-1: Overvoltage Category II, Pollution Degree 2
Electromagnetic Environment	IEC 61326-1: Basic