

# Manual Supplement

Manual Title:	4180, 4181 Technical Guide	Supplement Issue:	<b>2</b>
Part Number:	Web-Only	Issue Date:	3/14
Print Date:	March 2013	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.



## Change #1, 359

On page 9, under **Specifications** add Footnote 4 to the **4181 Temperature range**:

- [4] Due to physical limitations of the product, the maximum temperature of the 4181 is limited for emissivity settings greater than 0.95. See Table 3 in section 3.5.1 of the 4180, 4181 Technical Guide for more information.

## Change #2, 365

On page 21, under **System Menu, IR CALibration**, remove:

SURFACE: <edit>