



Certificate of Compliance

Certificate: 2342846

Master Contract: 162029

Project: 2404106

Date Issued: March 3, 2011

Issued to: **Fluke Corporation**
6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 8721 05 - LABORATORY EQUIPMENT - Electrical

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Calibrator, Fluke Model 5080A, cord-connected, rated 110/120, 220/240V, 47-63Hz, 600VA MAX.

Notes:

The above model is Equipment Class I, Pollution Degree 2, Measurement Category I.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD)

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 2342846

Master Contract: 162029

Project: 2404106

Date Issued: March 3, 2011

ISA-82.02.01 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment – General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements