

Manual Supplement

Manual Title:	5080A Getting Started	Supplement Issue:	1
Part Number:	3502941	Issue Date:	3/13
Print Date:	April 2010	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1

On page 1, under **Safety Information**, delete the first sentence and all the bullets.

On page 2, under the **Symbols** table, delete the CAT I information and add the following:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 5, under **General Specifications**, delete the Safety and EMC and add the following:

Safety IEC 61010-1: 300 V CAT II, Pollution Degree 2

Electromagnetic Environment IEC 61326-1: Controlled

Electromagnetic Compatibility Applies to use in Korea only Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.