

Certificate of Compliance

Certificate: 2450346 Master Contract: 162029

Project: 2696182 **Date Issued:** March 21, 2014

Issued to: Fluke Corporation

6920 Seaway Blvd Everett, WA 98203

USA

Attention: Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cíndy Johnson

Issued by: Cindy Johnson, AScT.

PRODUCTS

CLASS 8721 06 - LABORATORY EQUIPMENT - Electrical

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US

Standards

Multi-Product Calibrator, Models 5502A, 5502E and 5522A with detachable cord, rated 100, 120, 220, and 240 V, 47-63 Hz, 5A for 100/120 and 2.5 A for 220/240V, 600 VA maximum.

Notes:

• The above models are Equipment Class I, Pollution Degree 2.

• Mode of Operation: Continuous.

• Environmental Conditions: 0°C to 50°C; Humidity <80% to 30 °C, <70% to 40 °C, <40% to 50 °C.

DQD 507 Rev. 2012-05-22 Page: 1



Certificate: 2450346 Master Contract: 162029

Project: 2696182 **Date Issued:** March 21, 2014

CONDITIONS OF ACCEPTABILITY

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2010, MOD)

UL Std. No. 61010-1 (3rdEdition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Reference Standards:

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

DQD 507 Rev. 2012-05-22 Page: 2