



# Certificate of Compliance

**Certificate:** 2248267 (LR 44340C)

**Master Contract:** 162029

**Project:** 2692194

**Date Issued:** March 5, 2014

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Paul Tanaka*

**Issued by:** Paul Tanaka, ASCT.

## **PRODUCTS**

**CLASS 8721 06** - LABORATORY EQUIPMENT - Electrical

**CLASS 8721 86** - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Part A: Electrical Power Standards, Models 6100B and 6105A (Master Unit Models), 6101B and 6106A (Auxiliary Unit Models); rated 100-240V, 47-63 Hz, 15A, 1250VA.

Part B: Fluke 6135A/PMU Calibration System, rated 100-240V, 47-63Hz, 3,500VA.

## **Notes:**

1. The above models are cord connected, Equipment Class 1, Pollution Degree 2, Installation Category II, Measurement Category None.
2. Mode of operation: Continuous.
3. Environmental Conditions: Normal: Refer to Manual for details.



**Certificate:** 2248267 (LR 44340C)

**Master Contract:** 162029

**Project:** 2692194

**Date Issued:** March 5, 2014

---

### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD)

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

#### Reference Standards

IEC Standard No. 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

### **CONDITIONS OF ACCEPTABILITY**

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) Equipment is only to be installed on floor with a weight rating of min 400kg/m<sup>2</sup>
- (4) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.
- (5) Equipment is not to be used with flammable liquids.