

Manual Supplement

Manual Title:	9140 Users Guide	Supplement Issue:	1
Part Number:	3729482	Issue Date:	5/15
Print Date:	January 2013	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

FLUKE®

Calibration

Change #1, 408

On page 1, replace the DC Symbol with:

≡ DC

On page 18, under **Setting the Temperature**, replace the paragraph under the steps:

When the set-point temperature is changed the controller switches the well heater on or off to raise or lower the temperature. The displayed well temperature gradually changes until it reaches the set-point temperature. The well may require 5 to 20 minutes to reach the set-point depending on the span. Another 5 to 10 minutes is required for the temperature to stabilize.