

Instrument Security Procedures

Models:

Fluke 110, Fluke 111, and Fluke 112

Product Name:

True-rms Multimeters

Instrument Description:

110 Series Digital Multimeters.

Memory Description:

Fluke 110, 111 and 112 have two memory devices: U3, micro controller, Renesas (Hitachi prior 2003) H8/3844, 32-kbyte masked ROM, 2-kbyte RAM, contains operating code for the product; U6, Serial 1k EEPROM, contains calibration constants.

Memory Cleaning Instructions:

The user controlled storage operation, other than calibration constants, is the storage of Min Max and Average. Min, Max and Average is cleared when the function switch is changed.