

# Manual Supplement

Manual Title: 114,115 and 117 Users Supplement Issue: **3**  
Part Number: 2572573 Issue Date: 9/06  
Print Date: July 2006 Page Count: 3  
Revision/Date:

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 3, add the following after the last bullet:

- **Do not use the Auto Volts function to measure voltages in circuits that could be damaged by this function's low input impedance ( $\approx 3 \text{ k}\Omega$ ) (114 and 117 only).**

## Change #2

On page 6, replace number ① with the following:

①	Input terminal for measuring ac and dc current to 10 A.
---	---

On page 21, under AC Amps and DC Amps, replace the **Range** with the following:

6.000 A 10.00 A <sup>[3]</sup> 20 A overload 30 seconds maximum, 10 minutes rest minimum.
---

Add the following note to the bottom of the table:

[3] >10 A unspecified

## Change #3

On page 18, under **General Specifications**, replace the first sentence with the following:

Accuracy is specified for 1 year after calibration, at operating temperatures of 18 °C to 28 °C, with relative humidity at 0 % to 90 %.

## Change #4

On page 19, change the following:

From: **IP Rating** (dust and water protection)...IP52 with no vacuum applied

To: **IP Rating** (dust and water protection)...IP42