

Manual Supplement

Manual Title: 116 Users
Part Number: 2538688
Print Date: July 2006
Revision/Date:

Supplement Issue: **3**
Issue Date: 9/06
Page Count: 3


This supplement contains information
necessary to ensure the accuracy of
the above manual.

Change #1

On page 3, add the following after the last bullet:

- **Do not use the Auto Volts function to measure voltages in circuits that could be damaged by this function's low input impedance ($\approx 3 \text{ k}\Omega$).**

On page 13, replace the figure with the following:



Change #2

On page 17, under **General Specifications**, replace the first sentence with the following:

Accuracy is specified for 1 year after calibration, at operating temperatures of 18 °C to 28 °C, with relative humidity at 0 % to 90 %.

Change #3

On page 19, change the following:

From: **IP Rating** (dust and water protection)...IP52 with no vacuum applied

To: **IP Rating** (dust and water protection)...IP42