

# Manual Supplement

Manual Title:	12B,18,7 Series Service	Supplement Issue:	<b>3</b>
Part Number:	602730	Issue Date:	12/99
Print Date:	August 1996	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1

On page 3-7, Table 3-2, Test No. 11,

Change: 400.0V dc  
To: 400.0V dc<sup>4</sup>

Add footnote 4 to the bottom of the table:

4. Model 7-300 meter range is 300 Vdc.

On page 3-8, Table 3-3, Test No. 4, 12, and 13, change Switch Position,

From: Right  
To: Center

On pages 4-3 and 4-5, Tables 4-2 and 4-3:

Delete both A1 entries in their entirety.

On all MP2 entries, remove the word WINDOW from the Description.

On page 4-8, Table 4-4, change the part number for the first U2,

From: 601575  
To: 601757

## Change #2

On page 3-6, replace Table 3-1 with the following:

Equipment	Minimum Specifications	Recommended Model
DMM Calibrator	DC Voltage: 0-600V Accuracy: $\pm 0.25\%$ AC Voltage: 0-600V Accuracy: $\pm 0.5\%$ Frequency: 50-400 Hz Resistance: 1.0 $\Omega$ – 40 M $\Omega$ Accuracy: $\pm 0.25\%$ Capacitance: 0-1.000 $\mu$ F Accuracy: $\pm 0.5\%$	Fluke 5520A
<b>ALTERNATE</b>		
Equipment	Minimum Specifications	Recommended Models
DMM Calibrator	DC Voltage: 0-600V Accuracy: $\pm 0.25\%$ AC Voltage: 0-600V Accuracy: $\pm 0.5\%$ Frequency: 50-400 Hz	Fluke 5700A
Decade Resistor	Resistance: 1.0 – 40 M $\Omega$ Accuracy: $\pm 0.25\%$	General Resistance RDS-77B and 41B
Decade Capacitor	Capacitance: 0-1.000 $\mu$ F Accuracy: $\pm 0.5\%$	GenRad 1412-BC