

Manual Supplement

Manual Title: 1520 Calibration
Part Number: 1592594
Print Date: November 2000
Revision/Date:

Supplement Issue: **4**
Issue Date: 10/02
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 15, in the **Replacement Parts and Optional Accessories** table,

Replace:

Fuse, 6 mm x 32 mm (0.25 x 1.25 in), 0.5 A, 750 V, Fast Acting	1556096
--	---------

With:

⚠ Fuse, 6 mm x 32 mm (0.25 x 1.25 in), 0.5 A, 660 V, Fast Acting, 50 A Minimum Interrupt rating	NA
OR	
⚠ Fuse, 6 mm x 32 mm (0.25 x 1.25 in), 0.5 A, 750 V, Fast Acting	1556096
⚠ For safety, use exact replacement only.	

Change #2

On page 8, under **Discharge Circuit Test**, replace the last sentence in step 3 with the following:

Check that the DMM reading for each position is between 1600 and 3600Ω.

Change #3

On page 15, in the **Replacement Parts and Optional Accessories** table change the part number for the Battery Cover,

From: 1274679

To: 1670088