

# Manual Supplement

Manual Title: 1587/1577 Users  
Part Number: 2401027  
Print Date: April 2005  
Revision/Date:

Supplement Issue: **2**  
Issue Date: 4/05  
Page Count: 1

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 30, under **1587 Lowpass Filter Voltage**, replace the last two columns with the following:

50 Hz to 60 Hz ± (% of Rdg + Digits)	60 Hz to 400 Hz ± (% of Rdg + Digits)
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (2 % + 3)	+ 2 % + 3 - 6 % - 3

## Change #2

On Page 4, under **Battery Saver (Sleep Mode)**, replace the last sentence of the first paragraph with the following:

The Meter comes out of Sleep mode when a key is pressed or when the rotary switch is turned.