

Manual Supplement

Manual Title: 1587/1577 Users
Part Number: 2401027
Print Date: April 2005
Revision/Date:

Supplement Issue: **3**
Issue Date: 6/05
Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 30, under **1587 Lowpass Filter Voltage**, replace the last two columns with the following:

50 Hz to 60 Hz ± (% of Rdg + Digits)	60 Hz to 400 Hz ± (% of Rdg + Digits)
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (1 % + 3)	+ 2 % + 3 - 6 % - 3
± (2 % + 3)	+ 2 % + 3 - 6 % - 3

Change #2

On Page 4, under **Battery Saver (Sleep Mode)**, replace the last sentence of the first paragraph with the following:

The Meter comes out of Sleep mode when a key is pressed or when the rotary switch is turned.

Change #3

On page 29, under **General Specifications**, replace the Electromagnetic Compatibility specification with the following:

In an RF field of 3 V/M, accuracy = specified accuracy except in temperature: accuracy = specified accuracy ± 5 °C (9 °F). (EN 61326-1:1997).