

Manual Supplement

Manual Title:	165X Users	Supplement Issue:	3
Print Date:	September 2003	Issue Date:	6/05
Revision/Date:	1, 6/04	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	165X
CD Rev. & Date:	1, 6/2004
CD PN:	1674214

Change #1

On page 31, under step number 6, replace the second bullet with the following:

- The secondary (lower) display shows the fault voltage (N to PE) related to the rated residual current.

Change #2

On page 64, under *Tripping Current Measurement/Ramp Test ($I_{\Delta N}$)*, add an * to 50% to 110% of RCD's rated current and add the following footnote:

*30% to 110% for Type A

Change #3 ECO W1056056

On page 59, under *Insulation Resistance Measurement*, Model 1651:

Change: 500-1000 V

To: 250-500-1000 V