

Manual Supplement

Manual Title: 1735 Users
Part Number: 2560330
Print Date: March 2006
Revision/Date: 1, 4/06

Supplement Issue: 1
Issue Date: 8/06
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 10, prior to **Basic Operation**, add the following:

Charging the Internal Battery

Before you use the 1735 Logger, charge the internal battery as follows:

1. Switch the 1735 Battery Charger to either 115V or 230V as appropriate.
2. With the Logger power off, attach the BC1735 battery charger to a power outlet and then to the 1735.
3. Charge the 1735 for 5 hours before using it for the first time.
4. On subsequent use, turn the Logger power on before you connect the BC1735 Battery Charger.

This helps insure that the fast charge mode is activated. If the Logger does not turn on due to a discharged battery, recharge the battery for 5 hours with the Logger power off as described in steps 2 and 3 above.

Change #2

On page 52, under **EMC** replace the Emission, with the following:

Emission: IEC/EN 61326-1:1997+A1:1998+A2:2000 class B