

Manual Supplement

| | | | |
|----------------|---------------------|-------------------|------|
| Manual Title: | 183/185 Calibration | Supplement Issue: | 1 |
| Part Number: | 1610436 | Issue Date: | 3/04 |
| Print Date: | May 2001 | Page Count: | 1 |
| Revision/Date: | 1, 6/02 | | |

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 22, Table 24, under the Display Reading column for the 183, replace steps 20 through 27:

| |
|--------------------|
| 45.87 to 46.13 |
| 459.58 to 460.42 |
| -460.42 to -459.58 |
| 4.5958 to 4.6042 |
| 45.958 to 46.042 |
| 459.58 to 460.42 |
| 0998.3 to 1001.7 |
| -1001.7 to -0998.3 |

Change #2

On page 24, Table 24, change the following and add the footnote:

From: Capacitance Test

To: **Capacitance Test

**Variations in test equipment can cause erroneous readings. Use a fixed value capacitance standard if instability occurs.