

# Manual Supplement

Manual Title:	187/189 Calibration	Supplement Issue:	1
Part Number:	1584337	Issue Date:	12/00
Print Date:	November 2000	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

# Change #1

On page 10, Table 2, make the following changes:

Change:

Serial Cable, Infrared		Fluke P/N 1578406
------------------------	--	-------------------

To:

Serial Cable, Infrared/W IR Cable Adapter		Fluke P/N 1590638
---	--	-------------------

Delete:

IR Cable Adapter		Fluke P/N 1576688
------------------	--	-------------------

On pages 19 & 20, Table 4, steps 27, 49 & 50 change **Display Lower & Upper Limits** to:

Step	Lower Limit	Upper Limit
27	-0.02	0.02
49	-4987.3	-5012.7
50	-0.2	0.2

On page 25, under **Temperature Calibration**, replace the second bullet with:

- Infrared Serial Cable Kit, Fluke P/N 1590638

Under Optional equipment, delete the following:

- IR Cable Adapter, Fluke P/N 1576688