

Manual Supplement

Manual Title:	27 Service	Supplement Issue:	1
Part Number:	690206	Issue Date:	3/05
Print Date:	November 1998	Page Count:	1
Revision/Date:	1, 9/03		

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 1-4, add the following Warning prior to the Caution.

⚠Warning

To prevent hazard, do not use or apply voltage in wet locations and be sure to dry this product properly prior to use.

The fluke model 27 case is water tight enclose effective to 2PSI (a water depth of 5 feet). If the product has been submerged or is wet by indications of surface moisture, remove the moisture by allowing the product to properly dry before use. Moisture or water located within the terminals, seams or product surface may result in a hazard for applied voltages.