

Manual Supplement

Manual Title:	5500A Operator	Supplement Issue:	1
Print Date:	December 1994	Issue Date:	1/07
Revision/Date:	11, 7/06	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	5500A/5520A
CD Rev. & Date:	2, 6/2006
CD PN:	1627768

Change #1, 39296

On page 1-21, under **Temperature Calibration (RTD) Specifications**, under **RTD Type** change:

From: Pt 395, 100 Ω

To: Pt 385, 100 Ω

Change #2

On page 4-7, under **Zeroing the Calibrator**, add the following note between the first and second paragraph:

Note

If the 5500A is operated outside the range of $t_{cal} \pm 5$ °C, then the temperature coefficient defined in the General Specifications, Chapter 1 of this manual, must be calculated and added to the absolute uncertainties. Zeroing the 5500A is still required.