

Manual Supplement

Manual Title:	5500A/EP Users	Supplement Issue:	1
Part Number:	602562	Issue Date:	3/99
Print Date:	August 1996	Page Count:	1
Revision/Date:	3, 4/98		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 1, in the second paragraph,

Change: Version 4.3

To: Version 5.12e

On page 3, make the following changes:

In the second paragraph,

Change: Version 4.3

To: Version 5.12e

Under **Installing the Software**, delete the *Note*.

On page 4, under Establishing Traceability Information, in the second paragraph,

Change: 4.2 modes

To: 5.12e modes

On page 16, under **Troubleshooting**, delete the second bullet.

On page 18, under **DC Voltage**, Cardinal Point Output,

Change: 0.99V to 1.01V, and -1.00V to -1.01V

To: \pm (0.99V to 1.101V)