

Manual Supplement

Manual Title: 561 Users
Part Number: 2562924
Print Date: February 2006
Revision/Date:

Supplement Issue: 1
Issue Date: 3/06
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 9, prior to Figure 8, add the following note:

Note

The 561 stays on for 10 minutes after trigger release while the probe is inserted with laser off.