



Certificate of Compliance

Certificate: 70011314

Master Contract: 162029

Project: 70011314

Date Issued: September 24, 2014

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Thomas Smith

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, AScT.

PRODUCTS

CLASS 8721 06 - LABORATORY EQUIPMENT - Electrical

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

3 Phase Power Calibrator, Fluke Calibration model: 6003A, rated: 115/230 V, 47-63 Hz, 2250 VA.

Notes:

1. The above model is cord connected, Equipment Class I, Pollution Degree 2.
2. Mode of operation: Continuous.
3. Environmental Conditions: Normal: 5 to 40 C, 2000 m max, 80% to temperatures up to 31 °C decreasing linearly to 50% RH at 40 °C.



Certificate: 70011314

Master Contract: 162029

Project: 70011314

Date Issued: September 24, 2014

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards:

EN 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

EN 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits