

# Manual Supplement

Manual Title: 77,75,73,70,23,21 Calibration  
Part Number: 659141  
Print Date: November 1997  
Revision/Date: 1, 4/99

Supplement Issue: 2  
Issue Date: 10/00  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1

On page 11, Table 5, under Decades of 1: change the Display Reading for the 23,77 as follows,

Step 4,	From: 9.994 to 10.06
	To: 9.94 to 10.06
Step 5,	From: 99.94 to 100.6
	To: 99.4 to 100.6

## Change #2

On page 9, step 2,

Change:	Referring to Table 2...
To:	Referring to Table 3...

On page 10, under **Resistance Test**, step 3,

Change:	.....value indicated in steps 1 through 6.
To:	.....value indicated in steps 1 through 7.