

Supplement

700P2X Series Instruction Sheet

Pat Number:	601093	Supplement Issue:	1
Print Date:	Aug. 1995	Issue Date:	8/00
Revision/Date:	1, 3/00	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1

Under **Recommended Measurement Technique**
add the following note:

Note

Low range pressure modules may be sensitive to gravity. For best results, pressure module 30 psi and below should be held at the same physical orientation from the time they are zeroed until the measurement is complete.