

Supplement

Title: 705 Inst.Sht. Supplement Issue: 1
Print Date: August 1998 Issue Date: 5/07
Revision/Date: 2, 6/03 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 705 & 707
CD Rev. & Date: 5/2003
CD PN: 2088974

Change #1

Under ***Specifications***, change the ***Input/Output Protection***, to:

Input/output protection on mA range: self resetting, nonreplaceable 0.1 A fuse.

Change #2, 41512

Under ***General Specifications***, following the **Weight** specification, add:

Protection Class: Pollution Degree II