

Manual Supplement

Manual Title: 724 Users
Print Date: February 2000
Revision/Date: 1, 8/03

Supplement Issue: 1
Issue Date: 9/03
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 41, add the following to the *Temperature, Thermocouples* table:

Type	Range	Measure ans Source Accuracies (ITS-90)
XK	-200 to 800 °C	-200 to -100 ± 0.5 °C -100 to 800 ± 0.6 °C
BP	0 to 2500 °C	0 to 800 ± 1.2 °C 800 to 2500 ± 2.5 °C