

Manual Supplement


Manual Title:	725Ex Calibration	Supplement Issue:	1
Part Number:	2406553	Issue Date:	3/05
Print Date:	January 2005	Page Count:	1
Revision/Date:			


This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:


1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.


Change #1

On page 8, under **General Specifications, Product Compliance**, change the following:

From: **CE**  II 1 G EEx ia [ia] IIB 171 °C
0344 KEMA 04ATEX1303X

 Class I Div. 1 Groups B,C, and D
LR110460 Class I Zone 0 Aex/Ex ia [ia] IIB 171 °C
2004.1573226

To: **CE**  II 1 G EEx ia IIB 171 °C
0344 KEMA 04ATEX1303X

 Class I Div. 1 Groups B,C, and D
LR110460 Class I Zone 0 Aex/Ex ia IIB 171 °C
2004.1573226