Manual Supplement

Manual Title: 725 Users Supplement Issue: **2**Print Date: October 1998 Issue Date: 5/07
Revision/Date: 3, 5/04 Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 725/726 Product Manuals

CD Rev. & Date: 4, 9/2005 CD PN: 1549615



725 Users Manual Supplement

Change #1

On page 59, under RTD Excitation (simulation) change the following:

From:

PT 200-385	0.15 to 3.0 ma
------------	----------------

To:

Under *Temperature, RT Ranges, and Accuracies (ITS-90)* replace the **Allowable excitation current (source)** with the following:

Ni120, Pt100-385, Pt100-392, Pt100-JIS: 0.15 to 3.0 mA Pt200-385, Pt500-385: 0.05 to 0.80 mA; Pt1000-385: 0.05 to 0.40 mA

12/04

725 Users Manual Supplement

Change #2, 40767

On page 60, under *General Specifications*, add the following:

Battery life Battery life under load value	25 hours typical 12 mA in 500 Ω = 12 hrs
Protection Class	Pollution Degree II

2 5/07