

# Manual Supplement

Manual Title:	744 Users	Supplement Issue:	2
Part Number:	691287	Issue Date:	10/00
Print Date:	September 1998	Page Count:	2
Revision/Date:	1, 2/99		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 78, replace step 24, and delete step 25.

24. Exercise the switch by varying the applied pressure. Adjust the switch until the set point is correct.

On page 112, change the second and third entries,

From: 11.00 Hz to 109.99 Hz                      0.01 Hz

To: 11.00 Hz to 109.99 Hz                      0.1 Hz

From: 110.0 Hz to 1099.9 Hz                      0.01 Hz

To: 110.0 Hz to 1099.9 Hz                      0.1 Hz

On page 113, change the **Type E**, third entry,

From: -200 to -100    0.5    0.8    0.3    0.4

To: -100 to 600    0.3    0.4    0.3    0.4

On page 116, change the **Range** on the 120 $\Omega$  Ni(672) Type,

From: -200 to 260

To: -80 to 260

On page 117, change the **Range** on the 100 $\Omega$  Pt(3916) Type,

From: 0 to 360

To: 0 to 630

## Change #2

On page 99, under ***Calibration Data***, replace the first sentence with:

The date the calibrator was last calibrated shows on the calibration sticker. The date the calibrator was last adjusted is shown on the last screen in Setup mode.