

Manual Supplement

Manual Title:	744 HART Mode User	Supplement Issue:	2
Part Number:	691292	Issue Date:	9/03
Print Date:	December 1998	Page Count:	2
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 7, Table 1, add the following after the Sensor (read only) row:


(Temperature) Sensor (read/write)	•	•		
--------------------------------------	---	---	--	--

On page 10, under Sensor, change the first sentence and add a new paragraph as follows:

- Change:** From the Active Device screen press the **Setup** and **Sensor** softkeys to access the read only Sensor Setup screen.
- To:** From the Active Device screen press the **Setup** and **Sensor** softkeys to access the Sensor Setup screen.
- Add:** You can also program the Sensor Type and Sensor Connection registers in a supported temperature transmitter using this screen. (Screen examples do not reflect Temperature device menu items.)

Change #2

On page 9, replace Figure 2, with the following:



pe03f.eps