Manual Supplement

Manual Title: 74X Series Calibration Supplement Issue: 1
Part Number: 602505 Issue Date: 4/06
Print Date: February 1997 Page Count: 1

Revision/Date: 2, 4/99

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

- 1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
- 2. No revision letter is indicated at the beginning of the change.



74X Series Calibration Manual Supplement

Change #1

On page 6, replace the bottom portion of the *Frequency Sourcing* table with the following:

Waveform Choices: Zero-symmetric sine or positive square wave, 50% duty cycle.

Amplitude: 0.1 to 10 V pk (square wave, 0 V reference)

0.1 to 10 V p-p (sinewave)

Amplitude Accuracy:

0 Hz to 1099 Hz: 3% of output + 0.5% f.s. (squarewave) **1.0 Hz to 1099 Hz:** 3% of output + 0.5% f.s (sinewave) **1.1 kHz to 10.9 kHz:** 10% of output + 0.5% f.s.

11 kHz to 50 kHz: 30% of output + 0.5% f.s.

Maximum Input Voltage: 30 V dc

Under *Temperature, Thermocouples* Type E, delete the entire third row.

4/06