

Manual Supplement

Manual Title: 77 Series IV Users
Part Number: 2695884
Print Date: September 2006
Revision/Date:

Supplement Issue: 1
Issue Date: 10/06
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1 - 38525

On page 4, under **MIN MAX AVG Recording Mode**, under the note replace the last part of both sentences with the following:

.....changes longer than 350 ms in duration.

On page 13, under **Function**, replace the MIN MAX AVG, row with the following:

MIN MAX AVG	For dc functions, accuracy is the specified accuracy of the measurement function ± 12 counts for changes longer than 350 ms in duration. For ac functions, accuracy is the specified accuracy of the measurement function ± 40 counts for changes longer than 350 ms in duration.
-------------	--

Change #2

On page 5, under **Manual Ranging and Autoranging**, replace the second paragraph under the note with the following:

*If you press **RANGE** while in MIN MAX_AVG, the meter beeps twice, indicating an invalid operation, and the range does not change.*