

Manual Supplement

Manual Title:	79/26 Series III Calibration Information	Supplement Issue:	1
Part Number:	659349	Issue Date:	2/01
Print Date:	December 1997	Page Count:	1
Revision/Date:	2, 3/00		

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 15, delete the following section in its entirety:

Testing the LO-Ohms Function

On page 20, Table 12, change the F2 part number,

From: F11 A, 1000 VAC/DC, Min Interrupt Rating 17 kA 943118

To: F11 A, 1000 VAC/DC, Min Interrup Rating 20 kA 803293