

# Manual Supplement

Manual Title:	80 Series V Getting Started	Supplement Issue:	1
Part Number:	2101973	Issue Date:	1/06
Print Date:	May 2004	Page Count:	1
Revision/Date:	1, 8/04		

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 10, Table 5, add the following:

<b>F8-Err</b>	Invalid model. Have Meter serviced.
<b>OPEn</b>	Open thermocouple is detected.