

Manual Supplement

Manual Title: 80 Series V Users
Print Date: May 2004
Revision/Date: 1, 8/04

Supplement Issue: 1
Issue Date: 1/06
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 80 Series V
CD Rev. & Date: 2, 3/2005
CD PN: 2101999

Change #1

On page 12, Table 5, add the following:

F8-Err	Invalid model. Have Meter serviced.
OPEn	Open thermocouple is detected.

On page 16, under **Measuring Temperature (87)**, in the paragraph following the caution,

Change: When there is no thermocouple connected, the display also reads OL.

To: When there is no thermocouple connected, the display shows OPEn for Meters above serial number (90710501) and OL for meters below serial number (90710501).

Note

To locate the serial number remove the Meter from the holster. The serial number is on the back of the Meter.