

Manual Supplement

| | | | |
|----------------|-------------------------|-------------------|------|
| Manual Title: | 80 Series V Calibration | Supplement Issue: | 1 |
| Part Number: | 2102915 | Issue Date: | 1/06 |
| Print Date: | September 2004 | Page Count: | 1 |
| Revision/Date: | 1, 2/05 | | |

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 16, Table 13, add the following:

| | |
|---------------|-------------------------------------|
| F8-Err | Invalid model. Have Meter serviced. |
| OPEn | Open thermocouple is detected. |