

Manual Supplement

Manual Title: 8508A Service
Print Date: September 2004
Revision/Date: 2, 3/13

Supplement Issue: **3**
Issue Date: 1/15
Page Count: 2











This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	8508A
CD Rev. & Date:	6, 3/2013
CD PN:	2150143

Change #1, 66660

On page 1-3, under **Safety Considerations**, remove the first paragraph.

On page 1-6, replace the **Symbols** table with:

Symbol	Description
	Risk of danger. Important information. See manual.
	Hazardous voltage. Voltage >30 V dc or ac peak might be present.
	AC (Alternating Current).
	Power ON / OFF
	Functional Earth ground.
	This product complies with the WEEE Directive (2002/96/EC) marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste. Product Category: With reference to the equipment types in the WEEE Directive Annex I, this product is classed as category 9 "Monitoring and Control Instrumentation" product. Do not dispose of this product as unsorted municipal waste. Go to Fluke's website for recycling information.
	Recycle.
	Product conforms with the requirements of the applicable EC directives.
	Intertek ETL listed mark for CSA C22.2 61010-1, and UL 61010-1.
	Conforms to relevant South Korean EMC Standards

On page 4-14, Figure 4-9 and Table 4-1, remove IEC 127 from the Fuse Rating.

On page 4-15, Table 4-2, remove IEC 127 from the Fuse Rating.

Change #2, 427

On page 1-8, under **General Specifications**, replace the **Safety** and **EMC** with:

Safety IEC 61010-1: Overvoltage CAT II, Pollution Degree 2, 1420 Vpk max transient into measurement terminals.

Electromagnetic Compatibility (EMC)

IEC 61326-1..... (Controlled EM Environment); CISPR 11, Group 1, Class A

Group 1 Equipment..... Group 1 has intentionally generated and/or use conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.

Class A Equipment is equipment suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes. Caution - There may be potential difficulties in ensuring electromagnetic compatibility in other environments, due to conducted and radiated disturbances. Emissions which exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

USA (FCC)..... 47 CFR 15 subpart B, this product is considered an exempt device per clause 15.103

Korea (KCC) Class A Equipment (Industrial Broadcasting & Communication Equipment)
This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

Change #3, 430

On page 1-19, under the **Note to Performance Specifications**:

Change: [18] >2 GΩ Relative Humidity Operating <80 % to 30°C <70 % to 40°C

To: [18] >10 GΩ Typical