


Manual Supplement

Manual Title:	87V EX Calibration	Supplement Issue:	1
Print Date:	April 2006	Issue Date:	1/07
Revision/Date:		Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 39785

On page 8, under ***Electrical Characteristics of the Terminals***, replace the Diode row with the following:

	1000 V rms	< 7.9 V dc	3.000 V dc	1.0 mA typical
---	------------	------------	------------	----------------