



Certificate of Compliance

Certificate: 1932435 (LR 44340)

Master Contract: 162029

Project: 2643976

Date Issued: September 25, 2013

Issued to: **Fluke Corporation**
6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, ASCT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

5-1/2 Digit Multimeter, Fluke Model 8808A and Tektronix Model DMM4020; cord-connected, rated 100/120/220/240V, 47- 440Hz, 15VA.

Notes:

1. The above models are Equipment Class 1, Pollution Degree 2, Installation Category II, and Measurement Category 600V CATII; 1000V CATI.
2. Mode of operation: Continuous.
3. Environmental Conditions: 0 to 50°C; 2000 m max; <90% rH (10°Cto 30°C), ; <75% rH (30°Cto 40°C), ; <45% rH (40°Cto 50°C).



Certificate: 1932435 (LR 44340)

Master Contract: 162029

Project: 2643976

Date Issued: September 25, 2013

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3RD Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards

IEC Standard No. 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC Standard No. 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.