

Manual Supplement

Manual Title: 8808A Users
Print Date: July 2007
Revision/Date:

Supplement Issue: 1
Issue Date: 11/07
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 8808A
CD Rev. & Date: 11/2007
CD PN: 2713002

Change #1

On page 1-8, under **Relative Humidity (non-condensing)**, replace Operating with the following:

Operating..... Uncontrolled (< 10°C)
<90 % (10 °C to 28 °C)
<75 % (28 °C to 40 °C)
<45 % (40 °C to 50 °C)