



Certificate of Compliance

Certificate: 1837899 (LR 44340)

Master Contract: 162029

Project: 2643139

Date Issued: October 31, 2013

Issued to: **Fluke Corporation**
6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Bench DMM, Models Fluke 8845A, Fluke 8846A, Tektronix DMM4040, and Tektronix DMM4050; detachable cord set, rated 100/120 V ac, 200/240 V ac, 47-440 Hz, 28 VA max.

Notes:

1. The above models are Equipment Class I, Pollution Degree 2, Measurement Category 600 V Cat II / 1000V Cat I.
2. Mode of operation: Continuous.
3. Environmental Conditions: 0°C to 55°C, Humidity 90% to temperatures up to 28 °C decreasing linearly to 50% rH at 55 °C; max 90% rh, non condensing.



Certificate: 1837899 (LR 44340)

Master Contract: 162029

Project: 2643139

Date Issued: October 31, 2013

CONDITIONS FOR CONSIDERATION

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + Update No. 1:2008

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards:

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits