

Manual Supplement

Manual Title: Ti40,Ti45,Ti50,Ti55 Users Supplement Issue: 1
Print Date: May 2006 Issue Date: 6/06
Revision/Date: Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: Ti40,Ti45,Ti50,Ti55
CD Rev. & Date: 5/2006
CD PN: 2662914

Change #1

On page D-1, Table D-1, replace the Detector entry with:

| | |
|------------|---------------------------------------------------------------------------------|
| Detector | 160 X 120 Focal Plane Array (FPA), Vanadium Oxide (VOX) Uncooled Microbolometer |
| Frame Rate | 7.5 Hz |