



INDIETRO

PESI > MASSE E PESIERE DI PRECISIONE CALIBRATI IN CLASSE E2/F1/F2

MASSE E PESIERE DI PRECISIONE CALIBRATI IN CLASSE E2, F1, F2

Errore Massimo tollerato per la verifica	VALORE NOMINALE DELLA MASSA	CLASSE E₁ ± mg	CLASSE E₂ ± mg	CLASSE F₁ ± mg	CLASSE F₂ ± mg	CLASSE M₁ ± mg	CLASSE M₂ ± mg	CLASSE M₃ ± mg
	50 Kg	25	75	250	750	2500	7500	25000
	20 Kg	10	30	100	300	1000	3000	10000
	10 Kg	5	15	50	150	500	1500	5000
	5 Kg	2.5	7.5	25	75	250	750	2500
	2 Kg	1.0	3.0	10	30	100	300	1000
	1 Kg	0.5	1.5	5	15	50	150	500
	500 g	0.25	0.75	2.5	7.5	25	75	250
	200 g	0.10	0.30	1.0	3.0	10	30	100
	100 g	0.05	0.15	0.5	1.5	5	15	50
	50 g	0.030	0.10	0.30	1.0	3.0	10	30
	20 g	0.025	0.080	0.25	0.8	2.5	8	25
	10 g	0.020	0.060	0.20	0.6	2	6	20
	5 g	0.015	0.050	0.15	0.5	1.5	5	15
	2 g	0.012	0.040	0.12	0.4	1.2	4	12
	1 g	0.010	0.030	0.10	0.3	1.0	3	10
	500 mg	0.008	0.025	0.08	0.25	0.8	2.5	--
	200 mg	0.006	0.020	0.06	0.20	0.6	2.0	--
	100 mg	0.005	0.015	0.05	0.15	0.5	1.5	--
	50 mg	0.004	0.012	0.04	0.12	0.4	--	--
	20 mg	0.003	0.010	0.03	0.10	0.3	--	--
	10 mg	0.002	0.008	0.025	0.08	0.25	--	--
	5 mg	0.002	0.006	0.020	0.06	0.20	--	--
	2 mg	0.002	0.006	0.020	0.06	0.20	--	--
	1 mg	0.002	0.006	0.020	0.06	0.20	--	--